

RF MPE REPORT

Report No.: 20240517G07861X-W3

Product Name: Patient Monitor

Model No.: Alpha PM22

FCC ID: 2A3S6-IMI500

Applicant: Infinium Medical, Inc

Address: 12151 62nd St. N. #5 Largo Florida, 33773, United States

Dates of Testing: 05/13/2024 - 06/14/2024

Issued by: CCIC Southern Testing Co., Ltd.

Lab Location: Electronic Testing Building, No.43, Shahe Road, Xili Street, Nanshan District, Shenzhen, Guangdong, China.

Tel: 86 755 26627338 **E-Mail:** manager@ccic-set.com

This test report consists of 8 pages in total. It may be duplicated completely for legal use with the approval of the applicant. It should not be reproduced except in full, without the written approval of our laboratory. The client should not use it to claim product endorsement by CCIC-SET. The test results in the report only apply to the tested sample. The test report shall be invalid without all the signatures of testing engineers, reviewer and approver. Any objections must be raised to CCIC-SET within 15 days since the date when the report is received. It will not be taken into consideration beyond this limit.

Test Report

Product.....: Patient Monitor

Brand Name.....: **INFINIUM®**

Trade Name: **INFINIUM®**

Applicant.....: Infinium Medical,Inc

Applicant Address.....: 12151 62nd St. N. #5 Largo Florida, 33773,United States

Manufacturer.....: Infinium Medical,Inc

Manufacturer Address.....: 12151 62nd St. N. #5 Largo Florida, 33773,United States

Test Standards.....: 47 CFR Part 2.1091

Test Result.....: Pass

Tested by: kim Li 2024.06.28

Kim Li, Test Engineer

Reviewed by.....: Sun Jiaohui 2024.06.28

Sun Jiaohui, Senior Engineer

Approved by.....: Chris You 2024.06.28

Chris You, Manager

Table of Contents

| | |
|--|----------|
| 1. GENERAL INFORMATION | 5 |
| 1.1. EUT Description | 5 |
| 1.2. EUT Description | 6 |
| 1.3. Laboratory Facilities | 6 |
| 1.4. Laboratory Location | 6 |
| 2. TECHNICAL REQUIREMENTS SPECIFICATION IN CFR TITLE 47 PART 2.1091 ... | 7 |
| 2.1. Exposure Limits | 7 |
| 2.2. Predication of MPE limit at a given distance | 7 |
| 2.3. Evaluation Results | 8 |
| 2.4. Conclusion | 8 |

| Change History | | |
|----------------|------------|-------------------|
| Issue | Date | Reason for change |
| 1.0 | 2024.06.28 | First edition |
| | | |

1. GENERAL INFORMATION

1.1. EUT Description

| | |
|---------------------------------|--|
| Product Name | Patient Monitor |
| EUT supports Radios application | 2.4G WIFI |
| Frequency Range(Tx) | 2.4G WIFI: 2.412GHz ~ 2.462GHz |
| Bandwidth | 802.11b/g/n-HT20: 20MHz 802.11n-HT40: 40MHz |
| Modulation Type | 2.4G WIFI: DSSS (802.11b), OFDM (802.11g/n) |
| Antenna gain | 2.4G WIFI: 3.0dBi |
| Antenna Type | Internal Antenna |

Note 1: The information of antenna gain and cable loss is provided by the manufacturer and our lab is not responsible for the accuracy of the antenna gain and cable loss information.

1.2. EUT Description

EUT has been tested according to the following standards.

| No. | Identity | Document Title |
|-----|---|---|
| 1 | 47 CFR Part 1 | Practice and Procedure |
| 2 | 47 CFR Part 2 | Frequency Allocations and Radio Treaty Matters; General Rules and Regulations |
| 3 | KDB 447498 D01 General RF Exposure Guidance v06 | RF Exposure Procedures and Equipment Authorization Policies for Mobile and Portable Devices |
| 4 | OET Bulletin 65 Edition 97-01 | Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields |

1.3. Laboratory Facilities

FCC-Registration No.: CN1283

CCIC Southern Testing Co., Ltd EMC Laboratory has been registered and fully described in a report filed with the FCC (Federal Communications Commission). The acceptance letter from the FCC is maintained in our files. Designation Number: CN1283, valid time is until Jun. 30th, 2025.

ISED Registration: 11185A-1

CCIC Southern Testing Co., Ltd. EMC Laboratory has been registered by Certification and Engineering Bureau of Industry Canada for the performance of radiated measurements with Registration No. 11185A on Aug. 04, 2016, valid time is until Jun. 30th, 2025.

CAB number: CN0064

A2LA Code: 5721.01

CCIC-SET is a third party testing organization accredited by A2LA according to ISO/IEC 17025. The accreditation certificate number is 5721.01.

1.4. Laboratory Location

| | |
|---------------|---|
| Company Name: | CCIC Southern Testing Co., Ltd. |
| Address: | Electronic Testing Building, No.43, Shahe Road, Xili Street, Nanshan District, Shenzhen, Guangdong, China |

2. Technical Requirements Specification in CFR Title 47 Part 2.1091

2.1. Exposure Limits

The criteria listed in the following table shall be used to evaluate the environment impact of human exposure to radio frequency (RF) radiation as specified in 1.1307(b).

Table 1 to § 1.1310(e)(1) - Limits for Maximum Permissible Exposure (MPE)

| Frequency Range (MHz) | Electric Field Strength (V/m) | Magnetic Field Strength (A/m) | Power Density (mW/cm ²) | Averaging Time (minutes) |
|--|-------------------------------|-------------------------------|-------------------------------------|--------------------------|
| (i) Limits for Occupational/Controlled Exposure | | | | |
| 0.3-3.0 | 614 | 1.63 | *(100) | < 6 |
| 3.0-30 | 1824/f | 4.89/f | *(900/f ²) | < 6 |
| 30-300 | 61.4 | 0.163 | 1.0 | < 6 |
| 300-1500 | / | / | f/300 | < 6 |
| 1500-100,000 | / | / | 5 | < 6 |
| (ii) Limits for General Population/Uncontrolled Exposure | | | | |
| 0.3-1.34 | 614 | 1.63 | *(100) | < 30 |
| 1.34-30 | 824/f | 2.19/f | *(180/f ²) | < 30 |
| 30-300 | 27.5 | 0.073 | 0.2 | < 30 |
| 300-1500 | / | / | f/1500 | < 30 |
| 1500-100,000 | / | / | 1.0 | < 30 |

Note: f = frequency in MHz. * = Plane-wave equivalent power density.

2.2. Predication of MPE limit at a given distance

Refer to formulas on page 19 of OET Bulletin 65, Edition 97-01.

$$S = \frac{PG}{4\pi R^2}$$

Where:

S = power density (in appropriate units, e.g. mW/cm²)

P = power input to the antenna (in appropriate units, e.g., mW)

G = numeric gain of the antenna in the direction of interest relative to an isotropic radiator

R = distance to the centre of radiation of the antenna (appropriate units, e.g., cm)

2.3. Evaluation Results

Worst-Case mode Conducted Output Power Results for 2.4G WLAN

| Band | Mode | Frequency (MHz) | Maximum Output Power (dBm) | Max Tune up power (dBm) | Max Tune up power (mW) |
|-----------|---------|-----------------|----------------------------|-------------------------|------------------------|
| 2.4G WIFI | 802.11b | 2437 | 16.40 | 16±1 | 50.12 |

Calculation results: Worst-Case mode

| Band | Max Tune up power (dBm) | Antenna Gain (dBi) | Distance (cm) | Result (mW/cm ²) | Power Density (mW/cm ²) | Ratio |
|-----------|-------------------------|--------------------|---------------|------------------------------|-------------------------------------|-------|
| 2.4G WIFI | 17.0 | 3.0 | 20 | 0.02 | 1.00 | 0.02 |

2.4. Conclusion

According to the KDB 447498 D01 General RF Exposure Guidance v06 section 7.2 determine the device is exclusion from SAR test.

** END OF REPORT **