

## FCC MPE calculation Report

Product name : Portable ultrasonic flowmeter signal converter with  
Bluetooth connectivity

Applicant : Krohne Altimeter

FCC ID : 2A3Y8-UFC300P

Test report No. : P000029886 MPE calculation report Ver 2.0



Report number: P000029886 MPE calculation report Ver 2.0

## Laboratory information

### Accreditation

*Telefication complies with the accreditation criteria for test laboratories as laid down in ISO/IEC 17025:2017. The accreditation covers the quality system of the laboratory as well as the specific activities as described in the authorized annex bearing the accreditation number L021 and is granted on 30 November 1990 by the Dutch Council For Accreditation (RvA: Raad voor Accreditatie).*

Telefication is designated by the FCC as an Accredited Test Firm for compliance testing of equipment subject to Certification under Parts 15 & 18. The Designation number is: NL0001.

Telefication is a Wireless Device Testing laboratory recognized by Innovation, Science and Economic Development Canada to test to Canadian radio equipment requirements.

The Industry Canada company number for Telefication is: 4173A.

Telefication is a registered Conformity Assessment body (CAB) under the Japan-EC MRA (Agreement on Mutual Recognition between Japan and the European Community). The registration number is: 201.

### Documentation

The test report must always be reproduced in full; reproduction of an excerpt only is subject to written approval of the testing laboratory. The documentation of the testing performed on the tested devices is archived for 10 years at Telefication Netherlands.

### Testing Location

<b>Test Site</b>	Kiwa Telefication BV
<b>Test Site location</b>	Wilmersdorf 50 7327 AC Apeldoorn The Netherlands  Tel. +31 88998 3393
<b>Test Site FCC</b>	NL0001
<b>CABID</b>	NL0001

## Revision History

Version	Date	Remarks	By
v1.0	20-01-2022	Release version	KK
v2.0	17-03-2022	Applicant information changed	KK

## Table of Contents

<b>Revision History</b> .....	<b>2</b>
<b>1 General Description</b> .....	<b>4</b>
1.1    Applicant .....	4
1.2    Manufacturer .....	4
1.3    Tested Equipment Under Test (EUT).....	4
1.4    SAR Measurement Evaluation.....	5
1.4.1    Maximum Output Power .....	5
1.4.2    MPE Limits .....	5
1.4.3    MPE calculation .....	6
1.5    Summary .....	6

## 1 General Description

### 1.1 Applicant

**Client name:** Krohne Altometer  
**Address** Kerkeplaat 12, Dordrecht, The Netherlands  
**Telephone:** +31-786306233  
**E-mail:** e.vuelban@krohne.com  
**Contact name:** Edgar Vuelban

### 1.2 Manufacturer

**Manufacturer name:** Krohne Altometer  
**Address:** Kerkeplaat 12, Dordrecht, The Netherlands  
**Telephone:** +31-786306233  
**E-mail:** e.vuelban@krohne.com  
**Contact name:** Edgar Vuelban

### 1.3 Tested Equipment Under Test (EUT)

**Product name:** Portable ultrasonic flowmeter signal converter with Bluetooth connectivity  
**Brand name:** OPTISONIC 6300 P  
**Product type:** Flowmeter converter  
**FCC ID:** 2A3Y8-UFC300P  
**Software version:** ER2.x  
**Hardware version:** ER2.x

## 1.4 SAR Measurement Evaluation

### 1.4.1 Maximum Output Power

The maximum radiated power including antenna gain is shown as below.

Technology	Output power (dBm)
BT classic	10.55

\* from Telefication test report no: P000029886 001 Ver 1.00

### 1.4.2 MPE Limits

Limits for occupational/controlled exposure

Frequency Range (MHz)	Electric field strength (V/m)	Magnetic field strength (A/m)	Power density (mW/cm <sup>2</sup> )	Averaging time (minutes)
0.3 – 3.0	614	1.63	100 (see note 1)	≤6
3.0 – 30	1842/f	4.89/f	900/f <sup>2</sup> (see note 1)	≤6
30 – 300	61.4	0.163	1.0	≤6
300 – 1500	--	--	f/300	≤6
1500 – 100000	--	--	5	≤6

Limits for general population/uncontrolled exposure

Frequency Range (MHz)	Electric field strength (V/m)	Magnetic field strength (A/m)	Power density (mW/cm <sup>2</sup> )	Averaging time (minutes)
0.3 – 1.34	614	1.63	100 (see note 1)	≤30
1.34 – 30	824/f	2.19/f	180/f <sup>2</sup> (see note 1)	≤30
30 – 300	27.5	0.073	0.2	≤30
300 – 1500	--	--	f/1500	≤30
1500 – 100000	--	--	1.0	≤30

Notes :

f = frequency in MHz

1: plane wave equivalent power density

### 1.4.3 MPE calculation

As declared by the Applicant, the EUT is a wireless device used in a fixed application, at least 20 cm from any body part of the user or nearby persons.

Calculation method of RF Safety Distance:

$$PD = \frac{P_{out} * G}{4\pi r^2} = \frac{P(eirp)}{4\pi r^2}$$

Where:

PD = Power Density in  $mW/cm^2$

Pout = Output power in mW

G = Gain of antenna

R = Distance between observation point and centre of the radiator in cm

#### Calculation results

Technology	Frequency (MHz)	Max radiated power (mW)	Distance (cm)	Power density ( $mW/cm^2$ )	Limit ( $mW/cm^2$ )
BT Classic	2440	11.3	20	0.00225	1

### 1.5 Summary

Since MPE calculation is below the limit SAR test is not required.