

---

# Safety Human Exposure

## 1.1 Radio Frequency Exposure Compliance

### 1.1.1 Electromagnetic Fields

RESULT:

Pass

#### Test Specification

|                           |   |
|---------------------------|---|
| Test item                 | : Medical wearable in the form of a wristwatch for Parkinson's disease monitoring |
| Identification / Type No. | : Flytta  |
| FCC ID                    | : 2A6KGFLYTTA-1   |
| Test standard             | : CFR47 FCC Part 2: Section 2.1093<br>FCC KDB Publication 447498 v06              |

#### Measurement Record:

Since maximum peak output power of the transmitter is  $-2.82\text{dBm} \approx 0.52\text{mW} < \frac{3 \cdot d}{\sqrt{f}} = 9.7 \text{ mW}$ .

Hence the EUT is excluded from SAR evaluation according to FCC KDB Publication 447498 D01 General RF Exposure Guidance v06.