

Custom Band: CW, FRONT
Room Ambient Temperature: 23.0°C, Liquid Temperature: 22.0°C

Exposure Conditions

Band	Custom Band	TSL Permittivity	56.6
Frequency [MHz] / Channel Number	13.600 / 13600	TSL Conductivity [S/m]	0.737
Group / UID	CW / 0--	Phantom Section / TSL	Flat / Head Simulating Liquid
Conversion Factor	15.04	Test Distance [mm]	0.00

DASY Configuration

Probe Calibration Date	EX3DV4 - SN7313 2025-02-24	Phantom	ELI V8.0 (20deg probe tilt)
DAE Calibration Date	DAE4 Sn1591 2025-02-18	TSL Type	HBBL-600-10000
Software Version	16.4.0.5005		

Scan Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	150.0 x 180.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	15.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	0.072	0.026
psSAR10g [W/Kg]	0.034	0.005
Power Drift [dB]		-0.30
Dist 3dB Peak [mm]		2.4
M2/M1 [%]		74.4

