



7/19/2023

Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD 21046

To whom it may concern:

Curbell Medical Products Inc., hereby requests a Permissive Change Class II of the radio equipment certification for the WIFI Wi Link Network Controller certification number, FCC ID: 2A86D-PD11.

This product differs from the originally approved product in the following manor:

- (a) Associated digital circuitry. – No changes made to this section.
- (b) Functional capabilities. – No changes to this section.
- (c) Antenna Characteristics. – No changes to Characteristics, just orientation on PCB for our use.
- (d) Cosmetic differences – No changes to this section.
- (e) Other – SAR testing done for use closer than 20cm from human body

I do attest that the Field strength and/or RF Output readings remained the same or are lower than the originally approved product and that they did not increase therefore qualifying this product for this application type.

Sincerely,

A handwritten signature in black ink, reading "Gregory H. Walczak", enclosed within a rectangular box.

Gregory Walczak
Sr. Compliance/Applications Engineer
Curbell Medical Products Inc.