

TEST REPORT

Application No.: HKEM2205000461PF
Applicant: C-MER RainsOptics Limited
Address of Applicant: Room 607, 6/F, 17W of Hong Kong Science Park, No. 17 Science Park West Avenue, Shatin, N.T., Hong Kong
Equipment Under Test (EUT):
EUT Name: Pocket Smartphone Ophthalmoscope
Model No.: 20D001
FCC ID: 2A8CG-FUNDUSLINK
IC: 28897- FUNDUSLINK
HVIN: FDL1.0
Standard(s) : 47 CFR Part 15, Subpart C
RSS-247 Issue 2
RSS-Gen Issue 5
Date of Receipt: 2022-06-20
Date of Test: 2022-06-20 to 2022-06-28
Date of Issue: 2022-06-28

Test Result:	The submitted sample was found to comply with the test requirement
---------------------	--



Law Man Kit
EMC Manager

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request and accessible at <http://www.sgs.com/en/Terms-and-conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-conditions.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. The document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Revision Record				
Version	Chapter	Date	Modifier	Remark
01		2022-06-28		Original

Authorized for issue by:				
				
		Chan Chun Lok /Project Engineer		Date: 2022-06-28
				
		Law Man Kit /Reviewer		Date: 2022-06-28

2 Test Summary

Radio Spectrum Technical Requirement				
Item	Standard	Method	Requirement	Result
RF Exposure	47 CFR Part 1.1307, 47 CFR Part 2.1093, KDB 447498 D04	KDB447498D04	KDB447498D04	PASS
RF Exposure	RSS102 Issue 5	RSS-102 Section 2.5.1	RSS-102 Section 2.5.1	PASS

Declaration of EUT Family Grouping:

N/A

Abbreviation:

Tx: In this whole report Tx (or tx) means Transmitter.

Rx: In this whole report Rx (or rx) means Receiver.

RF: In this whole report RF means Radiated Frequency.

CH: In this whole report CH means channel.

Volt: In this whole report Volt means Voltage.

Temp: In this whole report Temp means Temperature.

Humid: In this whole report Humid means humidity.

Press: In this whole report Press means Pressure.

N/A: In this whole report not application.

3 Contents

	Page
1 COVER PAGE	1
2 TEST SUMMARY.....	3
3 CONTENTS	4
4 GENERAL INFORMATION.....	5
4.1 DETAILS OF E.U.T.....	5
4.2 DESCRIPTION OF SUPPORT UNITS.....	6
4.3 MODULATION CONFIGURE.....	6
4.4 TEST LOCATION	7
4.5 TEST FACILITY	7
4.6 DEVIATION FROM STANDARDS	7
4.7 ABNORMALITIES FROM STANDARD CONDITIONS	7
5 RADIO SPECTRUM TECHNICAL REQUIREMENT	8
5.1 RF EXPOSURE	8
5.1.1 <i>Test Requirement:</i>	8
5.1.1 <i>EUT RF Exposure Evaluation.....</i>	<i>Error! Bookmark not defined.</i>
5.2 RF EXPOSURE	9
5.2.1 <i>Test Requirement:</i>	9
5.2.2 <i>Conclusion</i>	10
6 PHOTOGRAPHS	11
6.1 EUT CONSTRUCTIONAL DETAILS (EUT PHOTOS)	11

4 General Information

4.1 Details of E.U.T.

Power supply:	Adaptor Model: UES06WNCPU-050100SPA Input: AC 100 V - 240 V, 50/60 Hz, 0.2 A Output: DC 5.0 V, 1.0 A or Battery Model: DTP123060 Output: DC 3.7 V
Test voltage:	DC 3.7 V
Cable:	Power Cable: 100.5 cm unshielded 2-wire USB
Antenna Gain:	Maximum 3.77 dBi at 2402 MHz, 2.27 dBi at 2442 MHz, 0.67 dBi at 2480
Antenna Type:	PCB Antenna
Bluetooth Version:	V5.2 LE
Channel Separation:	2MHz
Modulation Type:	GFSK
Number of Channels:	40
Operation Frequency:	2402MHz to 2480MHz
Power Class:	<10mW
Series No.:	BAA220400010
Firmware Version:	FDL1.00
Hardware Version:	FDL1.0

Frequency List:

Channel	Frequency (MHz)	Channel	Frequency (MHz)
0	2402	20	2442
1	2404	21	2444
2	2406	22	2446
3	2408	23	2448
4	2410	24	2450
5	2402	25	2452
6	2414	26	2454
7	2416	27	2456
8	2418	28	2458
9	2420	29	2460
10	2422	30	2462
11	2424	31	2464
12	2426	32	2466
13	2428	33	2468
14	2430	34	2470
15	2432	35	2480
16	2434	36	2474
17	2436	37	2476
18	2438	38	2478
19	2440	39	2480

The frequencies under test are bolded.

4.2 Description of Support Units

Description	Manufacturer	Model No.	Serial No.
Laptop	DELL	P75F	475LXQ2
PuTTY.exe	C-MER RainsOptics Limited	N/A	N/A

Note: The laptop and the software PuTTY.exe were for the control of the engineering mode.

4.3 Modulation configure

RF software:	PuTTY.exe			
Modulation	Packet	Packet Type	Packet Size	Power
GFSK	Default	Default	Default	Neg4dBm
Remark:				
1. Neg4dBm value was set in test software as maximum output power setting.				

4.4 Test Location

All tests were performed at:

SGS Hong Kong Limited
Unit 2 and 3, G/F, Block A, Po Lung Centre,
11 Wang Chiu Road, Kowloon Bay, Kowloon, Hong Kong
Tel: +852 2305 2570 Fax: +852 2756 4480

No tests were sub-contracted.

4.5 Test Facility

The test facility is recognized, certified, or accredited by the following organizations:

- **IAS Accreditation (Lab Code: TL-817)**

SGS Hong Kong Limited has met the requirements of AC89, IAS Accreditation Criteria for Testing Laboratories, and has demonstrated compliance with ISO/IEC Standard 17025:2017, General requirements for the competence of testing and calibration laboratories. This organization is accredited to provide the services specified in the scope of accreditation maintained on the IAS website (www.iasonline.org).

The report must not be used by the client to claim product certification, approval, or endorsement by IAS, NIST, or any agency of the Federal Government.

- **FCC Recognized Accredited Test Firm(CAB Registration No.: 514599)**

SGS Hong Kong Limited has been accredited and fully described in a report filed with the (FCC) Federal Communications Commission. The acceptance letter from the FCC is maintained in our files. Designation Number: HK0015, Test Firm Registration Number: 514599.

- **Industry Canada (Site Registration No.: 26103; CAB Identifier No.: HK0015)**

SGS Hong Kong Limited has been recognized by Department of Innovation, Science and Economic Development (ISED) Canada as a wireless testing laboratory. The acceptance letter from the ISED is maintained in our files. CAB Identifier No: HK0015, Site Registration Number: 26103.

4.6 Deviation from Standards

None

4.7 Abnormalities from Standard Conditions

None

5 Radio Spectrum Technical Requirement

5.1 RF Exposure

5.1.1 Test Requirement:

CFR 47 Part 1.1307, KDB 447498 D04 Interim General RF Exposure Guidance v01

Limit:

According to FCC Part1.1307 (b)(3): For single RF sources (i.e., any single fixed RF source, mobile device, or portable device, as defined in paragraph (b)(2) of this section): A single RF source is exempt if:

(A) The available maximum time-averaged power is no more than 1 mW, regardless of separation distance. This exemption may not be used in conjunction with other exemption criteria other than those in paragraph (b)(3)(ii)(A) of this section. Medical implant devices may only use this exemption and that in paragraph (b)(3)(ii)(A);

5.1.2 Conclusion

The worst-case e.i.r.p for FCC;

Operation mode	Channel	Frequency (MHz)	Conduct power (including Tune-up tolerance) (dBm)	Antenna Gain (dBi)	Worst Case e.i.r.p (mW)	Exemption Threshold (mW)
GFSK	Low	2402	-3.9	3.77	0.97	1
GFSK	Middle	2442	-3.6	2.27	0.74	1
GFSK	High	2480	-3.6	0.67	0.51	1

Thus, the EUT fulfil the 1mW exemption criteria.

5.2 RF Exposure

5.2.1 Test Requirement:

RSS-102 Section 2.5.1

Limit:

All transmitters are exempt from routine SAR and RF exposure evaluations provided that they comply with the requirements of sections 2.5.1 or 2.5.2 of RSS-102 Issue 5, March 2015. If the equipment under test (EUT) meets the requirements of sections 2.5.1 or 2.5.2, applicants are only required to submit a properly signed declaration of compliance.

SAR evaluation is required if the separation distance between the user and/or bystander and the antenna and/or radiating element of the device is less than or equal to 20 cm, except when the device operates at or below the applicable output power level (adjusted for tune-up tolerance) for the specified separation distance defined in Table 1.

Table 1: SAR evaluation — Exemption limits for routine evaluation based on frequency and separation distance

Frequency (MHz)	Exemption Limits (mW)				
	At separation distance of ≤5 mm	At separation distance of 10 mm	At separation distance of 15 mm	At separation distance of 20 mm	At separation distance of 25 mm
≤300	71 mW	101 mW	132 mW	162 mW	193 mW
450	52 mW	70 mW	88 mW	106 mW	123 mW
835	17 mW	30 mW	42 mW	55 mW	67 mW
1900	7 mW	10 mW	18 mW	34 mW	60 mW
2450	4 mW	7 mW	15 mW	30 mW	52 mW
3500	2 mW	6 mW	16 mW	32 mW	55 mW
5800	1 mW	6 mW	15 mW	27 mW	41 mW

Frequency (MHz)	Exemption Limits (mW)				
	At separation distance of 30 mm	At separation distance of 35 mm	At separation distance of 40 mm	At separation distance of 45 mm	At separation distance of ≥50 mm
≤300	223 mW	254 mW	284 mW	315 mW	345 mW
450	141 mW	159 mW	177 mW	195 mW	213 mW
835	80 mW	92 mW	105 mW	117 mW	130 mW
1900	99 mW	153 mW	225 mW	316 mW	431 mW
2450	83 mW	123 mW	173 mW	235 mW	309 mW
3500	86 mW	124 mW	170 mW	225 mW	290 mW
5800	56 mW	71 mW	85 mW	97 mW	106 mW

Output power level shall be the higher of the maximum conducted or equivalent isotropically radiated power (e.i.r.p.) source-based, time-averaged output power. For controlled use devices where the 8 W/kg for 1 gram of tissue applies, the exemption limits for routine evaluation in Table 1 are multiplied by a factor of 5. For limb-worn devices where the 10-gram value applies, the exemption limits for routine evaluation in Table 1 are multiplied by a factor of 2.5. If the operating frequency of the device is between two frequencies located in Table 1, linear interpolation shall be applied for the applicable separation distance. For test separation distance less than 5 mm, the exemption limits for a separation distance of 5 mm can be applied to determine if a routine evaluation is required.

For medical implants devices, the exemption limit for routine evaluation is set at 1 mW. The output power of a medical implants device is defined as the higher of the conducted or e.i.r.p to determine whether the device is exempt from the SAR evaluation.

5.2.2 Conclusion

By using linear interpolation to determine the limit for the worse separation distance of ≤ 5 mm at 2402 MHz:

$$\text{Limit} = 4 + (2450-2402) \times ((7-4) / (2450-1900)) = 4.26 \text{ mW}$$

2442 MHz

$$\text{Limit} = 4 + (2450-2442) \times ((7-4) / (2450-1900)) = 4.04 \text{ mW}$$

2480 MHz:

$$\text{Limit} = 4 + (2480-2450) \times ((2-4) / (3500-2450)) = 3.94 \text{ mW}$$

The e.i.r.p is calculated with the formula:

$$10^{\text{((Conductive power + Antenna gain)/10)}}$$

Refer to the conductive power measurement in report HKEM220500046102:

Operation mode	Channel	Frequency (MHz)	Conduct power (including Tune-up tolerance) (dBm)	Antenna Gain (dBi)	Worst Case e.i.r.p (mW)
GFSK	Low	2402	-3.9	3.77	0.97
GFSK	Middle	2442	-3.6	2.27	0.74
GFSK	High	2480	-3.6	0.67	0.51

Which are below the exemption limit.

6 Photographs

6.1 EUT Constructional Details (EUT Photos)

Refer to the appendices external, internal and setup photos.

- End of the Report -