

# EMC TEST REPORT

**Applicant** Alliedstar Medical Equipment Co., Ltd.

**FCC ID** 2A8SG-S300W

**Product** Alliedstar Sensa Intraoral Scanner

**Brand** Alliedstar Sensa

**Model** S300W

**Report No.** EFTA25080158-IE-02-E1

**Issue Date** September 9, 2025

Eurofins TA Technology (Shanghai) Co., Ltd. tested the above equipment in accordance with the requirements in **FCC Code CFR47 Part15B (2024)/ ANSI C63.4-2014**. The test results show that the equipment tested is capable of demonstrating compliance with the requirements as documented in this report.

---

*Prepared by: Liu Wei*

*Approved by: Xu Kai*

**Eurofins TA Technology (Shanghai) Co., Ltd.**

*Building 3, No.145, Jintang Rd, Pudong Shanghai, P.R.China*

*TEL: +86-021-50791141/2/3*

*FAX: +86-021-50791141/2/3-8000*

## Table of Contents

1	Test Laboratory .....	4
1.1	Notes of the Test Report.....	4
1.2	Test Facility .....	4
1.3	Testing Location.....	4
2	General Description of Equipment Under Test .....	5
2.1	Applicant and Manufacturer Information .....	5
2.2	General Information .....	5
2.3	Applied Standards .....	6
2.4	Test Mode .....	7
3	Test Case Results .....	8
3.1	Radiated Emission .....	8
3.2	Conducted Emission .....	13
5	Uncertainty Measurement.....	16
6	Main Test Instruments .....	17
ANNEX A: The EUT Appearance .....		18
ANNEX B: Test Setup Photos .....		19
ANNEX C: Product Change Description .....		20

## Summary of measurement results

Number	Test Case	Clause in FCC Rules	Conclusion
1	Radiated Emission	FCC Part15.109, ANSI C63.4-2014	PASS
2	Conducted Emission	FCC Part15.107, ANSI C63.4-2014	PASS
Date of Testing: August 8, 2025 ~ August 9, 2025			
Date of Sample Received: June 18, 2025			
Note:			
1. All indications of Pass/Fail in this report are opinions expressed by Eurofins TA Technology (Shanghai) Co., Ltd. based on interpretations and/or observations of test results.			
Measurement Uncertainties were not taken into account and are published for informational purposes only.			

**S300W (Report No.: EFTA25080158-IE-02-E1) is a variant model of S300W (Report No.: R2403A0226-E1V1).**

**This product only changes Product Name, Brand, HW Version, Battery and PCB (A resistor at the fan was modified (R128 resistor, with the resistance value changed from 178K to 240K)). This report tests Radiated Emission and Conducted Emission and is recorded in the report. The detailed product change description please refer to the *Difference Declaration Letter*.**

# 1 Test Laboratory

## 1.1 Notes of the Test Report

This report shall not be reproduced in full or partial, without the written approval of **Eurofins TA Technology (Shanghai) Co., Ltd.** The results documented in this report apply only to the tested sample, under the conditions and modes of operation as described herein. Measurement Uncertainties were not taken into account and are published for informational purposes only. This report is written to support regulatory compliance of the applicable standards stated above.

## 1.2 Test Facility

### **FCC (Designation number: CN1179, Test Firm Registration Number: 446626)**

Eurofins TA Technology (Shanghai) Co., Ltd. has been listed on the US Federal Communications Commission list of test facilities recognized to perform measurements.

### **A2LA (Certificate Number: 3857.01)**

Eurofins TA Technology (Shanghai) Co., Ltd. has been listed by American Association for Laboratory Accreditation to perform measurement.

## 1.3 Testing Location


Company: Eurofins TA Technology (Shanghai) Co., Ltd.  
Address: Building 3, No.145, Jintang Rd, Pudong Shanghai, P.R.China  
City: Shanghai  
Post code: 201201  
Country: P. R. China  
Contact: Xu Kai  
Telephone: +86-021-50791141/2/3  
Fax: +86-021-50791141/2/3-8000  
Website: <https://www.eurofins.com/electrical-and-electronics>  
E-mail: [Kain.Xu@cpt.eurofinscn.com](mailto:Kain.Xu@cpt.eurofinscn.com)

## 2 General Description of Equipment Under Test

### 2.1 Applicant and Manufacturer Information

<b>Applicant</b>	Alliedstar Medical Equipment Co., Ltd.
<b>Applicant address</b>	No.222, West Section 3, Waihuan Rd, Yanjiang District, Ziyang, China
<b>Manufacturer</b>	Alliedstar Medical Equipment Co., Ltd.
<b>Manufacturer address</b>	No.222, West Section 3, Waihuan Rd, Yanjiang District, Ziyang, China

### 2.2 General Information

EUT Description			
Device Type	Portable Device		
Model	S300W		
Lab internal SN	EFTA25080158-IE-01/S01		
HW Version	D		
SW Version	1.0		
Power Rating	DC 3.6V from battery or DC 12V from Adapter.		
Connecting I/O Port(s)	Please refer to the User's Manual.		
Antenna Type	Internal Antenna		
Frequency	Band	Tx (MHz)	Rx (MHz)
	Wi-Fi 5G (U-NII-1)	5150 ~ 5250	5150 ~ 5250
	Wi-Fi 5G (U-NII-3)	5725 ~ 5850	5725 ~ 5850
EUT Accessory			
Adapter	Manufacturer: DONGGUAN SHILONG FUHUA ELECTRONIC CO., LTD. Model: UES24LCP-120200SPA Input: 100-240V~50/60Hz 500mA Output: 12.0V  2.0A 24.0W		
Battery	Manufacturer: Shenzhen Ryder Electronics Co., Ltd. Model: Li-18650-3.6V 3400mAh -PCM-NTC DC 3.6V, 3400mAh		
Auxiliary Test Equipment			
PC	PC Manufacturer: DELL Model: 0000053 (SN: 7ZX7R93)		
Note: 1. The EUT is sent from the applicant to Eurofins TA and the information of the EUT is declared by the applicant.			

## 2.3 Applied Standards

According to the specifications of the manufacturer, it must comply with the requirements of the following standards:

### Test standards

**FCC Code CFR47 Part15B (2024)**

**ANSI C63.4-2014**

## 2.4 Test Mode

Test Mode	
Mode 1	Adapter + Charge base + EUT +Charging

Test Type	Test Mode	Worst Mode
Radiated Emission	Mode 1	Mode 1
Conducted Emission	Mode 1	Mode 1
During the test, the preliminary test was performed in all modes, the test data of the worst-case condition was recorded in this report.		

### 3 Test Case Results

#### 3.1 Radiated Emission

##### Ambient Condition

Temperature	Relative humidity
15°C ~ 35°C	30% ~ 60%

##### Methods of Measurement

The EUT is placed on a non-metallic table 0.8m above the horizontal metal reference ground plane. The distance between EUT and receive antenna should be 10 meters below 1GHz; 3 meters for above 1GHz. During the test, the EUT was operating in its typical mode. The test method is according to ANSI C63.4-2014. Sweep the whole frequency band through the range from 30MHz to the 5th harmonic of the carrier. During the test, the height of receive antenna shall be moved from 1 to 4 meters, and the antenna shall be performed under horizontal and vertical polarization. The turn table shall be rotated from 0 to 360 degrees for detecting the maximum of radiated signal level.

The data of cable loss and antenna factor has been calibrated in full testing frequency range before the testing. During the test, the EUT is worked at maximum output power.

Set the spectrum analyzer in the following:

Below 1GHz:

RBW=100 kHz / VBW=300 kHz / Sweep=AUTO

Above 1GHz:

(a) PEAK Detector: RBW=1MHz / VBW=3MHz/ Sweep=AUTO

(b) AVERAGE Detector: RBW=1MHz / VBW=3MHz / Sweep=AUTO


The radiated emission was measured in the following position: EUT stand-up position (Z axis), lie-down position (X, Y axis). The worst emission was found in lie-down position (X axis) and the worst case was recorded.



## Test Setup


Below 1GHz

Distance 10m



Note: Area side: 21m x 12m

Above 1GHz



Note: Area side: 2.4mX3.6m

Antenna Tower meets ANSI C63.4 requirements for measurements above 1 GHz by keeping the antenna aimed at the EUT during the antenna's ascent/ descent along the antenna mast.

## Limits

### Class B

Frequency (MHz)	Field Strength (dB $\mu$ V/m)	Detector
30 -88	30.0	Quasi-peak
88-216	33.5	Quasi-peak
216 – 960	36.0	Quasi-peak
960-1000	44.0	Quasi-peak
1000-5 <sup>th</sup> harmonic of the highest frequency or 40GHz, which is lower	54 74	Average Peak

### Frequency range of radiated measurements


Highest frequency generated or used in the device or on which the device operates or tunes (MHz)	Upper frequency of measurement range (MHz)
Below 1.705	30
1.705-108	1000
108-500	2000
500-1000	5000
Above 1000	5th harmonic of the highest frequency or 40 GHz, whichever is lower.

## Test Results

Sweep the whole frequency band through the range from 30MHz to the 5th harmonic of the carrier. The Emissions in the frequency band 18GHz – 40GHz is more than 20dB below the limit are not reported.

The following graphs display the maximum values of horizontal and vertical by software.

A symbol (  $\text{dB}\mu\text{V}/\text{m}$  ) in the test plot below means (dB $\mu\text{V}/\text{m}$ )




Radiated Emission from 30MHz to 1GHz

Frequency (MHz)	Quasi-Peak (dB $\mu\text{V}/\text{m}$ )	Limit (dB $\mu\text{V}/\text{m}$ )	Margin (dB)	Height (cm)	Polarization	Azimuth (deg)	Correct Factor (dB)
41.20	18.13	40.00	21.87	102.0	V	156.00	19
66.21	16.43	40.00	23.57	102.0	V	130.00	17
166.28	14.67	43.50	28.83	102.0	V	308.00	16
298.69	38.60	46.00	7.40	102.0	H	78.00	21
311.91	22.93	46.00	23.07	101.0	H	63.00	21
957.48	29.23	46.00	16.77	102.0	H	298.00	32

Remark: 1. Correction Factor = Antenna factor + Insertion loss(cable loss+amplifier gain)

2. Margin = Limit – Quasi-Peak



Note: The signal beyond the limit is carrier.  
Radiated Emission from 1GHz to 18GHz

Frequency (MHz)	MaxPeak (dBμV/m)	Average (dBμV/m)	Limit (dB μ V/m)	Margin (dB)	Meas. Time (s)	Height (m)	Pol	Azimuth (deg)	Corr. (dB/m)
1503.63	32.45	---	74.00	41.55	1000.00	125.0	H	241.00	-12
1614.13	---	20.03	54.00	33.97	1000.00	176.0	V	255.00	-12
2472.63	36.63	---	74.00	37.37	1000.00	103.0	H	288.00	-7
2572.50	---	24.10	54.00	29.90	1000.00	224.0	V	1.00	-7
4183.25	41.15	---	74.00	32.85	1000.00	111.0	V	238.00	-3
4217.25	---	28.02	54.00	25.98	1000.00	109.0	H	293.00	-2
6767.25	---	33.80	54.00	20.20	1000.00	194.0	V	170.00	4
6767.25	46.94	---	74.00	27.06	1000.00	125.0	V	344.00	4
10462.63	49.88	---	74.00	24.12	1000.00	102.0	V	312.00	8
11085.25	---	37.54	54.00	16.46	1000.00	125.0	H	24.00	9
13943.38	55.48	---	74.00	18.52	1000.00	125.0	H	6.00	12
14383.25	---	41.62	54.00	12.38	1000.00	119.0	H	28.00	11

Remark: 1. Correction Factor = Antenna factor + Insertion loss (cable loss + amplifier gain)  
2. Margin = Limit – MaxPeak / Average

## 3.2 Conducted Emission


### Ambient Condition

Temperature	Relative humidity
15°C ~ 35°C	30% ~ 60%

### Methods of Measurement

The EUT is placed on a non-metallic table of 80cm height above the horizontal metal reference ground plane. During the test, the EUT was operating in its typical mode. The test method is according to ANSI C63.4-2014. Connect the AC power line of the EUT to the L.I.S.N. Use EMI receiver to detect the average and Quasi-peak value. RBW is set to 9 kHz, VBW is set to 30kHz. The measurement result should include both L line and N line.

### Test Setup



Note: Power Supply is AC Power source and it is used to change the voltage 120V/60Hz.

### Limits


Frequency (MHz)	Class A (dBμV)		Class B (dBμV)	
	Quasi-peak	Average	Quasi-peak	Average
0.15 - 0.5	79	66	66 to 56 *	56 to 46*
0.5 - 5	73	60	56	46
5 - 30	73	60	60	50

\*: Decreases with the logarithm of the frequency.

Note: The EUT should meet CLASS B limit.

## Test Results

Following plots, Blue trace uses the peak detection; Green trace uses the average detection.



Frequency [MHz]	QPK Level [dBμV]	QPK Limit [dBμV]	QPK Margin [dB]	AVG Level [dBμV]	AVG: CAV Limit [dBμV]	AVG Margin [dB]	Correction [dB]	Line	Meas. BW [kHz]	Meas. Time [s]
0.159	42.80	65.52	22.72	26.76	55.52	28.76	20.90	L1	9.000	1.000
0.299	29.75	60.28	30.53	17.59	50.28	32.69	20.94	L1	9.000	1.000
0.731	14.62	56.00	41.38	9.34	46.00	36.66	20.41	L1	9.000	1.000
4.070	15.33	56.00	40.67	10.01	46.00	35.99	19.42	L1	9.000	1.000
9.778	15.91	60.00	44.09	10.63	50.00	39.37	19.42	L1	9.000	1.000
29.418	21.86	60.00	38.14	17.29	50.00	32.71	19.80	L1	9.000	1.000

**Remark: Correct factor=cable loss + LISN factor**

L line

Conducted Emission from 150 kHz to 30 MHz



Frequency [MHz]	QPK Level [dBµV]	QPK Limit [dBµV]	QPK Margin [dB]	AVG Level [dBµV]	AVG: CAV Limit [dBµV]	AVG Margin [dB]	Correction [dB]	Line	Meas. BW [kHz]	Meas. Time [s]
0.164	41.40	65.28	23.89	25.16	55.28	30.12	20.90	N	9.000	1.000
0.299	29.57	60.28	30.71	17.59	50.28	32.69	20.95	N	9.000	1.000
0.780	14.64	56.00	41.36	9.20	46.00	36.80	20.36	N	9.000	1.000
4.493	15.30	56.00	40.70	9.97	46.00	36.03	19.42	N	9.000	1.000
14.372	15.40	60.00	44.60	10.13	50.00	39.87	19.51	N	9.000	1.000
29.126	18.00	60.00	42.00	11.66	50.00	38.34	19.88	N	9.000	1.000

**Remark: Correct factor=cable loss + LISN factor**

N line

Conducted Emission from 150 kHz to 30 MHz

## 5 Uncertainty Measurement

Case	Uncertainty	Factor k
Radiated Emission 30MHz – 200MHz	3.39 dB	1.96
Radiated Emission 200MHz – 1GHz	3.82 dB	1.96
Conducted Emission	2.57 dB	2



## 6 Main Test Instruments

Name of Equipment	Manufacturer	Type/Model	Serial Number	Calibration Date	Expiration Time
Radiated Emission					
EMI Test Receiver	R&S	ESCI3	100948	2025-05-07	2026-05-06
Signal Analyzer	R&S	FSV40	101186	2025-05-06	2026-05-05
TRILOG Broadband Antenna	SCHWARZBECK	VULB 9163	1023	2023-07-14	2026-07-13
Horn Antenna	SCHWARZBECK	BBHA 9120D	430	2024-07-18	2027-07-17
Amplifier	R&S	SCU18F	101022	2025-05-06	2026-05-05
Software	R&S	EMC32	9.26.01	/	/
Conducted Emission					
Artificial main network	R&S	ENV216	102191	2024-12-02	2026-12-01
EMI Test Receiver	R&S	ESR	101667	2025-05-06	2026-05-05
Software	R&S	EMC32	10.35.10	/	/
Software	R&S	EMC32	10.35.10	/	/

## ANNEX A: The EUT Appearance

The EUT Appearance are submitted separately.

## ANNEX B: Test Setup Photos

The Test Setup Photos are submitted separately.

## ANNEX C: Product Change Description

The Product Change Description are submitted separately.

\*\*\*\*\* END OF REPORT \*\*\*\*\*