



Declaration of Conformity

1009.25B0
Universal remote control



Contents

1	Country specific labeling information.....	3
1.1	Canada	3
1.2	European Union.....	4
1.3	USA	7

1 Country specific labeling information

1.1 Canada

	[fr]	[en]
ISED	Contient IC : 29667-NINAW106 L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes : 1. L'appareil ne doit pas produire de brouillage; 2. L'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.	Contains IC: 29667-NINAW106 This device contains licence-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's licence-exempt RSS(s). Operation is subject to the following two conditions: 1. This device may not cause interference. 2. This device must accept any interference, including interference that may cause undesired operation of the device.

[fr] Les produits peuvent être livrés dans les variantes suivantes / avec les composants suivants :

Numéro de produit	Nom du produit / de la marque
1009.25B0	Télécommande universelle

[en] The products can be delivered in the following variants / with the following components:

Product-No.	Product- / Trade Name
1009.25B0	Universal remote control

1.2 European Union



EU Declaration of Conformity for Medical Device

Document ID: ELO-ID: 1334573 v02

Name and Address of the Manufacturer: MAQUET GmbH
Kehler Str. 31
D-76437 Rastatt

Single Registration Number: DE-MF-000009119

On our sole responsibility, we hereby declare that the product(s)

Product- / Trade Name: Universal Remote Control

Intended Purpose: The hand-held control devices are optional operating devices for Maquet operating tables. Communication with the operating table takes place either via an interface cable or wireless.

Reference-No.: See Annex I_EN

Basic UDI-DI (acc. to Part C of Annex VI): 40467680112AB3

Classification (acc. to Annex VIII): Class I

comply with the relevant provisions of the following Regulation(s) and Directive(s):

Regulation (EU) 2017/745 on Medical Devices

Conformity Assessment Procedure: Acc. to Annex II and Annex III of Regulation (EU) 2017/745

Common Specifications used: N/A

EU Declaration of Conformity for Medical Device
ELO-ID: 1334573 v02



Directive 2011/65/EU on the restriction of the use of certain substances in electrical and electronic equipment

Directive 2014/53/EU on Radio Equipment

Conformity Assessment Procedure:	Acc. to Modul A of Directive 2014/53/EU]
Standards applied	EN 62479 EN 50566 EN 301489-1 EN 301489-17 EN 300328

This declaration of conformity is valid from date of issue until 25 May 2025.

Rastatt, 2024-07-16

Christian Burgstahler, Manager Regulatory Affairs
Signed on behalf of MAQUET GmbH

EU Declaration of Conformity for Medical Device
ELO-ID: 1334573 v02

**ANNEX I_EN**

The products can be delivered in the following variants / with the following components:

Product-No.	Product- / Trade Name	Basic UDI-DI
1009.25B0	Universal Remote Control	40467680112AB3

1.3 USA

	[en]
	Changes or modifications not expressly approved by the party responsible for compliance voids the user's authority to operate this equipment. Contains FCC ID: 2A98NNINAW106 This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures: ▪ Reorient or relocate the receiving antenna. ▪ Increase the separation between the equipment and receiver. ▪ Connect the equipment into an outlet on a circuit different from that to which the receiver is connected. ▪ Consult the dealer or an experienced radio/TV technician for help.

[en] The products can be delivered in the following variants / with the following components:

Product-No.	Product- / Trade Name
1009.25B0	Universal remote control



 **MAQUET GmbH** · Kehler Str. 31 · 76437 Rastatt · GERMANY · Phone: +49 7222 932 -0

www.getinge.com

IDC 1009.25 EN 02 2024-07-23

V02 04 23-07-2024