

TEST REPORT

Applicant: Hunan JIUYEE Medical Technology Co., Ltd

2nd floor, No. 5 plant, North zone of bonded zone,

Address: high tech zone, Hengyang City, Hunan Province,

China 421008

Equipment Type: Smart Bluetooth Hearing Aid(ITE)

Model Name: AIR C2 (refer to section 2.3)

Brand Name: JIUYEE

FCC ID: 2A9JP-AIRC1-C2R

Test Standard: 47 CFR Part 2.1093 KDB 447498 D04 v01

Sample Arrival Date: May 16, 2023

Test Date: Aug. 15, 2023 - Sep. 12, 2023

Date of Issue: Sep. 19, 2023

ISSUED BY:

Shenzhen BALUN Technology Co., Ltd.

Tested by: Zhang Jiwei Checked by: Xu Rui Approved by: Tolan Tu

(Testing Director)

Tolan lu

2hang Jiwei Xu Rui

Tel: +86-755-66850100 E-mail: qc@baluntek.com Page No. 1/10



Revision History

VersionIssue DateRevisions ContentRev. 01Aug. 31, 2023Initial Issue

Rev. 02 Sep. 19, 2023 1, Update FCC ID;

2. Update test value;

3. Update the EUT Name

TABLE OF CONTENTS

1	GENER	AL INFORMATION	3
	1.1	Test Laboratory	3
	1.2	Test Location	3
2	PRODU	JCT INFORMATION	4
	2.1	Applicant Information	4
	2.2	Manufacturer Information	4
	2.3	General Description for Equipment under Test (EUT)	4
	2.4	Ancillary Equipment	4
	2.5	Technical Information	5
3	SUMMA	ARY OF TEST RESULT	6
	3.1	Test Standards	6
4	DEVICE	CATEGORY AND LEVELS LIMITS	7
5	ASSES	SMENT RESULT	9
	5.1	Output Power	9
	5.2	Tune-up power	9
	5.3	RF Exposure Evaluation Result	9
	5.4	Conclusion	9



1 GENERAL INFORMATION

1.1 Test Laboratory

Name	Shenzhen BALUN Technology Co., Ltd.		
A ddraga	Block B, 1/F, Baisha Science and Technology Park, Shahe Xi Road,		
Address	Nanshan District, Shenzhen, Guangdong Province, P. R. China		
Phone Number	+86 755 6685 0100		

1.2 Test Location

Name	Shenzhen BALUN Technology Co., Ltd.		
	☑ Block B, 1/F, Baisha Science and Technology Park, Shahe Xi		
	Road, Nanshan District, Shenzhen, Guangdong Province, P. R.		
Logotion	China		
Location	☐ 1/F, Building B, Ganghongji High-tech Intelligent Industrial Park,		
	No. 1008, Songbai Road, Yangguang Community, Xili Sub-district,		
	Nanshan District, Shenzhen, Guangdong Province, P. R. China		
Accreditation	The laboratory is a testing organization accredited by FCC as a		
Certificate	accredited testing laboratory. The designation number is CN1196.		



2 PRODUCT INFORMATION

2.1 Applicant Information

Applicant	Hunan JIUYEE Medical Technology Co., Ltd
A dalara a a	2nd floor, No. 5 plant, North zone of bonded zone, high tech zone,
Address	Hengyang City, Hunan Province, China 421008

2.2 Manufacturer Information

	Manufacturer	Hunan JIUYEE Medical Technology Co., Ltd		
	A ddraga	2nd floor, No. 5 plant, North zone of bonded zone, high tech zone,		
Add	Address	Hengyang City, Hunan Province, China 421008		

2.3 General Description for Equipment under Test (EUT)

EUT Name	Smart Bluetooth Hearing Aid(ITE)		
Model Name Under Test	AIR C2		
Series Model Name	AIR C1		
Description of Model	The model changed for different market and customer, the others are		
name differentiation	the same. (this information provided by the applicant)		
Hardware Version	V2.2		
Software Version	V1.0.3		
Dimensions (Approx.)	N/A		
Weight (Approx.)	N/A		

2.4 Ancillary Equipment

	Battery		
	Brand Name	MICpower	
	Model No.	M1054S3-Q03	
Ancillary Equipment 1	Serial No.	N/A	
	Capacity	48 mAh	
	Rated Voltage	3.85 V	
	Limit Charge Voltage	4.4 V	

Report No.: BL-SZ2380058-701



2.5 Technical Information

Network and Wireless	Bluetooth (BR+EDR+BLE)
connectivity	Bidetootii (BR+EDR+BLE)

The requirement for the following technical information of the EUT was tested in this report:

Operating Mode	Bluetooth		
Frequency Range	Bluetooth	2400 ~ 2483.5 MHz	
Antenna Type	Bluetooth	FPC Antenna	
Exposure Category	General Population/Uncontrolled Exposure		
Product Type	Portable Device		

Report No.: BL-SZ2380058-701



3 SUMMARY OF TEST RESULT

3.1 Test Standards

No. Identity Document Title		Document Title
1 47 CFR Part 2.1093 Radiofrequency radiation exposure evaluation: portable		Radiofrequency radiation exposure evaluation: portable devices
2 KDB 447498 D04 v01 KDB 447498 D04 Interim General		KDB 447498 D04 Interim General RF Exposure Guidance v01



4 DEVICE CATEGORY AND LEVELS LIMITS

Portable Derives:

CFR Title 47 §2.1093(b)

(b) For purposes of this section, a portable device is defined as a transmitting device designed to be used so that the radiating structure(s) of the device is/are within 20 centimeters of the body of the user.

FCC KDB 447498 Derives:

According with FCC KDB 447498 D04, Appendix B, The SAR-based exemption formula applies for single fixed, mobile, and portable RF sources with available maximum time-averaged power or effective radiated power (ERP), whichever is greater, of less than or equal to the threshold Pth (mW).

This method shall only be used at separation distances from 0.5 cm to 40 cm and at frequencies from 0.3 GHz to 6 GHz (inclusive). The following table shows the power threshold from 5mm to 50mm.

Power Thresholds (mW)					
Fraguency	At separation				
Frequency	distance of				
(MHz)	≤5 mm	10 mm	15 mm	20 mm	25 mm
300	39 mW	65 mW	88 mW	110 mW	129 mW
450	22 mW	44 mW	67 mW	89 mW	112 mW
835	9 mW	25 mW	44 mW	66 mW	90 mW
1900	3 mW	12 mW	26 mW	44 mW	66 mW
2450	3 mW	10 mW	22 mW	38 mW	59 mW
3600	2 mW	8 mW	18 mW	32 mW	49 mW
5800	1 mW	6 mW	14 mW	25 mW	40 mW
	At separation				
Frequency	distance of				
(MHz)	30 mm	35 mm	40 mm	45 mm	50 mm
300	148 mW	166 mW	184 mW	201 mW	217 mW
450	135 mW	158 mW	180 mW	203 mW	226 mW
835	116 mW	145 mW	175 mW	207 mW	240 mW
1900	92 mW	122 mW	157 mW	195 mW	236 mW
2450	83 mW	111 mW	143 mW	179 mW	219 mW
3600	71 mW	96 mW	125 mW	158 mW	195 mW
5800	58 mW	80 mW	106 mW	136 mW	169 mW



Note:

- Maximum power is the source-based time-average power and represents the maximum RF output power including tune-up tolerance among production units
- 2. Per KDB 447498 D04, for larger devices, the test separation distance of adjacent edge configuration is determined by the closest separation between the antenna and the user.
- 3. Per KDB 447498 D04, standalone SAR test exclusion threshold is applied; If the distance of the antenna to the user is < 5mm, 5mm is used to determine SAR exclusion threshold
- 4. Per KDB 447498 D04, for separation distances from 0.5 cm to 40 cm and at frequencies from 0.3 GHz to 6 GHz (inclusive), the threshold Pth (mW) is given by Following:

$$P_{th}(mW) = \begin{cases} ERP_{20cm}(d/20cm)^x & d \leq 20cm \\ ERP_{20cm} & 20cm \leq d \leq 40cm \end{cases}$$

where

$$x = -log_{10} \left(\frac{60}{ERP_{20cm} \sqrt{f}} \right)$$

- a. f(GHz) is the RF channel transmit frequency in GHz
- b. d is the separation distance (cm), The result is rounded to one decimal place for comparison
- c. ERP_{20cm} are determined by:

$$ERP_{20cm}(mW) = f(x) = \begin{cases} 2040f & 0.3GHz \le f < 1.5GHz \\ 3060 & 1.5GHz \le f \le 6GHz \end{cases}$$



ASSESSMENT RESULT

5.1 Output Power

Bluetooth				
Mode BR+EDR BLE				
Peak Power (dBm)	1.86	0.05		
Antenna Gain (dBi)	-1.85	-1.85		
EIRP (dBm)	0.01	-1.80		
Note: This table listed the worst case power value, please refer to BL-SZ2380058-601~602 report for more details.				

5.2 Tune-up power

Mode	Conducted Power Range (dBm)	EIRP Range (dBm)	ERP Range (dBm)	
Bluetooth	[0.00,2.00]	[-1.85,0.15]	[-4.00,-2.00]	

Note1: ERP= EIRP -2.15dB

Note2: According KDB 447498 D04, used the greater of maximun conducted power and ERP to compare with the threshold

value Pth.

5.3 RF Exposure Evaluation Result

Mode	Distance (mm)	Calculation Frequency (GHz)	Tune-up limit power (dBm)	Tune-up limit power (mW)	Threshold Value (mW)	Verdict
Bluetooth	5	2.48	2.00	1.58	2.72	Compliance

5.4 Conclusion

This EUT is deemed to comply with the reference level limits, therefore the basic restrictions are compliant with human exposure limits.



Statement

- 1. The laboratory guarantees the scientificity, accuracy and impartiality of the test, and is responsible for all the information in the report, except the information provided by the customer. The customer is responsible for the impact of the information provided on the validity of the results.
- 2. The report without China inspection body and laboratory Mandatory Approval (CMA) mark has no effect of proving to the society.
- 3. For the report with CNAS mark or A2LA mark, the items marked with "☆" are not within the accredited scope.
- 4. This report is invalid if it is altered, without the signature of the testing and approval personnel, or without the "inspection and testing dedicated stamp" or test report stamp.
- 5. The test data and results are only valid for the tested samples provided by the customer.
- 6. This report shall not be partially reproduced without the written permission of the laboratory.
- 7. Any objection shall be raised to the laboratory within 30 days after receiving the report.

--END OF REPORT--