

eBELTER

Declaration letter for model difference

We:

Shenzhen Belter Health Measurement and Analysis Technology Co., Ltd

702,704,block C, tsinghua unis science park, hi-tech industrial park north, nanshan Shenzhen
China

Officially notify TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch that the Models ePA-09B4, PA-10, IVBPA have the same technical construction including circuit diagram, PCB Layout, components and component layout, all electrical construction and mechanical construction.

The difference are the motherboard compared with IVBPA is to remove LED display and display driver for ePA-09B4, and the main board compared with IVBPA is to remove the Bluetooth transmission part for PA-10.

Applicant: Shenzhen Belter Health Measurement and Analysis Technology Co., Ltd

2015.7.7
(Date)

