

FCC CLASS II PERMISSIVE CHANGE REQUEST LETTER

Reason for Amendment (current / obsolete)	Revision History		Approved Date
	From	To	
Initial Release (current)	1.0	1.0	Feb-20-2012

FCC Class II Permissive Change Request Letter

Date: 7/9/2013

To FCC:

RE: FCC Permissive II Change Request for Company: Abbott Point of Care FCC ID: 2AAEX-SDABGN

We are submitting an application for a class II permissive change to the FCC approval of the Company name: Abbott Point of Care, product description: SDIO Wireless Module (FCC: **2AAEX-SDABGN**, Original Grant Date: 07/08/2013). The transmitter module itself has not changed. Here are the changes:

- 1) adding portable host - DragonFly Hand-held Blood Analyzer Unit
- 2) BT will be disabled by software - no hardware changes

Sincerely,



Joseph Kiceina / Director, Regulatory Affairs
400 College Road East, Princeton, NJ 08540
joseph.kiceina@apoc.abbott.com
PH: 609-454-9494