

Innokas Yhtymä Oy Vihikari 10 FI-90440 Kempele FINLAND +358 8 562 3100 FAX +358 8 562 3111

FCC IDENTIFIER:

2ABKEVC150

PRODUCT:

VC150

To whom it may concern

Innokas Yhtyma Oy is applying updated FCC Grant with U-NII-3 Grant Note Code 38 for a patient monitor. The device is Carescape VC150 Vital Signs Monitor.

The enclosed documents constitute a formal submittal and application for a Class II Permissive change for the patient monitor pursuant to the following rules:

Subpart E of Part 15 of FCC Rules (CFR 47), UNII Devices

This permissive change is to remove note 49 on the grant dated 06/10/2016 based on the results of new testing in accordance with \$15.407(b)(4)(i).

There have been no changes to Carescape VC150 Vital Signs Monitor impacting RF features and the old test results are still valid. The new testing confirms that all requirements of the new rules are met. Devices under the old rules have not been manufactured, marketed or sold after 1-JUN-2016.

The device is serial number controlled. Innokas Yhtyma Oy can distinguish between devices shipped under the new rules from the serial number

Sincerely yours,

Kempele, February 19th, 2020

Kotipalo

Innokas Yhtymä Oy

Tiina Kotipalo

Tima

Director of Quality and Regulatory Affairs