



RF Exposure

Test Requirement:	FCC 47CFR 15.247(i)
Product:	Upper Arm Blood Pressure Monitor
Brand Name:	Braun
Model No.:	BUA4075BUS
FCC ID:	2ABRGBUA4075
Test Date:	2025-06-18
Mode of Operation:	BT Tx mode

Requirements:

In 15.247(i), an equipment shall be operated in a manner that ensures that the public is not exposed to radio frequency energy levels in excess of the limits in §§ 1.1310 and 2.1093 of this chapter.

Applications to the Commission for construction permits, licenses to transmit or renewals thereof, equipment authorizations or modifications in existing facilities must contain a statement confirming compliance with the limits unless the facility, operation, or transmitter is categorically excluded, as discussed below. Technical information showing the basis for this statement must be submitted to the Commission upon request.

According to KDB447498 D04 General RF Exposure Guidance v01, Appendix B

B.2 Blanket 1 mW Blanket Exemption

The 1 mW Blanket Exemption of § 1.1307(b)(3)(i)(A) applies for single fixed, mobile, and portable RF sources with available maximum time-averaged power of no more than 1mW, regardless of separation distance.

Result: PASS

The maximum time-averaged power = 0.291mW (at frequency = 2402.0MHz) < 1mW

Remark: maximum time-averaged power = -5.361dBm = 0.291mW