

FCC ID: 2ABS8CBM0114IL32

D.L.S. ELECTRONIC SYSTEMS, INC.
166 South Carter Street
Genoa City, WI 53128

July 2, 2014 revised August 8, 2014

FCC Authorization & Evaluation Division
7435 Oakland Mills Road
Columbia, Maryland 21046

Gentlemen:

We are submitting for the **Rehabilitation Institute of Chicago**, the **Complete Controller, Model Number: CCDB110**, with an **FCC ID: 2ABS8CBM0114IL32** for **Certification with Single Modular Approval**. The Rehabilitation Institute of Chicago is purchasing an existing module with FCC modular approval, but using it within 20cm of the human body. As this does not conform to the grant notes of the module they are purchasing, the Rehabilitation Institute of Chicago is certifying the module themselves.

The original module has been tested and complies to **Part 15, Subpart C**. If there are any questions in regard to this application, please call D.L.S. Electronic Systems, Inc. at (262) 279-0210. Enclosed you will find a copy of the original test report with the information and results of tests that is required under Part 15, Section 15.247.(Further specifics for the original module, including the model numbers included in the existing certification, can be found in the Operational Descriptions.)

Thank you for your assistance in this matter.

Sincerely,



William M. Stumpf
DLS/js