



Surgiris  
80, rue de la gare  
59170 CROIX  
FRANCE

To: Curtis-Straus LLC  
Certification Department  
527 Great Road  
Littleton, MA 01460

### Limited Modular approval letter

Dear application examiner,

Surgiris, would like to have you authorization as a (Limited) modular approval on the SCY00-040 module, with FCC ID: 2AC7OCTRL000. The requirements of CFR 47 part 15.212 have been checked and shown in the following statement,

Modular transmitter requirement	Yes	No	Comment
(i) The radio elements of the modular transmitter must have their own shielding. The physical crystal and tuning capacitors may be located external to the shielded radio elements.	X		The SCY00-040 module will only be installed in SURGIRIS end products. Some control measurements are performed on host equipment to ensure the compliance for both intentional and unintentional emission requirements.
(ii) The modular transmitter must have buffered modulation/data inputs (if such inputs are provided) to ensure that the module will comply with Part 15 requirements under conditions of excessive data rates or over-modulation.	X		
(iii) The modular transmitter must have its own power supply regulation.	X		
(iv) The modular transmitter must comply with the antenna and transmission system requirements of Sections 15.203, 15.204(b) and 15.204(c). The antenna must either be permanently attached or employ a "unique" antenna coupler (at all connections between the module and the antenna, including the cable). The "professional installation" provision of Section 15.203 is not applicable to modules but can apply to limited modular approvals under paragraph (b) of this section.	X		
(v) The modular transmitter must be tested in a stand-alone configuration, i.e., the module must not be inside another device during testing for compliance with Part 15 requirements....	X		
(vi) The modular transmitter must be equipped with either a permanently affixed label or must be capable of electronically displaying its FCC identification number.	X		
If the FCC ID is not visible when the module is installed inside another device, then the outside of the device into which the module is installed must also display a label referring to the			

enclosed module. This exterior label can use wording such as the following: "Contains Transmitter Module FCC ID: xxxxx" or "Contains FCC ID: xxxxxxxxxxxx." The exact wording is not specified in our Rules (since modules are not specifically addressed), so you may use similar wording which expresses the same meaning.			
(vii) The modular transmitter must comply with any specific rules or operating requirements that ordinarily apply to a complete transmitter and the manufacturer must provide adequate instructions along with the module to explain any such requirements. A copy of these instructions must be included in the application for equipment authorization.	X		
(viii) The modular transmitter must comply with any applicable RF exposure requirements in its final configuration.	X		

Best regards,

**S.A.S. SURGIRIS**  
 80, rue de la Gare  
 59170 CROIX  
 Tél. 03 20 16 07 93  
 Fax 03 20 95 46 19  
 N° SIRET 437 760 512 00044  
 TVA FR 20 437 760 812  
