

SBO Hearing A/S

Federal Communication Commission

Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau

Innovation, Science and Economic Development Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2, Canada

June 25th, 2025

Declaration for IC and FCC Label Exemption for CL_AU5_BTEMPP hearing aid module

With reference to Industry Canada, RSP-100 issue 12, chapter 5 Labelling requirements and KDB 784748 D01v09r02 GUIDELINES FOR LABELLING AND USER INFORMATION FOR DEVICES SUBJECT TO PART 15 AND PART 18, we, SBO Hearing A/S hereby declare an exemption regarding the labelling of the IC and FCC certification number on the CL_AU5_BTEMPP wireless hearing aid families.

The SBO Hearing A/S IC and FCC certification number: **11936A-AU5BTEMPP** and **2ACAH-AU5BTEMPP** are located next to the device model number in the user manual and on the product packaging material. The exemption declaration is based on the fact that the hearing aids will be too small (behind ear product size max. 30.0 x 14.0 x 8.0 mm exclusive acoustic tube) for the IC and FCC certification number to be legible without the aid of magnification while printed on a surface of the hearing aid area not exposed to wear and user functions.

For your information, our TCB is Cetecom Advanced GmbH in Germany.

Signed for
SBO Hearing A/S



Per Klaus Nielsen
Principal Certification Specialist, Regulatory Affairs
Email: pkni@sbohearing.com