

Date: September 30, 2019  
Federal Communications Commission  
Authorization and Evaluation Devision  
7435 Oakland Mills Road  
Columbia, MD 21046

Attn: OET Dept.

Ref: FCC Class II Permissive change for FCC ID:2ACHK-01070189  
Original module Grant Date: 09/10/2018  
The five C2PC module Grant Date: 09/04/2019  
Applicant: iRay Technology Co. Ltd.

Dear Examiner,

This is to request a Class II permissive change for FCC ID: 2ACHK-01070189, originally granted on 09/10/2018. The five C2PC Granted on 09/04/2019.

The major change filed under this application is:

1. Please refer to the below table.
1. There were no hardware changes to the module
2. There were no other hardware and software changes to the host

Different	Prototype	Variant
Model	Mars1417V-TSI	Focus 35C / TRIMAX 35C
Product name	Wireless Digital Flat Panel Detector	Detector
Brand name	/	Carestream®/TRIMAX®
Color	White 	Grey 
DC input port	3PIN+4PIN 	3PIN 

Label	<p><b>Wireless Digital Flat Panel Detector</b></p> <p>Model: Mars1417V-TSI</p> <p>Power: Adapter Port Input 24V=0.75A Battery Port Input 11.5V=1.5A iRay Technology Taicang Ltd. No.33 Xinggang Road, Taicang Port Economic and Technological Development Zone, Jiangsu, China 200X-XX www.iraygroup.com</p> <p><b>EC REP</b> iRay Europe GmbH In den Dorfwiesen 14, 71720 Oberstenfeld Germany</p> <p><b>FC</b> Contains FCC ID: 2ACHK-01070189 Contains IC: 25116-01070189</p> <p><b>SN</b></p>  <p><b>Focus 35C Detector Rx only</b> <b>Focus 35C 系统探测器</b></p> <p><b>MODEL</b> 型号: Focus 35C <b>SC</b> 2169</p> <p><b>SN</b></p> <p><b>CMIT ID:</b> Adapter Port Input: 24 VDC, 0.75 A 适配器输入: 直流24伏特, 0.75安培 Battery Port Input: 11.5 VDC, 1.5 A 电池输入: 直流11.5伏特, 1.5安培</p> <p><b>Manufactured in China for:</b> 生产企业名称: Carestream Health, Inc. 企业注册地址: 150 Verona Street, Rochester, NY, USA 14608</p> <p><b>CE</b> 0197 <b>IPX1</b></p> <p>P/N AJ4287 REV A</p>	<p><b>Focus 35C Detector Rx only</b> <b>Focus 35C 系统探测器</b></p> <p><b>MODEL</b> 型号: Focus 35C <b>SC</b> 2169</p> <p><b>SN</b></p> <p><b>CMIT ID:</b> Adapter Port Input: 24 VDC, 0.75 A 适配器输入: 直流24伏特, 0.75安培 Battery Port Input: 11.5 VDC, 1.5 A 电池输入: 直流11.5伏特, 1.5安培</p> <p><b>Manufactured in China for:</b> 生产企业名称: Carestream Health, Inc. 企业注册地址: 150 Verona Street, Rochester, NY, USA 14608</p> <p><b>CE</b> 0197 <b>IPX1</b></p> <p>P/N AJ4287 REV A</p>	<p><b>TRIMAX 35C Detector Rx only</b> <b>TRIMAX 35C 系统探测器</b></p> <p><b>MODEL</b> 型号: TRIMAX 35C <b>SC</b> 2173</p> <p><b>SN</b></p> <p><b>CMIT ID:</b> Adapter Port Input: 24 VDC, 0.75 A 适配器输入: 直流24伏特, 0.75安培 Battery Port Input: 11.5 VDC, 1.5 A 电池输入: 直流11.5伏特, 1.5安培</p> <p><b>Manufactured in China for:</b> 生产企业名称: Carestream Health, Inc. 企业注册地址: 150 Verona Street, Rochester, NY, USA 14608</p> <p><b>CE</b> 0197 <b>IPX1</b></p> <p>P/N AJ4291 REV A</p>
-------	--	--	---

Sincerely yours,



Name: Wei Pan  
Title: RA manager  
Company: iRay Technology Co. Ltd.