



**Dentsply
Sirona**

DAC Universal
Dental reprocessing device




Sirona Dental Systems GmbH
Fabrikstrasse 31
64625 Bensheim
Germany

2019-02-22
Type: MK IV

REF 6633213
SN 0001021

Pressure vessel:

Pressurised air supply 5-8 bar
Max working pressure 3,5 bar
Max working temperature 137 °C
Chamber volume <2 liter
Main supply 100 - 127 V AC
220 - 240 V AC
50/60 Hz
1300 W



0123

+E27666332131/\$\$+70001021/16D20190222%

FCC ID: 2AD7W-DACUNIMK4 IC: 12730A-DACUNIMK4

This device complies with part 15 of the FCC rules.

Operation is subject to the following two conditions: (1)

This device may not cause harmful interference, and (2)
this device must accept any interference received, including
interference that may cause undesired operation.