



**F2 Labs**  
16740 Peters Road  
Middlefield, Ohio 44062  
United States of America  
[www.f2labs.com](http://www.f2labs.com)

## **CERTIFICATION TEST REPORT**

**Manufacturer:** Centurion Medical Products  
100 Centurion Way  
Williamston, Michigan 48895 USA

**Applicant:** Same As Above

**Product Name:** Compass Monitor Accessory Plug (MAP)

**Product Description:** The Centurion Compass CAST-MAP system consists of two parts. The Compass CAST disposable digital wireless blood pressure monitor (transmitter) and the Compass MAP wireless sensor (receiver) which plugs into the Blood Pressure monitor port of a medical multifunction monitor. When a connection is made between the CAST and the MAP, a detailed blood pressure waveform is presented on the multi-function monitor. The Compass CAST and MAP are both implemented using Low Energy Bluetooth radios.

**Model:** CWM001 (Compass MAP)

**FCC ID:** 2AF4Z-CWM001

**Testing Commenced:** Nov. 10, 2015

**Testing Ended:** Nov. 11, 2015

**Test Results:** In Compliance

The EUT complies with the EMC requirements when manufactured identically as the unit tested in this report, including any required modifications. Any changes to the design or build of this unit subsequent to this testing may deem it non-compliant.

**Standards:**

- KDB447498



Order Number: F2LQ7746C

Manufacturer: Centurion Medical Products  
Model: CWM001 (Compass MAP)

**Evaluation Conducted by:**

Ken Littell, Director of EMC & Wireless Operations

**Report Reviewed by:**

Wendy Fuster, President

F2 Labs  
26501 Ridge Road  
Damascus, MD 20872  
Ph 301.253.4500  
Fax 301.253.5179

F2 Labs  
16740 Peters Road  
Middlefield, OH 44062  
Ph 440.632.5541  
Fax 440.632.5542

This test report may be reproduced in full; partial reproduction only may be made with the written consent of F2 Labs. The results in this report apply only to the equipment tested.



Order Number: F2LQ7746C

Manufacturer: Centurion Medical Products

Model: CWM001 (Compass MAP)

## TABLE OF CONTENTS

Section	Title	Page
1	ADMINISTRATIVE INFORMATION	4
2	SUMMARY OF TEST RESULTS/MODIFICATIONS	5
3	ENGINEERING STATEMENT	6
4	EUT INFORMATION AND DATA	7
5	RF EXPOSURE FOR DEVICE >20cm FROM HUMAN	8



Order Number: F2LQ7746C

Manufacturer: Centurion Medical Products

Model: CWM001 (Compass MAP)

## 1 ADMINISTRATIVE INFORMATION

### 1.1 Measurement Location:

F2 Labs in Middlefield, Ohio. Site description and attenuation data are on file with the FCC's Sampling and Measurement Branch at the FCC Laboratory in Columbia, MD.

### 1.2 Measurement Procedure:

All measurements were performed according to KDB558074.

### 1.4 Document History

Document Number	Description	Issue Date	Approved By
F2LQ7746C-11E	First Issue	Feb. 11, 2016	W. Fuster



Order Number: F2LQ7746C

Manufacturer: Centurion Medical Products  
Model: CWM001 (Compass MAP)

## 2 SUMMARY OF TEST RESULTS

Test Name	Standard(s)	Results
RF Exposure for Device >20cm from Human	KDB447498	Complies
Exemption for 5mm distance from Human	KDB447498	Complies

Modifications Made to the Equipment
None



**Order Number: F2LQ7746C**

**Manufacturer: Centurion Medical Products**

**Model: CWM001 (Compass MAP)**

### **3      ENGINEERING STATEMENT**

This report has been prepared on behalf of Centurion Medical Products to provide documentation for the testing described herein. This equipment has been tested and found to comply with KDB447498. The test results found in this test report relate only to the item(s) tested.



Order Number: F2LQ7746C

Manufacturer: Centurion Medical Products

Model: CWM001 (Compass MAP)

## 4 EUT INFORMATION AND DATA

### 4.1 Equipment Under Test:

Product: Compass Monitor Accessory Plug (MAP)  
Model: CWM001

Serial No.: None Spec.  
FCC ID: 2AF4Z-CWM001

### 4.2 Trade Name:

Centurion Medical Products

### 4.3 Power Supply:

5VDC from an AC/DC power supply

### 4.4 Applicable Rules:

- KDB447498

### 4.5 Equipment Category:

Radio Transmitter-DTS

### 4.6 Antenna:

0dBi Integral Antenna

### 4.7 Accessories:

N/A

### 4.8 Test Item Condition:

The equipment to be tested was received in good condition.



Order Number: F2LQ7746C

Manufacturer: Centurion Medical Products

Model: CWM001 (Compass MAP)

## 5. RF EXPOSURE FOR DEVICE >20cm FROM HUMAN

### 5.1 Requirements:

Limit: 1mW/cm<sup>2</sup>

Formula used for result: 
$$\frac{\text{E.I.R.P.}}{4 \pi R^2}$$

Results: E.I.R.P. = 6.06mW

6.06mW at the 2402 MHz High Channel  
which is the highest.

$$\frac{6.06\text{mW}}{4 \pi R^2} = \frac{6.06\text{mW}}{5026.55} = .0012\text{mW/cm}^2$$

**NOTE: The EUT meets the SAR Test Exclusion for use of 5mm distance since EIRP of 6.06mW is under the 10mW requirement for exclusion.**