


Instruction Manual

Thank you for purchasing our products.
Please read the instructions carefully before using it.


Detail Specs



SenseON

Detail Specs


Sensor: Clinical Grade Cardio-Pulmonary Sensor
 Battery Type: 3.7V rechargeable Lithium battery
 Battery Life: 2hrs per day up to 7 days
 Connection: Bluetooth 4.0
 Sticker Materials: Athletic Kinesio Tape
 Body Materials: Food Grade Silicon
 Waterproof Grade: IP67, Dust Tight & Water Immersion Proof



Features

Cardio-Pulmonary Function Tracking

Clinical grade Cardio-Pulmonary function sensor for accurate heart and respiratory rate monitoring.



Features


Rich sports biometric data

Professional heart and respiration function metrics and other quantitative workout data.

Features

Personalized smart sport assistant


Formulated exercise intensity and quantity are recommended to make it more effective, healthier and happier for athletic training, physical fitness or health recovery.



Features

Convenient and comfortable experience

Using of replaceable athletic Kinesio Tape ensures great comfort for several hours without peeling-off.




Warranty card

Product model	
Product number	
Purchase date	
Customer name	


Warranty record

1. Please provide this card when repairing or replacement needed.
 2. Free repairing or replacement are provided for any malfunction resulted from poor quality of original components or manufacturing in one year since the product is purchased.

Get the APP




Use steps




① Put SenseON on the adhesive tape

Use steps




② Tear off release paper of adhesive tape

Use steps



③ Stick SenseON on the left chest

Use steps



④ Double tap on SenseON

Company Information

Brand: CardioCycle
 Product name: SenseON
 Design by CardioCycle in Shenzhen, Assembled in China
 Company: Shenzhen CardioCycle Technology Co., Ltd.
 Company Address: 4228 Building 4, Shenzhen International Software Park, Nanshan District, Shenzhen, P.R. China
 Work time: Mon to Fri 09:30-18:00
 Web site: www.cardiocycle.net

Federal Communication Commission (FCC) Radiation Exposure Statement

Power is so low that no RF exposure calculation is needed.

FCC statements:

This device complies with part 15 of the FCC rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

NOTE: The manufacturer is not responsible for any radio or TV interference caused by unauthorized modifications or changes to this equipment. Such modifications or changes could void the user's authority to operate the equipment.

NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.