



February 24,2016

WLL Report #14358-02 Rev 1

MC10 INC.  
10 Maguire Rd.  
Building #3, First Floor  
Lexington, MA 02421

Model: MC10 Inc. BioStampRC Sensor

FCC ID: 2AGYUBRCS01, IC ID: 6648A-ER84

Washington Laboratories, Ltd. performed a SAR Evaluation on the MC10 Inc. BioStampRC Sensor. This product is a portable transmitter, operating between 2402MHz and 2480MHz using Bluetooth LE technology. It is designed to be worn by the user with a separation distance  $\leq$  5mm and therefore SAR must be considered.

#### FCC Exclusion Results:

The EUT was tested under FCC Part 15.247 as a DTS device and the maximum measured power was determined as 0.55mW(-2.6dBm) radiated EIRP, The Production tolerance is  $\pm$ 1dB for a total maximum output of 0.69mW (-1.6dBm), This will be rounded to 1mW per FCC Guidance.. Documentation of the measurement technique used can be found in Washington Laboratory Report 14358-01.

Referring to the FCC Document, 447498 D01 General RF Exposure Guidance v05r01, Mobile and Portable Devices RF Exposure Procedures and Equipment Authorization Policies, section 4.1 allows for the exclusion of SAR Testing if the device meets the following requirements:

$$[(\text{max. power of channel, including tune-up tolerance, mW}) / (\text{min. test separation distance, mm})] \cdot [\sqrt{f(\text{GHz})}] \leq 3.0 \text{ for 1-g SAR and } \leq 7.5 \text{ for 10-g extremity SAR}$$

Where,

- $f(\text{GHz})$  is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation
- The result is rounded to one decimal place for comparison
- When the minimum *test separation distance* is  $< 5$  mm, a distance of 5 mm is used

Using the above formula and test data for the highest power channel:

$$(1 \div 5) \times \sqrt{2.402} = 0.3 < \leq 3.0$$

$$(1 \div 5) \times \sqrt{2.442} = 0.3 < \leq 3.0$$

$$(1 \div 5) \times \sqrt{2.480} = 0.3 < \leq 3.0$$

## IC Exclusion Results from RSS102 issue 5

### Exemption Limits for Routine Evaluation – SAR Evaluation

SAR evaluation is required if the separation distance between the user and/or bystander and the antenna and/or radiating element of the device is less than or equal to 20 cm, except when the device operates at or below the applicable output power level (adjusted for tune-up tolerance) for the specified separation distance defined in Table 1.

**Table 1: SAR evaluation – Exemption limits for routine evaluation based on frequency and separation distance<sup>4,5</sup>**

Frequency (MHz)	Exemption Limits (mW)				
	At separation distance of <b>≤5 mm</b>	At separation distance of <b>10 mm</b>	At separation distance of <b>15 mm</b>	At separation distance of <b>20 mm</b>	At separation distance of <b>25 mm</b>
≤300	71 mW	101 mW	132 mW	162 mW	193 mW
450	52 mW	70 mW	88 mW	106 mW	123 mW
835	17 mW	30 mW	42 mW	55 mW	67 mW
1900	7 mW	10 mW	18 mW	34 mW	60 mW
2450	4 mW	7 mW	15 mW	30 mW	52 mW
3500	2 mW	6 mW	16 mW	32 mW	55 mW
5800	1 mW	6 mW	15 mW	27 mW	41 mW

If the frequency falls between two levels a linear interpolation shall be used:

For the EUT frequencies in question the Limits are as follows (as the conducted levels are higher than the radiated levels , conducted levels shall be used), Tune-up tolerance =±1dB:

Frequency	Limit	EUT Levels (Conducted levels)
2.402GHz	4.2mW	0.87mW
2.442GHZ	4.0mW	1.10mW
2.480GHz	3.9mW	1.23mW

The EUT fulfills the exclusion requirements for both FCC and Industry Canada.

Therefore, the EUT does not require SAR Testing.

Testing was performed on an Open Area Test Site (OATS) of Washington Laboratories, Ltd, 7560 Lindbergh Drive, Gaithersburg, MD 20879. Site description and site attenuation data have been placed on file with the FCC's Sampling and Measurements Branch at the FCC laboratory in Columbia, MD. Washington Laboratories, Ltd. has been accepted by the FCC and approved by ACLASS under Certificate AT-1448 as an independent FCC test laboratory.

These tests are accredited and meet the requirements of ISO/IEC 17025 as verified by the ANSI-ASQ National Accreditation Board/ACCLASS. Refer to certificate and scope of accreditation AT-1448.

Should you have any questions, please do not hesitate to contact us.

Sincerely,

A handwritten signature in blue ink, appearing to read "James Ritter".

James Ritter  
Compliance Engineer

A handwritten signature in blue ink, appearing to read "Steven D. Koster".

Steven D. Koster  
President