

August 17, 2016

TUV SUD BABT FCB Octagon House, Segensworth Road, Fareham, Hampshire, PO15 5RL

Attention: Director of Certification RE: SAR Exclusion for Portable Device

FCC ID: 2AH33373100

Hereby declare that the NantHealth GlowCap Device which is defined as a portable device under the Code of Federal Regulations 47 CFR paragraph 2.1093 is categorically excluded from routine evaluation of radiofrequency exposure evaluation for portable device. The device satisfies the SAR Exclusion Threshold as per Clause 4.3.1 of KDB 447498 D01v06 using the formula:

$$3.0 \geq \left(\frac{\text{max power in mW}}{\text{min. test separation distance in mm}}\right) \left(\sqrt{f_{(GHz)}}\right)$$

$$3.0 \ge \left(\frac{2.65 \text{ mW}}{(5 \text{ mm})}\right) \left(\sqrt{2.402}\right)$$

 $3.0 \ge 0.82$

Sincerely,

Alex Cilai

Authorized Signatory

Title: EMC/Wireless Test Engineer