



Date: May 19, 2020

Attn: Reviewing Engineer  
Federal Communications Commission  
7435 Oakland Mills Road  
Columbia, MD 21046

To Whom It May Concern:

The TytoCare Ltd., Tyto Device G2, FCC ID: 2AHKN-5G2, and label are so small that it cannot include the statement specified in FCC Part 15 Sub-part A, Section 15.19 (a) (3) "This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference and
- (2) This device must accept any interference received, including interference that may cause undesired operation."

Due to this reason, we are using the provisions of Part 15 Sub-part A, Section 15.19 (a) (5) and this statement appears in a prominent location in the user manual.

Please contact me if there is any information you may need.

Thank you,

Signature:

A handwritten signature in blue ink, appearing to read "Assaf Shklar". The signature is written in a cursive, flowing style.

Assaf Shklar  
VP Operations