



YYYY MM
Date of Manufacture

Label P/N: 082-0004-050 Rev: 0

Imaging Dynamics Company Ltd.

3510, 29 Street NE Calgary,
Alberta, Canada, T1Y 7E5

Digital Radiographic Detector

Model: 1417WC

Input: 24VDC, 30VA

Serial
Number:

823073003224

100-1000-001

Made in CANADA

MEDICAL IMAGING EQUIPMENT - CLASSIFIED WITH RESPECT
TO SHOCK, FIRE, AND MECHANICAL HAZARDS.

This equipment complies with IEC 60601-1, IEC
60601-1-2 and 21 CFR Subchapter J



TYPE B



C-UL-US

FCC ID : 2AJDJ1417WC

This equipment complies with IEC 60601-1, IEC
60601-1-2 and 21CFR Subchapter J

This device complies with Part 15 of the FCC Rules.

Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference
- (2) This device must accept any interference received, including interference that may cause undesired operation.

IC:21810-1417WC

This Class A digital apparatus complies with Canadian
ICES-003.

5150-5250 MHz is indoor use only.