



Regulatory

You may need to open the app to see notifications.

Do not bring your ProAir Dihaler inhaler into an MRI or CT/x-ray scan room.

Do not expose your inhaler to other known sources of Electromagnetic Interference (such as diathermy and RF ID systems)

The electronic module complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

(1) The electronic module may not cause harmful interference, and

(2) The electronic module must accept any interference received, including interference that may cause undesired operation.

Changes or modifications to the electronic module not expressly approved by Teva could void the user's authority to operate the equipment.