

SAR Test Exclusion Evaluation

FCC ID: 2AJZ6HT100B

Product Name: Blood Glucose Meter

Model No: (1)Tyson Bio HT100-B (2)JT100-B


The standalone RFExp exclusion shall be refer to FCC § 1.1307 (b)(3)(i)(A) SAR-Based and MPE-based exemption which device determined the All distance from antenna to user/bystander.

DSSS

Frequency (GHz)	Tune-up max power (mW)	Exclusion Power (mW)
2.402	0.8	1

The EUT's maximum time-averaged power is less than 1mW so it preclude RFExp requirement.

Sincerely Yours,



Mr. Johnny Hsueh
Section Manager
AUDIX Technology Corporation