

Appendix B: Tissue Stimulating Liquids, System Checks and FCC System Validation

B.1. SAR System Check

Prior to SAR assessment, the system is verified to $\pm 10\%$ of the SAR measurement on the reference dipole at the time of calibration by the calibration facility. SAR System Validation was performed and complies with requirements per IEC/IEEE 62209-1528

Table 1 System Check Results (SAR)

System	Frequency (MHz)	Tissue Type	Date	Amb. Temp. (°C)	Tissue Temp (°C)	Input Power (dBm)	Verification Source SN	Probe SN	DAE SN	Measured 1g SAR (W/Kg)	1W Target 1g SAR (W/Kg)	1W Normalized 1g SAR (W/Kg)	1g SAR Deviation	Measured 10g SAR (W/Kg)	1W Target 10g SAR (W/Kg)	1W Normalized 10g SAR (W/Kg)	10g SAR Deviation
Delta	750	Head	05/29/2025	22.4	22.8	17	1235	7853	1844	0.418	8.690	8.340	-4.03%	0.274	5.690	5.467	-3.92%
Delta	750	Head	06/02/2025	23.3	22.6	17	1235	7853	1844	0.412	8.690	8.220	-5.40%	0.270	5.690	5.387	-5.32%
Delta	750	Head	06/04/2025	23.3	23	17	1235	7853	1844	0.422	8.690	8.420	-3.11%	0.275	5.690	5.487	-3.57%
Delta	835	Head	05/27/2025	23.1	22.8	17	4d311	7853	1844	0.503	9.820	10.036	2.20%	0.326	6.370	6.505	2.11%
Delta	835	Head	06/04/2025	23.3	23	17	4d311	7853	1844	0.501	9.820	9.996	1.79%	0.325	6.370	6.485	1.80%
Beta	1750	Head	05/28/2025	23.5	22.6	17	1205	7859	1905	1.740	36.500	34.718	-4.88%	0.938	19.200	18.716	-2.52%
Beta	1750	Head	06/04/2025	23.1	22.8	17	1205	7859	1905	1.720	36.500	34.319	-5.98%	0.926	19.200	18.476	-3.77%
Beta	1900	Head	06/02/2025	23.3	22.8	17	5d252	7859	1905	1.920	39.800	38.309	-3.75%	1.010	21.000	20.152	-4.04%
Beta	2450	Head	05/19/2025	23.5	22.8	17	1112	7859	1862	2.580	50.600	51.478	1.73%	1.210	23.700	24.143	1.87%
Beta	2450	Head	06/04/2025	23.1	22.8	17	1112	7859	1905	2.470	50.600	49.283	-2.60%	1.160	23.700	23.145	-2.34%

B.2. Dielectric Parameters of the TSL

Table 2 SAR Tissue Dielectric Parameters

Date	Tissue Type	Liquid Temp (°C)	Frequency (MHz)	Conductivity Measured (σ)	Conductivity Target (σ)	Deviation	Permittivity measured (εr)	Permittivity Target (εr)	Deviation
5/19/2025	Head	22.8	2300	1.72	1.67	3.32%	37.6	39.5	-4.62%
5/19/2025	Head	22.8	2310	1.73	1.68	3.21%	37.6	39.4	-4.59%
5/19/2025	Head	22.8	2320	1.74	1.68	3.11%	37.6	39.4	-4.58%
5/19/2025	Head	22.8	2350	1.76	1.71	2.66%	37.6	39.4	-4.56%
5/19/2025	Head	22.8	2400	1.79	1.76	2.23%	37.5	39.3	-4.59%
5/19/2025	Head	22.8	2450	1.83	1.80	1.70%	37.4	39.2	-4.59%
5/19/2025	Head	22.8	2480	1.85	1.83	1.17%	37.3	39.2	-4.67%
5/27/2025	Head	22.8	835	0.94	0.90	4.29%	40.5	41.5	-2.34%
5/27/2025	Head	22.8	850	0.94	0.92	2.95%	40.5	41.5	-2.45%
5/27/2025	Head	22.8	875	0.95	0.94	1.31%	40.4	41.5	-2.66%
5/27/2025	Head	22.8	900	0.96	0.97	-0.71%	40.3	41.5	-2.87%
5/27/2025	Head	22.8	935	0.98	0.99	-1.14%	40.2	41.5	-2.96%
5/28/2025	Head	22.6	1710	1.36	1.35	0.62%	40.7	40.1	1.41%
5/28/2025	Head	22.6	1750	1.38	1.37	0.76%	40.6	40.1	1.40%
5/28/2025	Head	22.6	1790	1.40	1.39	0.71%	40.6	40.0	1.39%
5/29/2025	Head	22.8	680	0.89	0.89	0.24%	41.5	42.3	-1.73%
5/29/2025	Head	22.8	695	0.89	0.89	0.67%	41.5	42.2	-1.69%
5/29/2025	Head	22.8	700	0.89	0.89	0.53%	41.5	42.2	-1.76%
5/29/2025	Head	22.8	710	0.90	0.89	1.22%	41.4	42.1	-1.66%
5/29/2025	Head	22.8	750	0.91	0.89	2.12%	41.3	41.9	-1.52%
5/29/2025	Head	22.8	770	0.92	0.89	2.87%	41.3	41.8	-1.30%
5/29/2025	Head	22.8	785	0.92	0.89	3.26%	41.2	41.7	-1.26%
5/29/2025	Head	22.8	800	0.93	0.90	3.60%	41.1	41.7	-1.28%
6/2/2025	Head	23.1	680	0.88	0.89	-0.79%	41.0	42.3	-2.96%
6/2/2025	Head	23.1	695	0.88	0.89	-0.34%	41.0	42.2	-2.92%
6/2/2025	Head	23.1	700	0.89	0.89	-0.47%	40.9	42.2	-2.98%
6/2/2025	Head	23.1	710	0.89	0.89	0.23%	40.9	42.1	-2.88%
6/2/2025	Head	23.1	750	0.90	0.89	1.10%	40.8	41.9	-2.75%
6/2/2025	Head	23.1	770	0.91	0.89	1.85%	40.7	41.8	-2.56%
6/2/2025	Head	23.1	785	0.91	0.89	2.25%	40.7	41.7	-2.54%
6/2/2025	Head	23.1	800	0.92	0.90	2.60%	40.6	41.7	-2.56%
6/2/2025	Head	22.8	1850	1.42	1.40	1.30%	40.1	40.0	0.32%
6/2/2025	Head	22.8	1880	1.43	1.40	2.41%	40.1	40.0	0.25%
6/2/2025	Head	22.8	1900	1.44	1.40	3.07%	40.1	40.0	0.15%
6/2/2025	Head	22.8	1920	1.46	1.40	3.94%	40.0	40.0	0.05%
6/2/2025	Head	22.8	1950	1.48	1.40	5.39%	40.0	40.0	0.00%
6/2/2025	Head	22.8	2000	1.50	1.40	7.48%	40.0	40.0	-0.12%
6/4/2025	Head	23	680	0.88	0.89	-0.30%	40.9	42.3	-3.16%
6/4/2025	Head	23	695	0.89	0.89	0.14%	40.9	42.2	-3.12%
6/4/2025	Head	23	700	0.89	0.89	0.00%	40.9	42.2	-3.19%
6/4/2025	Head	23	710	0.89	0.89	0.68%	40.8	42.1	-3.08%
6/4/2025	Head	23	750	0.91	0.89	1.48%	40.7	41.9	-2.94%
6/4/2025	Head	23	770	0.91	0.89	2.25%	40.7	41.8	-2.75%
6/4/2025	Head	23	785	0.92	0.89	2.66%	40.6	41.7	-2.72%
6/4/2025	Head	23	800	0.92	0.90	2.99%	40.5	41.7	-2.74%
6/4/2025	Head	23	815	0.93	0.90	3.64%	40.5	41.6	-2.64%
6/4/2025	Head	23	835	0.94	0.90	4.11%	40.4	41.5	-2.54%
6/4/2025	Head	23	850	0.94	0.92	2.71%	40.4	41.5	-2.62%
6/4/2025	Head	23	875	0.95	0.94	1.02%	40.3	41.5	-2.82%
6/4/2025	Head	23	900	0.96	0.97	-0.96%	40.3	41.5	-3.00%
6/4/2025	Head	23	935	0.97	0.99	-1.48%	40.2	41.5	-3.05%
6/4/2025	Head	22.8	1600	1.27	1.28	-0.83%	40.6	40.3	0.61%
6/4/2025	Head	22.8	1650	1.30	1.31	-1.06%	40.5	40.2	0.54%
6/4/2025	Head	22.8	1700	1.32	1.34	-1.31%	40.4	40.2	0.63%
6/4/2025	Head	22.8	1710	1.33	1.35	-1.31%	40.4	40.1	0.64%
6/4/2025	Head	22.8	1750	1.36	1.37	-1.09%	40.3	40.1	0.66%
6/4/2025	Head	22.8	1790	1.38	1.39	-1.01%	40.3	40.0	0.67%
6/4/2025	Head	22.8	2400	1.78	1.76	1.53%	39.4	39.3	0.26%
6/4/2025	Head	22.8	2450	1.82	1.80	1.17%	39.3	39.2	0.27%
6/4/2025	Head	22.8	2480	1.84	1.83	0.66%	39.2	39.2	0.22%
6/4/2025	Head	22.8	2500	1.86	1.85	0.31%	39.2	39.1	0.20%
6/4/2025	Head	22.8	2550	1.90	1.91	-0.56%	39.1	39.1	0.14%
6/4/2025	Head	22.8	2600	1.94	1.96	-1.20%	39.1	39.0	0.11%

The above measured tissue parameters were used in the DASY software. The DASY software was used to perform interpolation to determine the dielectric parameters at the SAR test device frequencies (per KDB Publication 865664 D01v01r04 and IEEE 1528-2013 6.6.1.2). The tissue parameters listed in the SAR test plots may slightly differ from the table above due to significant digit rounding in the software.

The SAR values were compensated for deviations between the measured and required tissue dielectric properties, as described in IEC/IEEE 62209-1528. The SAR values were applied to only scale up the measured SAR values, and not downward, per KDB Publication 865664 D01v04r04.

B.3. FCC System Validation

Per FCC KDB Publication 865664 D02 Section 2.3 a) states “SAR system validation status and system verification results should be documented in a separate section of the SAR report, or as an attachment, to confirm measurement accuracy.”

The SAR systems used for evaluating this device were validated against its performance specifications prior to the SAR measurements.

Reference dipoles were used with the required tissue-equivalent media for system validation, according to the procedures outlined in FCC KDB Publication 865664 D01 and IEEE 1528-2013. Since SAR probe calibrations are frequency dependent, each probe calibration point was validated at a frequency within the valid frequency range of the probe calibration point.

Per FCC KDB 865664 D02, “the validation status should be documented according to the validation date(s), measurement frequencies, SAR probes, calibrated signal type(s) and tissue dielectric parameters.” A tabulated summary of the system validation status is provided accordingly:

Table 3 System Validation

System	Frequency (MHz)	Date	Probe	DAE	Probe CalF		Cond. (σ)	Perm (ϵ_r)	CW Validation			Mod Validation		
					Freq (MHz)	Tissue Type			Sensitivity	Probe Linearity	Probe Isotropy	Mod Type	Duty Factor	PAR
Delta	750	11/18/2024	7853	1844	750	Head	0.906	43.8	PASS	PASS	PASS	N/A	N/A	N/A
Delta	835	11/18/2024	7853	1844	835	Head	0.936	43.6	PASS	PASS	PASS	GMSK	PASS	N/A
Beta	1750	05/28/2025	7859	1905	1750	Head	1.38	40.6	PASS	PASS	PASS	N/A	N/A	N/A
Beta	1900	05/28/2025	7859	1905	1900	Head	1.47	40.4	PASS	PASS	PASS	GMSK	PASS	N/A
Beta	2450	05/13/2025	7859	1862	2450	Head	1.72	38.4	PASS	PASS	PASS	OFDM	N/A	PASS
Beta	2450	05/21/2025	7859	1905	2450	Head	1.72	37.8	PASS	PASS	PASS	OFDM	N/A	PASS

NOTE: The probes have been calibrated for both CW and modulated signals. Modulations in the table above represent test configurations for which the measurement system has been validated per FCC KDB Publication 865664 D01 for scenarios when CW probe calibrations are used with other signal types.

SAR systems were additionally validated for modulated signals with a periodic duty cycle or with a high PAR (peak to average ratio) >5 dB, such as OFDM according to FCC KDB Publication 865664 D01 v01r04.

B.4. Sample TSL Compositions

TSL recipes are proprietary to SPEAG. Since the composition is approximate to the actual liquids utilized, the manufacturer data sheets are provided below.




Figure 1 - Head TSL Calibration Certificate Example