

Test Report 21-1-0179501T014a



18 2023-Mar-16 Number of pages: Date of Report:

CETECOM GmbH Sentec AG **Testing company:** Applicant:

> Im Teelbruch 116 45219 Essen Germany Tel. + 49 (0) 20 54 / 95 19-0 Fax: +49 (0) 20 54 / 95 19-150

Product: Patient Monitor

Model: tCOM+

FCC ID: 2ALTU-TCOM IC: 22941-TCOM

Testing has been carried out in accordance with: **FCC Regulations**

Title 47 CFR, Chapter I, Subchapter A, Part 15

Subpart B Unintentional Radiators

§ 15.107 Conducted limits § 15.109 Radiated emission limits

ISED-Regulations

Radio Standards Specification

RSS-Gen, Issue 5

General Requirements for Compliance of Radio Apparatus

ICES-003, Issue 7

Information Technology Equipment (including Digital Apparatus)

Tested Technology:

Test Results: ☑ The EUT complies with the requirements in respect of all parameters subject to the test.

The test results relate only to devices specified in this document

Signatures:

Ninovic Perez

Test Lab Manager

Authorization of test report

Hicham Laayouni Test Manager

Responsible of test report



Table of Contents

Ta	ble o	of Annex	3
1	(General information	4
	1.1	Disclaimer and Notes	4
	1.2	Attestation	4
	1.3	Summary of Test Results	5
	1.4	Summary of Test Methods	5
2	,	Administrative Data	6
	2.1	Identification of the Testing Laboratory	6
	2.2	General limits for environmental conditions	6
	2.3	Test Laboratories sub-contracted	6
	2.4	Organizational Items	6
	2.5	Applicant's details	6
	2.6	Manufacturer's details	6
	2.7	Equipment under Test (EUT)	7
	2.8	Untested Variant (VAR)	7
	2.9	Auxiliary Equipment (AE)	7
	2.10	0 Connected cables (CAB)	7
	2.1	1 Software (SW)	7
	2.1	2 EUT set-ups	7
	2.13	3 EUT operation modes	8
3	ı	Equipment under test (EUT)	8
	3.1	General Data of Main EUT as Declared by Applicant	8
	3.2	Modifications on Test sample	8
4	ı	Measurements	9
	4.1	Conducted interference voltage (AC mains)	9
	4.2	Radiated field strength emissions 30 MHz – 1 GHz	11
	4.3	Equipment lists	13
	4.4	Radiated field strength emissions above 1 GHz	14
	4.5	Equipment lists	16
5	ı	Results from external laboratory	17
6	(Opinions and interpretations	17
7	ı	List of abbreviations	17
8	ı	Measurement Uncertainty valid for conducted/radiated measurements	18
9	,	Versions of test reports (change history)	18



	Table of Annex				
Annex No.	Contents	Reference Description	Total Pages		
Annex 1 Test result diagrams		TR21-1-0179501T014a-A1	7		
Annex 2 Internal photographs of EUT					
Annex 3	External photographs of EUT	TR21-1-0179501T014a-A3	6		
Annex 4	Test set-up photographs	TR21-1-0179501T014a-A4	4		
The listed attachments are separate documents.					

TR21-1-0179501T014a 3/18



_

1 General information

1.1 Disclaimer and Notes

The test results of this test report relate exclusively to the test item specified in this test report as specified in chapter 2.7. CETECOM does not assume responsibility for any conclusions and generalizations drawn from the test results with regard to other specimens or samples of the type of the equipment represented by the test item.

The test report may only be reproduced or published in full. Reproduction or publication of extracts from the report requires the prior written approval of CETECOM.

The testing service provided by CETECOM has been rendered under the current "General Terms and Conditions for CETECOM". CETECOM will not be liable for any loss or damage resulting from false, inaccurate, inappropriate or incomplete product information provided by the customer.

Under no circumstances does the CETECOM test report include any endorsement or warranty regarding the functionality, quality or performance of any other product or service provided.

Under no circumstances does the CETECOM test report include or imply any product or service warranties from CETECOM, including, without limitation, any implied warranties of merchantability, fitness for purpose, or non-infringement, all of which are expressly disclaimed by CETECOM.

All rights and remedies regarding vendor's products and services for which CETECOM has prepared this test report shall be provided by the party offering such products or services and not by CETECOM.

In no case this test report can be considered as a Letter of Approval.

This test report is electronically signed and valid without handwritten signature. For verification of the electronic signatures, the public keys can be requested at the testing laboratory.

The test report must always be reproduced in full; reproduction of an excerpt only is subject to written approval of the testing laboratory. The documentation of the testing performed on the tested devices is archived for 10 years at CETECOM.

Also we refer on special conditions which the applicant should fulfill according §2.927 to §2.948, special focus regarding modification of the equipment and availability of sample equipment for market surveillance tests.

1.2 Attestation

I declare that all measurements were performed by me or under my supervision and that all measurements have been performed and are correct to my best knowledge and belief to Industry Canada standards. All of the above requirements are met in accordance with enumerated standards.

TR21-1-0179501T014a 4/18



1.3 Summary of Test Results

Test case	Reference	Reference	Reference	Page	Remark	Result
	in FCC 🛛	in ISED 🛮	in RSS-GEN ⊠			
AC-Power Lines Conducted	§15.107	ICES-003,	RSS Gen, Issue 5,	10		passad
<u>Emissions</u>		Issue 7	Chapter 8.8	10		passed
Radiated field strength emissions	§15.109	ICES-003,	RSS-Gen., Issue 5			
30 MHz – 1 GHz	§15.33	Issue 7	Chapter 8.9,	12		passed
	§15.35		Chapter 7.3			
Radiated field strength emissions	§15.109	ICES-003,	RSS-Gen., Issue 5			
above 1 GHz	§15.33	Issue 7	Chapter 8.9,	14		passed
	§15.35		Chapter 7.3			

PASSED The EUT complies with the essential requirements in the standard.

FAILED The EUT does not comply with the essential requirements in the standard.

N/A Test case does not apply to the test object.

NP The test was not performed by the CETECOM Laboratory.

Decision Rule: CETECOM GmbH follows ILAC G8:2019 chapter 4.2.1 (Simple Acceptance Rule).

1.4 Summary of Test Methods

Test case	Test method
AC-Power Lines Conducted Emissions	ANSI C63.4-2014 chapter 7
Radiated field strength emissions 30 MHz – 1 GHz	ANSI C63.4-2014 chapter 8.2.3
Radiated field strength emissions above 1 GHz	ANSI C63.4-2014 chapter 8.3

TR21-1-0179501T014a 5/18



2 Administrative Data

2.1 Identification of the Testing Laboratory

Company name: CETECOM GmbH

Address: Im Teelbruch 116 45219 Essen - Kettwig

Germany

Responsible for testing laboratory: Dipl.-Ing. Ninovic Perez

Accreditation scope: DAkkS Webpage: FCC ISED

IC Lab company No. / CAB ID: 3462D / DE0005

Test location: CETECOM GmbH; Im Teelbruch 116; 45219 Essen - Kettwig

2.2 General limits for environmental conditions

Temperature:	22±2 °C
Relative. humidity:	45±15% rH

2.3 Test Laboratories sub-contracted

Company name: --

2.4 Organizational Items

Responsible test manager: Hicham Laayouni

Receipt of EUT: 20.12.2022

Date(s) of test: 2022-Dec-21 to 2022-Dec-21

Version of template: 22.0901

2.5 Applicant's details

Applicant's name: Sentec AG

Address: Ringstrasse 39

4106 Therwil

Switzerland

Contact Person: Oliver Friedli

Contact Person's Email: oliver.friedli@sentec.com

2.6 Manufacturer's details

Manufacturer's name:	Sentec AG
Address:	Ringstrasse 39
	4106 Therwil
	Schweiz

TR21-1-0179501T014a 6/18



2.7 Equipment under Test (EUT)

EUT No.*)	Sample No.	Product	Model	Туре	SN	HW	SW
EUT 1	21-1-01795S18_C01	Patient Monitor	tCOM+	N/A	000014	103164 0001B	V00.01.0 3

^{*)} EUT short description is used to simplify the identification of the EUT in this test report.

2.8 Untested Variant (VAR)

VAR	Sample No.	Product	Model	Туре	SN	HW	SW
No.*)							

^{*)} The listed additional untested model variant(s) (VAR) is/are not object of evaluation of compliance. For further information please see Annex 5: Declaration of applicant of model differences.

2.9 Auxiliary Equipment (AE)

AE	Sample No.	Auxiliary Equipment	Model	SN	HW	SW
No.*)						
AE 1	21-1-01795S21_C01	Sensor	V-Sign Sensor	343241	N/A	N/A
AE 2	21-1-01795S25_C01	Flash Drive	USB flash drive (type C)	N/A	N/A	N/A
AE 3	21-1-01795S26_C01	Flash Drive	USB flash drive (type C)	N/A	N/A	N/A
AE 4	21-1-01795S30_C01	Power Supply	Power Supply	N/A	N/A	N/A
AE 5	21-1-01795S33_C01	Sensor Membrane Changer	Membrane Changer MC- R3 LOT 220946	N/A	N/A	N/A

^{*)} AE short description is used to simplify the identification of the auxiliary equipment in this test report. If the table above does not show any other line than the headline, no AE was used during testing nor was taken into account for evaluation

2.10 Connected cables (CAB)

CAB No.*)	Sample No.	Cable Type	Connectors / Details	Length
CAB 1	21-1-01795S22_C01	Cable		15 cm
CAB 2	21-1-01795S27_C01	Cable		100 cm
CAB 3	21-1-01795S28_C01	Cable		cm

^{*)} CAB short description is used to simplify the identification of the connected cables in this test report. If the table above does not show any other line than the headline, no cable was used during testing nor was taken into account for evaluation

2.11 Software (SW)

SW	Sample No.	SW Name	Description	SW Status
No.*)				

^{*)} SW short description is used to simplify the identification of the used software in this test report. If the table above does not show any other line than the headline, no SW was used during testing nor was taken into account for evaluation.

2.12 EUT set-ups

set-up no.*)	Combination of EUT and AE	Description
1	EUT 1 + AE 1 + AE 2 + AE 3 + AE 4 + AE 5 + CAB 1 + CAB 2 + CAB 3	Used for all measurements

TR21-1-0179501T014a 7 / 18

If the table above does not show any other line than the headline, no untested variants are available.



^{*)} EUT set-up no. is used to simplify the identification of the EUT set-up in this test report.

2.13 EUT operation modes

EUT operating mode no.*)	Operating modes	Additional information
	Monitoring function on	Normal monitoring function switched on,
Op. 1	+	Ping between EUT and laptop via Ethernetcable,
	Ping	all radios are off

^{*)} EUT operating mode no. is used to simplify the test report.

3 Equipment under test (EUT)

3.1 General Data of Main EUT as Declared by Applicant

Firmware	☐ for normal use ☐ Special version for test execution					
Power supply	□ AC Mains	single Line (L1/N) 120 V 60 Hz				
	☐ DC Mains					
	☐ Battery					
Operational conditions	T _{nom} =XX °C					
EUT sample type	Production					
Weight	2.500 kg					
Size [LxWxH]	27.0 cm x 16.0 cm x 16.0 cm					
Interfaces/Ports	Interfaces/Ports					
For further details refer Applicants Declaration & following technical documents						

3.2 Modifications on Test sample

Additions/deviations or exclusions	

TR21-1-0179501T014a 8/18



4 Measurements

4.1 Conducted interference voltage (AC mains)


4.1.1 Description of the general test setup and methodology

The measurement is performed as follows:

The entire frequency range is swept with Peak and Average detectors on all lines. The test result curves are based on this Preview Test. If there are peaks with margin less than 10 dB to the limits for emissions, the Final Test is started. All frequencies at which margin to the limits is less than 10 dB are scanned with Quasi peak or Average detectors on corresponding lines around those frequencies. These final values are entered to the test results curve and tables.

Diagrams show the peak values as a sum of measured lines in maxhold mode.

Schematic:



Testing method:

The measurement is made according to relevant reference clauses: (See Tables Summary of Test Results and Summary of Test Methods on page 5)

Exploratory, preliminary measurements

As a first step, determines the worst-case phase line (neutral or phase) as well as the most critical operating mode of the equipment. A complete frequency-sweep with PK-Detector is performed on each current-carrying conductor.

Final measurement on critical frequencies

For power phases and critical frequencies (Margin to AV- or QP limit lower than 3 dB) as a second step includes measurements with receivers detector set to Quasi-Peak and Average.

Formula:

 $\begin{array}{cccc} V_C = V_R + C_L & \text{(1)} & V_C = \text{measured Voltage} - \text{corrected value} \\ M = L_T - V_C & \text{(2)} & V_R = \text{Receiver reading} \\ & C_L = \text{Cable loss} \\ & M = \text{Margin} \\ & L_T = \text{Limit} \end{array}$

All units are dB-units, positive margin means value is below limit.

TR21-1-0179501T014a 9 / 18



4.1.2 Test receiver settings

Detector	Peak	Average	Quasi peak
Min. attenuation 10 dB		10 dB	10 dB
Resolution bandwidth	bandwidth 9 kHz 9 kH		9 kHz
Dector Meas-time	10 ms	Pre-measurement: 10 ms Final measurement: 1 s	1 s
Step size	3 kHz Pre-measurement: 3 kHz Final measurement: selected frequencies		Selected frequencies
Preamp	Off	Off	Off

4.1.3 Measurement Location

4.1.4 Limit

Frequency Range [MHz]	Class B ⊠		Class A	
	QUASI-Peak [dBμV] AVERAGE [dBμV]		QUASI-Peak [dBμV]	AVERAGE [dBμV]
0.15 - 0.5	66 to 56*	56 to 46*	79	66
0.5 – 5	56	46	73	60
5 – 30	60 50		73	60

4.1.5 **Result**

Diagram	Set-up	Mode	Power Line	Max [dBμV]	Detector	Result
1.01	01	01	N/L1	44.17 @23.97MHz	QP	pass

Remark: for more information and graphical plot see annex A1 TR21-1-0179501T014a-A1

TR21-1-0179501T014a 10/18




4.2 Radiated field strength emissions 30 MHz – 1 GHz

4.2.1 Description of the general test setup and methodology, see below example:

Evaluating the field emissions have to be done first by an exploratory emissions measurement and a final measurement for most critical frequencies. The tests are performed in a NSA-compliant semi anechoic room (SAR) recognized by the regulatory commissions.

Schematic:



Testing method:

The measurement is made according to relevant reference clauses: (See Tables *Summary of Test Results* and *Summary of Test Methods* on page 5)

Exploratory, preliminary measurements

The EUT and its associated accessories are placed on a non-conductive position manipulator (tipping device) of 0.8 m height which is placed on the turntable. By rotating the turntable (range 0° to 360°, step 90°) and the EUT itself either on 3-orthogonal axis (portable equipment) or 2-orthogonal axis (defined operational position of EUT) the emission spectrum and its characteristics was recorded with an EMI-receiver, broadband antenna and software.

Measurement antenna: horizontal and vertical, heights: 1,0 m and 1,82 m as worst-case determined by an exploratory emission measurements. The results are documented in a diagram. Critical frequencies (low margin to limit) are saved within a table for further investigations. If various operating modes are supported, further investigations are made to find the worst-case of them. Also the interconnection cables and equipment position were varied in order to maximize the emissions.

Final measurement on critical frequencies

Based on the exploratory measurements, the most critical frequencies are re-measured by main-taining the EUT's worst-case operation mode, cable position, etc. either on 10m OATS or 3m semi-anechoic room.

First a frequency zoom around the critical frequency is done to locate the frequency more precisely. After this step, for all identified critical frequencies, the maximum peak was determined.

Following parameters were varied: the turntable angle continuously in the range 0 to 360 degree, the EUT itself either over 3-orthogonal axis (not defined usage position) or 2-orthogonal axis (defined usage position). The measurement antenna height between 1 m and 4 m.

TR21-1-0179501T014a 11 / 18



On the determined worst-case position, a final measurement with necessary bandwidth and detector according standard has been carried out

Formula:

 $E_C = E_R + AF + C_L + D_F - G_A$ (1) AF = Antenna factor

C_L = Cable loss

 $M = L_T - E_C$ (2) $D_F = Distance correction factor (if used)$

E_C = Electrical field – corrected value

E_R = Receiver reading

G_A = Gain of pre-amplifier (if used)

L_T = Limit

M = Margin

All units are dB-units, positive margin means value is below limit.

4.2.2 Sample calculation

Raw- Value [dBuV/m]	Antenna factor	Distance Correction [dB]	Cable Loss	Preamplifier	Resulting correction value [dB]	Final result [dBuV/m]	Remarks
32.7	22.25		3.1		25.35	58.05	

Remark: This calculation is based on an example value at 800.4 MHz

4.2.3 Measurement Location

Test site 120901 - SAC - Radiated Emission <1GHz
--

4.2.4 Limit

Frequency Range	Class B	☑ (3 meters)	Class A	☐ (10 meters)		
[MHz]	Limit [μV/m]	Limit [dBμV/m]	Limit [μV/m]	Limit [dBμV/m]	Detector	RBW / VBW [kHz]
30 - 88	100	40.0	90	39.0	Quasi peak	100 / 300
88 - 216	150	43.5	150	43.5	Quasi peak	100 / 300
216 - 960	200	46.0	210	46.4	Quasi peak	100 / 300
960 - 1000	500	54.0	300	49.5	Quasi peak	100 / 300

4.2.5 Result

Diagram	Set-up	Mode Maximum Level [dBμV/m]		Result
			Frequency Range 30 – 1000 MHz	
3.01	01	01	41.82 dBμV/m @ 184.99MHz	

Remark: for more information and graphical plot see annex A1 TR21-1-0179501T014a-A1

TR21-1-0179501T014a 12 / 18



4.3 Equipment lists

ID	Description	Manufacturer	SerNo	CheckType	Last Check	Interval	Next Check
	120901 - SAC - Radiated Emission <1GHz			calchk	cal: 2015-Jul-21	cal: 10Y	cal: 2025-Jul-21
					chk: 2021-Jul-27	chk: 12M	chk: 2022-Jul-27
20341	Digital Multimeter Fluke 112	Fluke Deutschland GmbH / Glottertal	81650455	cal	cal: 2022-May-18	cal: 24M	cal: 2024-May-18
20442	Semi Anechoic Chamber	ETS-Lindgren Gmbh / Taufkirchen	-	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20482	filter matrix Filter matrix SAR 1	CETECOM GmbH	-	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20574	Biconilog Hybrid Antenna BTA-L	Frankonia GmbH / Heideck	980026L	cal	cal: 2022-Jun-15	cal: 36M	cal: 2025-Jun-15
20620	Test Receiver ESU26	Rohde & Schwarz Messgerätebau GmbH /	100362	cal	cal: 2022-Jun-08	cal: 12M	cal: 2023-Jun-08
		Memmingen					
20885	Power Supply EA3632A	Agilent Technologies Deutschland GmbH	75305850	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
25038	Loop Antenna HFH2-Z2	Rohde & Schwarz Messgerätebau GmbH /	879824/13	cal	cal: 2022-Jul-04	cal: 24M	cal: 2024-Jul-04
		Memmingen					

4.3.1 Legend

10Y

Note / remarks Interval of calibration & Verification 12M 12 months 24M 24 months 36M 36 months

10 Years

Abbreviation Check Type Description	
cnn	Calibration and verification not necessary
cal	Calibration
calchk	Calibration plus intermediate Verification
chk	Verification
cpu	Verification before usage

TR21-1-0179501T014a 13/18




4.4 Radiated field strength emissions above 1 GHz

4.4.1 Description of the general test setup and methodology, see below example:

Evaluating the emissions have to be done first by an exploratory emissions measurement and a final measurement for most critical frequencies. The tests are performed in a CISPR 18-1-4:2010 compliant fully anechoic room (FAR) recognized by the regulatory commission. The measurement distance was set to 3 meter for frequencies up to 18 GHz and 2 meter above 18 GHz. A logarithmic periodic antenna is used for the frequency range 30 MHz to 1 GHz. Horn antennas are used for frequency range 1 GHz to 40 GHz. The EUT is aligned within 3 dB beam width of the measurement antenna with three orthogonal axis measurements on the EUT.

Schematic:



Testing method:

The measurement is made according to relevant reference clauses: (See Tables Summary of Test Results and Summary of Test Methods on page 5)

Exploratory, preliminary measurements

The EUT and its associated accessories are placed on a non-conductive position manipulator (tipping device) of 1.55 m height which is placed on the turntable. By rotating the turntable (range 0° to 360°, step 15°) and the EUT itself either on 3-orthogonal axis (portable equipment) or 2-orthogonal axis (defined operational position of EUT) the emission spectrum and its characteristics was recorded with an EMI-receiver, broadband antenna and software.

The measurements are performed in horizontal and vertical polarization of the measurement antennas. The results are documented in a diagram. Critical frequencies (low margin to limit) are saved within a table for further investigations. If various operating modes are supported, further investigations are made to find the worst-case of them. Also the interconnection cables and equipment position were varied in order to maximize the emissions.

Final measurement on critical frequencies

Based on the exploratory measurements, the most critical frequencies are re-measured by main-taining the EUT's worst-case operation mode, cable position, etc.

First a frequency zoom around the critical frequency is done to locate the frequency more precisely. After this step, for all identified critical frequencies, the maximum peak was determined.

Following parameters were varied: the turntable angle continuously in the range 0 to 360 degree, the EUT itself over 3-orthogonal axis and the height for EUT with large dimensions or three axis scan for portable/small equipment.

TR21-1-0179501T014a 14 / 18



On the determined worst-case position, a final measurement with necessary bandwidth and detector according standard has been carried out.

Formula:

 $E_C = E_R + A_F + C_L + D_F - G_A$ (1) $E_C = E_C + E_C +$

 E_R = Receiver reading

 $M = L_T - E_C$ (2) M = Margin

 $L_T = Limit$

 A_F = Antenna factor

C_L = Cable loss

D_F = Distance correction factor (if used) G_A = Gain of pre-amplifier (if used)

All units are dB-units, positive margin means value is below limit.

4.4.2 Sample calculation

Raw- Value [dBuV/m]	Antenna factor	Distance Correction [dB]	Cable Loss + Preamplifier	Resulting correction value [dB]	Final result [dBuV/m]	Remarks
29.37	41.20		24.28	16.92	46.3	CableLoss and PreAmp data in one data correction file

Remark: This calculation is based on an example value at 10 GHz

4.4.3 Measurement Location

Test site 120907 - FAC2 - Radiated Emissions	
--	--

4.4.4 Limit

	Radiated emissions limits (3 meters)						
Frequency Range Limit Limit Detector RBW / VB $[MHz]$ $[\mu V/m]$ $[dB\mu V/m]$ $[kHz]$							
Above 1000	500	54	Average	1000 / 3000			
Above 1000	5000	74	Peak	1000 / 3000			

4.4.5 Result

Dia	gram	Set-up	Mode	Maximum Level [dBμV/m] Frequency Range 1 – 12.4 GHz	Result
4.0	1a	01	01	59.23@ 11.89 GHz	pass

Remark: for more information and graphical plot see annex A1 TR21-1-0179501T014a-A1

Diagram	Set-up	Mode	Maximum Level [dBμV/m] Frequency Range 12.4 – 18 GHz	Result
4.01b	01	01	52.02@ 13.40 GHz	pass

Diagram	Set-up	Mode	Maximum Level [dBμV/m] Frequency Range 18 – 40 GHz	Result
4.01c	01	01	56.65@ 38.83 GHz	pass

TR21-1-0179501T014a 15/18



4.5 Equipment lists

ID	Description	Manufacturer	SerNo	CheckType	Last Check	Interval	Next Check
	120907 - FAC2 - Radiated Emissions			chk			
					chk: 2021-Aug-30	chk: 12M	chk: 2022-Aug-30
20005	AC - LISN 50 Ohm/50µH ESH2-Z5	Rohde & Schwarz Messgerätebau GmbH /	861741/005	cal	cal: 2022-May-19	cal: 12M	cal: 2023-May-19
		Memmingen					
20133	Horn Antenna 3115 (Meas 1)	EMCO Elektronik GmbH	9012-3629	cal	cal: 2020-Apr-08	cal: 36M	cal: 2023-Apr-08
20302	Horn Antenna BBHA9170 (Meas 1)	Schwarzbeck Mess-Elektronik OHG / Schönau	155	cpu			
					chk: 2020-Apr-15	chk: 12M	
20412	Fully Anechoic Chamber 2	ETS-Lindgren Gmbh / Taufkirchen	without	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20729	FS-Z140	Rohde & Schwarz Messgerätebau GmbH	101004	cal	cal: 2020-May-26	cal: 36M	cal: 2023-May-26
20730	FS-Z110	Rohde & Schwarz Messgerätebau GmbH	101468	cal	cal: 2020-Jun-19	cal: 36M	cal: 2023-Jun-19
20731	FS-Z75	Rohde & Schwarz Messgerätebau GmbH /	101022	cal	cal: 2022-May-18	cal: 36M	cal: 2025-May-18
		Memmingen					
20732	Signal- and Spectrum Analyzer FSW67	Rohde & Schwarz Messgerätebau GmbH /	104023	cal	cal: 2022-Jun-08	cal: 12M	cal: 2023-Jun-08
		Memmingen					
20733	Harmonic Mixer FS-Z220	RPG-Radiometer Physics GmbH	101009	cal	cal: 2021-May-27	cal: 36M	cal: 2024-May-27
20734	Harmonic Mixer FS-Z325	RPG-Radiometer Physics GmbH	101005	cal	cal: 2021-May-27	cal: 36M	cal: 2024-May-27
20765	Pickett-Potter Horn Antenna FH-PP 40-60	RPG-Radiometer Physics GmbH / Meckenheim	010001	cal	cal: 2020-Sep-15	cal: 36M	cal: 2023-Sep-15
20767	Pickett-Potter Horn Antenna FH-PP 140-220	RPG-Radiometer Physics GmbH / Meckenheim	010011	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20811	Horn Antenna ASY-SGH-124-SMA	Antenna Systems Solutions S.L	29F14182337	cal	cal: 2021-Oct-20	cal: 36M	cal: 2024-Oct-20
20812	Pickett-Potter Horn Antenna FH-PP-325	RPG-Radiometer Physics GmbH	10024	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20813	Pickett-Potter Horn Antenna FH-PP 075	RPG-Radiometer Physics GmbH / Meckenheim	10006	cal	cal: 2020-Sep-09	cal: 36M	cal: 2023-Sep-09
20814	Pickett-Potter Horn Antenna FH-PP 140	RPG-Radiometer Physics GmbH	10008	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20815	Pickett-Potter Horn Antenna FH-PP 110	RPG-Radiometer Physics GmbH	10014	cal	cal: 2020-Sep-04	cal: 36M	cal: 2023-Sep-04
20816	SGH Antenna SGH-26-WR10	Anteral S.L.	1144	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20817	Waveguide Rectangular Horn Antenna SAR-	ERAVAN	13254-01	cal	cal: 2020-Jul-29	cal: 36M	cal: 2023-Jul-29
	2309-22-S2						
20836	1-18 GHz Amplifier	Wright Technologies, Inc., Inc.	0001	chk			
						chk: 36M	
20877	JS42-08001800-16-8P Verstärker	Miteq Inc.	2079991 / 2079992	chk			
					chk: 2020-Feb-27	chk: 36M	chk: 2020-May-27
20907	Waveguide WR-15 attenuator STA-30-15-M2	SAGE Millimeter Inc.	13256-01	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20908	Waveguide WR 10 attenuator STA-30-10-M2	SAGE Millimeter Inc.	13256-01	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20909	Waveguide Horn Antenna PE9881-24	Pasternack Enterprises, Inc.	37/2016	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20910	Frequency Multiplier 936VF-10/385	MI-Wave, Millimeter Wave Products Inc.	142	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20911	Frequency Multiplier 938WF-10/387	MI-Wave, Millimeter Wave Products Inc.	141	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20912	Low noise Amplifier Module 0.5-4GHz	RF-Lambda Europe GmbH	19041200083	cnn	cal: -	cal: -	cal: -
					chk: -	chk: -	chk: -
20913	Phase Amplitude Stable Cable Assembly DC-	RF-Lambda Europe GmbH	AC19040001	cnn	cal: -	cal: -	cal: -
	40GHz				chk: -	chk: -	chk: -
25457	DRG Horn Antenna SAS-574	A.H. Systems, Inc. / Chatsworth	383	cal	cal: 2022-Mar-28	cal: 36M	cal: 2025-Mar-28

Tools used in 'P3M1'

4.5.1 Legend

Note / remarks	Interval of calibration & Verification
12M	12 months
24M	24 months
36M	36 months
10Y	10 Years

Abbreviation Check Type	Description
cnn	Calibration and verification not necessary
cal	Calibration
calchk	Calibration plus intermediate Verification
chk	Verification
cpu Verification before usage	

TR21-1-0179501T014a 16/18



5	Results from 6	Results from external laboratory							
No	one	-							
6	Opinions and	interpretations							
No	one	-							
7	List of abbrevi	iations							
No	one								

TR21-1-0179501T014a 17/18



8 Measurement Uncertainty valid for conducted/radiated measurements

The reported uncertainties are calculated based on the standard uncertainty multiplied with the appropriate coverage factor \mathbf{k} , such that a confidence level of approximately 95% is achieved. For uncertainty determination, each component used in the concrete measurement set-up was taken in account and it contribution to the overall uncertainty according its statistical distribution calculated.

Measurement type	Frequen of meas Start [MHz]	cy range urement Stop [MHz]	Calculated Uncertainty based on confidence level of 95.54%	Remarks
Magnetic field strength	0.009	30	4.86	Magnetic loop antenna, Pre-amp on
	30	100	4.57	without Pre-Amp
	30	100	4.91	with PreAmp
	100	1000	4.02	without Pre-Amp
	100	1000	4.26	with PreAmp
	1000	18000	4.36	without Pre-Amp
	1000	18000	5.23	with PreAmp
RF-Output power (eirp)	18000	33000	4.92	Schwarzbeck BBHA9170 (#20302) Antenna set-up non-waveguide antenna)
Unwanted emissions (eirp)	33000	50000	4.17	Set-up for Q-Band (WR-22), non-wave guide antenna
[dB]	40000	60000	4.69	Set-up U-Band (WR-19), non-waveguide antenna
1. 1	50000	75000	4.06	External Mixer set-up V-Band (WR-15)
	75000	110000	4.17	External Mixer set-up W-Band (WR-6)
	90000	140000	5.49	External Mixer set-up F-Band (WR-8)
	140000	225000	6.22	External Mixer set-up G-Band (WR-5)
	225000	325000	7.04	External Mixer set-up (WR-3)
	325000	500000	8.84	External Mixer set-up (WR-2.2)
	1000	18000	2.85	Typical set-up with microwave generator and antenna, value for 7GHz calculated
	18000	33000	4.66	Typical set-up with microwave generator and antenna
Radiated Blocking	33000	50000	3.48	WR-22 set-up
[dB]	50000	75000	3.73	WR-15 set-up
	75000	110000	4.26	WR-6 set-up
	75000	110000	4.20	wr-o set-up
Frequency Error	40000	77000	276.19	calculated for 77 GHz (FMCW) carrier
[kHz]	6000	7000	33.92	calculated for 6.5GHz UWB Ch.5
	30	6000	1.11	Power measurement with Fast-sampling-detector
	30	6000	1.20	Power measurement with Spectrum-Analyzer
	30	6000	1.20	Power Spectrum-Density measurement
	30	7500	1.20	Conducted Spurious emissions:
	0.009	30	2.56	5. Conducted Spurious emissions:
TS 8997	2.4	2.48	1.95 ppm	6a. Bandwidth / 2-Marker Method for 2.4GHz ISM
conducted Parameters	5.18	5.825	7.180 ppm	6b. Bandwidth / 2-Marker Method for 5GHz WLAN
	5.18	5.825	1.099 ppm	7 Frequency (Marker method) for 5GHz WLAN
	30	6000	0.11561µs	8 Medium-Utilization factor / Timing
	30	6000	1.85	9 Blocking-Level of companion device
	30	6000	1.62	9 Blocking Generator level
Conducted emissions	0.009	30	3.57	
Conducted emissions				

9 Versions of test reports (change history)

Version	Applied changes	Date of release
	Initial release	2023-Mar-16

End Of Test Report

TR21-1-0179501T014a 18 / 18