

**Calibration Laboratory of**  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

#### References

- [1] ANSI-C63.19-2007  
American National Standard for Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids.
- [2] ANSI-C63.19-2011  
American National Standard, Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids.

#### Methods Applied and Interpretation of Parameters:

- *Coordinate System:* y-axis is in the direction of the dipole arms. z-axis is from the basis of the antenna (mounted on the table) towards its feed point between the two dipole arms. x-axis is normal to the other axes. In coincidence with the standards [1], the measurement planes (probe sensor center) are selected to be at a distance of 10 mm (15 mm for [2]) above the top metal edge of the dipole arms.
- *Measurement Conditions:* Further details are available from the hardcopies at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. The forward power to the dipole connector is set with a calibrated power meter connected and monitored with an auxiliary power meter connected to a directional coupler. While the dipole under test is connected, the forward power is adjusted to the same level.
- *Antenna Positioning:* The dipole is mounted on a HAC Test Arch phantom using the matching dipole positioner with the arms horizontal and the feeding cable coming from the floor. The measurements are performed in a shielded room with absorbers around the setup to reduce the reflections. It is verified before the mounting of the dipole under the Test Arch phantom, that its arms are perfectly in a line. It is installed on the HAC dipole positioner with its arms parallel below the dielectric reference wire and able to move elastically in vertical direction without changing its relative position to the top center of the Test Arch phantom. The vertical distance to the probe is adjusted after dipole mounting with a DASY5 Surface Check job. Before the measurement, the distance between phantom surface and probe tip is verified. The proper measurement distance is selected by choosing the matching section of the HAC Test Arch phantom with the proper device reference point (upper surface of the dipole) and the matching grid reference point (tip of the probe) considering the probe sensor offset. The vertical distance to the probe is essential for the accuracy.
- *Feed Point Impedance and Return Loss:* These parameters are measured using a HP 8753E Vector Network Analyzer. The impedance is specified at the SMA connector of the dipole. The influence of reflections was eliminated by applying the averaging function while moving the dipole in the air, at least 70cm away from any obstacles.
- *E-field distribution:* E field is measured in the x-y-plane with an isotropic ER3D-field probe with 100 mW forward power to the antenna feed point. In accordance with [1] and [2], the scan area is 20mm wide, its length exceeds the dipole arm length (180 or 90mm). The sensor center is 10 mm (15 mm for [2]) (in z) above the metal top of the dipole arms. Two 3D maxima are available near the end of the dipole arms. Assuming the dipole arms are perfectly in one line, the average of these two maxima (in subgrid 2 and subgrid 8) is determined to compensate for any non-parallelity to the measurement plane as well as the sensor displacement. The E-field value stated as calibration value represents the maximum of the interpolated 3D-E-field, in the plane above the dipole surface.
- *H-field distribution:* H-field is measured with an isotropic H-field probe with 100mW forward power to the antenna feed point, in the x-y-plane. The scan area and sensor distance is equivalent to the E-field scan. The maximum of the field is available at the center (subgrid 5) above the feed point. The H-field value stated as calibration value represents the maximum of the interpolated H-field, 10mm above the dipole surface at the feed point.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

**Measurement Conditions**

DASY system configuration, as far as not given on page 1.

|   |                      |          |
|---|----------------------|----------|
| <b>DASY Version</b>                       | DASY5                | V52.10.0 |
| <b>Phantom</b>                            | HAC Test Arch        |          |
| <b>Distance Dipole Top - Probe Center</b> | 10, 15 mm            |          |
| <b>Scan resolution</b>                    | dx, dy = 5 mm        |          |
| <b>Frequency</b>                          | 2450 MHz $\pm$ 1 MHz |          |
| <b>Input power drift</b>                  | < 0.05 dB            |          |

**Maximum Field values at 2450 MHz**

| <b>H-field 10 mm above dipole surface</b> | <b>condition</b>   | <b>interpolated maximum</b>  |
|---|--------------------|------------------------------|
| Maximum measured                          | 100 mW input power | 0.502 A/m $\pm$ 8.2 % (k=2)  |
| <b>E-field 10 mm above dipole surface</b> |                    |                              |
| Maximum measured above high end           | 100 mW input power | 138.6 V/m = 42.84 dBV/m      |
| Maximum measured above low end            | 100 mW input power | 135.6 V/m = 42.64 dBV/m      |
| Averaged maximum above arm                | 100 mW input power | 137.1 V/m $\pm$ 12.7 % (k=2) |
| <b>E-field 15 mm above dipole surface</b> | <b>condition</b>   | <b>Interpolated maximum</b>  |
| Maximum measured above high end           | 100 mW input power | 92.0 V/m = 39.28 dBV/m       |
| Maximum measured above low end            | 100 mW input power | 85.7 V/m = 38.66 dBV/m       |
| Averaged maximum above arm                | 100 mW input power | 88.9 V/m $\pm$ 12.8 % (k=2)  |

**Appendix (Additional assessments outside the scope of SCS 0108)****Antenna Parameters**

| Frequency | Return Loss | Impedance                   |
|-----------|-------------|-----------------------------|
| 2250 MHz  | 17.4 dB     | $64.8 \Omega + 4.2 j\Omega$ |
| 2350 MHz  | 29.6 dB     | $52.1 \Omega - 2.6 j\Omega$ |
| 2450 MHz  | 27.6 dB     | $54.2 \Omega - 1.0 j\Omega$ |
| 2550 MHz  | 31.5 dB     | $50.8 \Omega - 2.6 j\Omega$ |
| 2650 MHz  | 18.2 dB     | $61.4 \Omega - 7.5 j\Omega$ |

**3.2 Antenna Design and Handling**


The calibration dipole has a symmetric geometry with a built-in two stub matching network, which leads to the enhanced bandwidth.

The dipole is built of standard semirigid coaxial cable. The internal matching line is open ended. The antenna is therefore open for DC signals.

Do not apply force to dipole arms, as they are liable to bend. The soldered connections near the feedpoint may be damaged. After excessive mechanical stress or overheating, check the impedance characteristics to ensure that the internal matching network is not affected.

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Impedance Measurement Plot



**DASY5 H-field Result**

Date: 22.08.2017

Test Laboratory: SPEAG Lab2

**DUT: HAC Dipole 2450 MHz; Type: CD2450V3; Serial: CD2450V3 - SN: 1021**

Communication System: UID 0 - CW ; Frequency: 2450 MHz

 Medium parameters used:  $\sigma = 0$  S/m,  $\epsilon_r = 1$ ;  $\rho = 1$  kg/m<sup>3</sup>

Phantom section: RF Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

**DASY52 Configuration:**

- Probe: H3DV6 - SN6065; ; Calibrated: 30.12.2016
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn781; Calibrated: 13.07.2017
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA; Serial: 1070
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

**Dipole H-Field measurement @ 2450MHz/H-Scan - 2450MHz d=10mm/Hearing Aid Compatibility Test (41x181x1):** Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 0.5270 A/m; Power Drift = -0.01 dB


PMR not calibrated. PMF = 1.000 is applied.

H-field emissions = 0.5022 A/m

Near-field category: M2 (AWF 0 dB)

PMF scaled H-field

| Grid 1 M2        | Grid 2 M2        | Grid 3 M2        |
|------------------|------------------|------------------|
| <b>0.379 A/m</b> | <b>0.428 A/m</b> | <b>0.420 A/m</b> |
| Grid 4 M2        | Grid 5 M2        | Grid 6 M2        |
| <b>0.455 A/m</b> | <b>0.502 A/m</b> | <b>0.494 A/m</b> |
| Grid 7 M2        | Grid 8 M2        | Grid 9 M2        |
| <b>0.370 A/m</b> | <b>0.405 A/m</b> | <b>0.400 A/m</b> |



### DASY5 E-field Result

Date: 22.08.2017

Test Laboratory: SPEAG Lab2

**DUT: HAC Dipole 2450 MHz; Type: CD2450V3; Serial: CD2450V3 - SN: 1021**

Communication System: UID 0 - CW ; Frequency: 2450 MHz

Medium parameters used:  $\sigma = 0$  S/m,  $\epsilon_r = 1$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: RF Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ER3DV6 - SN2336; ConvF(1, 1, 1); Calibrated: 30.12.2016;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn781; Calibrated: 13.07.2017
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA; Serial: 1070
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

**Dipole E-Field measurement @ 2450MHz/E-Scan - 2450MHz d=10mm/Hearing Aid Compatibility Test (41x181x1):** Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 86.46 V/m; Power Drift = 0.00 dB

Applied MIF = 0.00 dB

RF audio interference level = 42.84 dBV/m

**Emission category: M1**

MIF scaled E-field

| Grid 1 M1<br>42.29 dBV/m | Grid 2 M1<br>42.84 dBV/m | Grid 3 M1<br>42.72 dBV/m |
|--------------------------|--------------------------|--------------------------|
| Grid 4 M1<br>41.16 dBV/m | Grid 5 M1<br>41.61 dBV/m | Grid 6 M1<br>41.37 dBV/m |
| Grid 7 M1<br>42.02 dBV/m | Grid 8 M1<br>42.64 dBV/m | Grid 9 M1<br>42.57 dBV/m |

**Dipole E-Field measurement @ 2450MHz/E-Scan - 2450MHz d=15mm/Hearing Aid Compatibility Test**  
(41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 86.21 V/m; Power Drift = 0.02 dB


Applied MIF = 0.00 dB

RF audio interference level = 39.28 dBV/m

**Emission category: M2**

MIF scaled E-field

| Grid 1 M2<br>38.99 dBV/m | Grid 2 M2<br>39.28 dBV/m | Grid 3 M2<br>39.2 dBV/m  |
|--------------------------|--------------------------|--------------------------|
| Grid 4 M2<br>38.53 dBV/m | Grid 5 M2<br>38.75 dBV/m | Grid 6 M2<br>38.64 dBV/m |
| Grid 7 M2<br>38.39 dBV/m | Grid 8 M2<br>38.66 dBV/m | Grid 9 M2<br>38.62 dBV/m |



## ANNEX F DAE CALIBRATION CERTIFICATE

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**C** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**Client **CTTL (Auden)**Certificate No: **DAE4-777\_Sep17**

### CALIBRATION CERTIFICATE

Object **DAE4 - SD 000 D04 BM - SN: 777**

Calibration procedure(s) **QA CAL-06.v29**  
Calibration procedure for the data acquisition electronics (DAE)


Calibration date: **September 08, 2017**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards             | ID #               | Cal Date (Certificate No.) | Scheduled Calibration  |
|-------------------------------|--------------------|----------------------------|------------------------|
| Keithley Multimeter Type 2001 | SN: 0810278        | 31-Aug-17 (No:21092)       | Aug-18                 |
| Secondary Standards           | ID #               | Check Date (in house)      | Scheduled Check        |
| Auto DAE Calibration Unit     | SE UWS 053 AA 1001 | 05-Jan-17 (in house check) | In house check: Jan-18 |
| Calibrator Box V2.1           | SE UMS 006 AA 1002 | 05-Jan-17 (in house check) | In house check: Jan-18 |

|                |                           |                                   |  |
|----------------|---------------------------|-----------------------------------|--|
| Calibrated by: | Name<br>Dominique Steffen | Function<br>Laboratory Technician | Signature<br> |
| Approved by:   | Sven Kühn                 | Deputy Manager                    |               |

Issued: September 8, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

**Calibration Laboratory of**  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

### Glossary

|                 |   |
|-----------------|---|
| DAE             | data acquisition electronics  |
| Connector angle | information used in DASY system to align probe sensor X to the robot coordinate system. |

### Methods Applied and Interpretation of Parameters

- *DC Voltage Measurement*: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle*: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
  - *DC Voltage Measurement Linearity*: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
  - *Common mode sensitivity*: Influence of a positive or negative common mode voltage on the differential measurement.
  - *Channel separation*: Influence of a voltage on the neighbor channels not subject to an input voltage.
  - *AD Converter Values with inputs shorted*: Values on the internal AD converter corresponding to zero input voltage
  - *Input Offset Measurement*: Output voltage and statistical results over a large number of zero voltage measurements.
  - *Input Offset Current*: Typical value for information; Maximum channel input offset current, not considering the input resistance.
  - *Input resistance*: Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
  - *Low Battery Alarm Voltage*: Typical value for information. Below this voltage, a battery alarm signal is generated.
  - *Power consumption*: Typical value for information. Supply currents in various operating modes.

**DC Voltage Measurement**

A/D - Converter Resolution nominal

High Range: 1LSB =  $6.1\mu\text{V}$  , full range = -100...+300 mVLow Range: 1LSB =  $61\text{nV}$  , full range = -1.....+3mV

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

| Calibration Factors | X                                  | Y                                  | Z                                  |
|---------------------|------------------------------------|------------------------------------|------------------------------------|
| High Range          | $405.400 \pm 0.02\% \text{ (k=2)}$ | $405.869 \pm 0.02\% \text{ (k=2)}$ | $405.579 \pm 0.02\% \text{ (k=2)}$ |
| Low Range           | $3.96640 \pm 1.50\% \text{ (k=2)}$ | $3.96264 \pm 1.50\% \text{ (k=2)}$ | $4.00499 \pm 1.50\% \text{ (k=2)}$ |

**Connector Angle**

|   |                          |
|---|--------------------------|
| Connector Angle to be used in DASY system | $97.0^\circ \pm 1^\circ$ |
|---|--------------------------|

**Appendix (Additional assessments outside the scope of SCS0108)**
**1. DC Voltage Linearity**

| High Range |         | Reading (µV) | Difference (µV) | Error (%) |
|------------|---------|--------------|-----------------|-----------|
| Channel X  | + Input | 200022.73    | -12.42          | -0.01     |
| Channel X  | + Input | 20003.49     | -1.25           | -0.01     |
| Channel X  | - Input | -19998.82    | 6.77            | -0.03     |
| Channel Y  | + Input | 200025.10    | -10.04          | -0.01     |
| Channel Y  | + Input | 20007.22     | 2.54            | 0.01      |
| Channel Y  | - Input | -20002.34    | 3.30            | -0.02     |
| Channel Z  | + Input | 200028.10    | -6.82           | -0.00     |
| Channel Z  | + Input | 20002.36     | -2.19           | -0.01     |
| Channel Z  | - Input | -20003.64    | 2.12            | -0.01     |

| Low Range |         | Reading (µV) | Difference (µV) | Error (%) |
|-----------|---------|--------------|-----------------|-----------|
| Channel X | + Input | 2000.54      | -0.37           | -0.02     |
| Channel X | + Input | 201.37       | 0.50            | 0.25      |
| Channel X | - Input | -199.19      | -0.20           | 0.10      |
| Channel Y | + Input | 1999.95      | -0.89           | -0.04     |
| Channel Y | + Input | 200.04       | -0.75           | -0.37     |
| Channel Y | - Input | -199.96      | -0.85           | 0.43      |
| Channel Z | + Input | 2001.05      | 0.20            | 0.01      |
| Channel Z | + Input | 199.88       | -0.86           | -0.43     |
| Channel Z | - Input | -200.02      | -0.88           | 0.44      |

**2. Common mode sensitivity**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

|           | Common mode<br>Input Voltage (mV) | High Range<br>Average Reading (µV) | Low Range<br>Average Reading (µV) |
|-----------|-----------------------------------|------------------------------------|-----------------------------------|
| Channel X | 200                               | 5.45                               | 3.79                              |
|           | -200                              | 3.93                               | 0.83                              |
| Channel Y | 200                               | 7.70                               | 7.39                              |
|           | -200                              | -9.52                              | -8.90                             |
| Channel Z | 200                               | 7.51                               | 6.49                              |
|           | -200                              | -9.21                              | -8.71                             |

**3. Channel separation**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

|           | Input Voltage (mV) | Channel X (µV) | Channel Y (µV) | Channel Z (µV) |
|-----------|--------------------|----------------|----------------|----------------|
| Channel X | 200                | -              | -1.61          | -2.84          |
| Channel Y | 200                | 8.30           | -              | 0.46           |
| Channel Z | 200                | 6.69           | 5.02           | -              |

**4. AD-Converter Values with inputs shorted**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

|           | High Range (LSB) | Low Range (LSB) |
|-----------|------------------|-----------------|
| Channel X | 15919            | 14652           |
| Channel Y | 16343            | 14477           |
| Channel Z | 16033            | 14911           |

**5. Input Offset Measurement**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input 10MΩ

|           | Average (µV) | min. Offset (µV) | max. Offset (µV) | Std. Deviation (µV) |
|-----------|--------------|------------------|------------------|---------------------|
| Channel X | -0.50        | -2.04            | 0.95             | 0.51                |
| Channel Y | 1.56         | 0.40             | 2.80             | 0.48                |
| Channel Z | 0.26         | -0.78            | 1.16             | 0.42                |

**6. Input Offset Current**

Nominal Input circuitry offset current on all channels: &lt;25fA

**7. Input Resistance (Typical values for information)**

|           | Zeroing (kOhm) | Measuring (MΩ) |
|-----------|----------------|----------------|
| Channel X | 200            | 200            |
| Channel Y | 200            | 200            |
| Channel Z | 200            | 200            |

**8. Low Battery Alarm Voltage (Typical values for information)**

| Typical values | Alarm Level (VDC) |
|----------------|-------------------|
| Supply (+ Vcc) | +7.9              |
| Supply (- Vcc) | -7.6              |

**9. Power Consumption (Typical values for information)**

| Typical values | Switched off (mA) | Stand by (mA) | Transmitting (mA) |
|----------------|-------------------|---------------|-------------------|
| Supply (+ Vcc) | +0.01             | +6            | +14               |
| Supply (- Vcc) | -0.01             | -8            | -9                |

**The photos of HAC test are presented in the additional document:**

Appendix to test report No.I18Z60544-SEM02/03

The photos of HAC test