



Impedance Plot for D2450V2-SN: 968 2450 Head

Calibrated impedance: $51.9\Omega + 3.63j\Omega$; Measurement impedance: $50.4\Omega - 3.7j\Omega$ (within 5Ω)



Date: 21.FEB.2024 19:04:46

Calibrated return loss: -27.9dB; Measurement return loss: -24.44dB (within 20%)



Date: 21.FEB.2024 19:04:29