



In Collaboration with

**s p e a g**  
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504  
E-mail: cntl@chinattl.com http://www.chinattl.cn

## DASY5 Validation Report for Head TSL

Date: 03.26.2019

Test Laboratory: CTTL, Beijing, China

**DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 5d170**

Communication System: UID 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium parameters used:  $f = 1900$  MHz;  $\sigma = 1.441$  S/m;  $\epsilon_r = 40.48$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Center Section

### DASY5 Configuration:

- Probe: EX3DV4 - SN3617; ConvF(8.14, 8.14, 8.14) @ 1900 MHz; Calibrated: 1/31/2019
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 2/6/2019
- Phantom: MFP\_V5.1C; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (2); SEMCAD X Version 14.6.12 (7450)


**System Performance Check/Zoom Scan (7x7x7) (7x7x7)/Cube 0:** Measurement grid:  
 $dx=5$ mm,  $dy=5$ mm,  $dz=5$ mm

Reference Value = 97.54 V/m; Power Drift = 0.08 dB

Peak SAR (extrapolated) = 18.9 W/kg

**SAR(1 g) = 9.9 W/kg; SAR(10 g) = 5.12 W/kg**

Maximum value of SAR (measured) = 15.6 W/kg



0 dB = 15.6 W/kg = 11.93 dBW/kg




In Collaboration with

**s p e a g**  
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504  
E-mail: cttl@chinattl.com http://www.chinattl.cn

### Impedance Measurement Plot for Head TSL





In Collaboration with

**s p e a g**  
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504  
E-mail: cttl@chinattl.com http://www.chinattl.cn

## DASY5 Validation Report for Body TSL

Date: 03.26.2019

Test Laboratory: CTTL, Beijing, China

**DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 5d170**

Communication System: UID 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium parameters used:  $f = 1900$  MHz;  $\sigma = 1.56$  S/m;  $\epsilon_r = 54.52$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Right Section

DASY5 Configuration:

- Probe: EX3DV4 - SN3617; ConvF(7.78, 7.78, 7.78) @ 1900 MHz; Calibrated: 1/31/2019
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 2/6/2019
- Phantom: MFP\_V5.1C ; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (2); SEMCAD X Version 14.6.12 (7450)

**System Performance Check/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid:**


$dx=5$ mm,  $dy=5$ mm,  $dz=5$ mm

Reference Value = 95.48 V/m; Power Drift = 0.04 dB

Peak SAR (extrapolated) = 18.6 W/kg

SAR(1 g) = 10.1 W/kg; SAR(10 g) = 5.28 W/kg

Maximum value of SAR (measured) = 15.7 W/kg





In Collaboration with

**s p e a k**

**CALIBRATION LABORATORY**

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China


Tel: +86-10-62304633-2079

Fax: +86-10-62304633-2504

E-mail: cttl@chinattl.com

http://www.chinattl.cn

## Impedance Measurement Plot for Body TSL





## D1900V2, Serial No. 5d170 Extended Dipole Calibrations

Referring to KDB 865664 D01 v01r02, if dipoles are verified in return loss ( $< -20\text{dB}$ , within 20% of prior calibration), and in impedance (within 5 ohm of prior calibration), the annual calibration is not necessary and the calibration interval can be extended.

1900V2 – serial no. 5d170


| 1900V2 – serial no. 5d170 |                  |           |                      |             |                           |             |                  |           |                      |             |                           |             |
|---------------------------|------------------|-----------|----------------------|-------------|---------------------------|-------------|------------------|-----------|----------------------|-------------|---------------------------|-------------|
| 1900 Head                 |                  |           |                      |             |                           |             | 1900 Body        |           |                      |             |                           |             |
| Date of Measurement       | Return-Loss (dB) | Delta (%) | Real Impedance (ohm) | Delta (ohm) | Imaginary Impedance (ohm) | Delta (ohm) | Return-Loss (dB) | Delta (%) | Real Impedance (ohm) | Delta (ohm) | Imaginary Impedance (ohm) | Delta (ohm) |
| 2019.3.26                 | -23.3            |           | 51.7                 |             | 6.7                       |             | -22.8            |           | 47.8                 |             | 6.7                       |             |
| 2020.3.25                 | -22.3            | 0.05      | 53.0                 | -1.26       | 7.4                       | -0.64       | -22.5            | 0.01      | 49.2                 | -1.37       | 7.41                      | -0.69       |

### <Justification of the extended calibration>


The return loss is  $< -20\text{dB}$ , within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

## Dipole Verification Data> D1900V2, serial no. 5d170

### 1900MHz – Head



### 1900MHz – Body







Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **Sporton**

Certificate No: **DAE4-1358\_Apr20**

## CALIBRATION CERTIFICATE

Object **DAE4 - SD 000 D04 BN - SN: 1358**

Calibration procedure(s) **QA CAL-06.v30**  
**Calibration procedure for the data acquisition electronics (DAE)**

Calibration date: **April 28, 2020**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ )°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards             | ID #        | Cal Date (Certificate No.) | Scheduled Calibration |
|-------------------------------|-------------|----------------------------|-----------------------|
| Keithley Multimeter Type 2001 | SN: 0810278 | 03-Sep-19 (No:25949)       | Sep-20                |

| Secondary Standards       | ID #               | Check Date (in house)      | Scheduled Check        |
|---------------------------|--------------------|----------------------------|------------------------|
| Auto DAE Calibration Unit | SE UWS 053 AA 1001 | 09-Jan-20 (in house check) | In house check: Jan-21 |
| Calibrator Box V2.1       | SE UMS 006 AA 1002 | 09-Jan-20 (in house check) | In house check: Jan-21 |

|                | Name          | Function              |
|----------------|---------------|-----------------------|
| Calibrated by: | Eric Hainfeld | Laboratory Technician |

|              |           |                |
|--------------|-----------|----------------|
| Approved by: | Sven Kühn | Deputy Manager |
|--------------|-----------|----------------|

Signature

Issued: April 29, 2020

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

## Glossary

|                 |                                                                                         |
|-----------------|-----------------------------------------------------------------------------------------|
| DAE             | data acquisition electronics                                                            |
| Connector angle | information used in DASY system to align probe sensor X to the robot coordinate system. |

## Methods Applied and Interpretation of Parameters

- *DC Voltage Measurement:* Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle:* The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
  - *DC Voltage Measurement Linearity:* Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
  - *Common mode sensitivity:* Influence of a positive or negative common mode voltage on the differential measurement.
  - *Channel separation:* Influence of a voltage on the neighbor channels not subject to an input voltage.
  - *AD Converter Values with inputs shorted:* Values on the internal AD converter corresponding to zero input voltage
  - *Input Offset Measurement:* Output voltage and statistical results over a large number of zero voltage measurements.
  - *Input Offset Current:* Typical value for information; Maximum channel input offset current, not considering the input resistance.
  - *Input resistance:* Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
  - *Low Battery Alarm Voltage:* Typical value for information. Below this voltage, a battery alarm signal is generated.
  - *Power consumption:* Typical value for information. Supply currents in various operating modes.



## DC Voltage Measurement

A/D - Converter Resolution nominal

High Range: 1LSB = 6.1 $\mu$ V , full range = -100...+300 mV

Low Range: 1LSB = 61nV , full range = -1.....+3mV

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

| Calibration Factors | X                         | Y                         | Z                         |
|---------------------|---------------------------|---------------------------|---------------------------|
| High Range          | 403.411 $\pm$ 0.02% (k=2) | 403.452 $\pm$ 0.02% (k=2) | 403.463 $\pm$ 0.02% (k=2) |
| Low Range           | 3.96158 $\pm$ 1.50% (k=2) | 3.98747 $\pm$ 1.50% (k=2) | 3.99174 $\pm$ 1.50% (k=2) |

## Connector Angle

|                                           |                   |
|-------------------------------------------|-------------------|
| Connector Angle to be used in DASY system | 113.5 ° $\pm$ 1 ° |
|-------------------------------------------|-------------------|

## Appendix (Additional assessments outside the scope of SCS0108)

### 1. DC Voltage Linearity

| High Range        | Reading ( $\mu\text{V}$ ) | Difference ( $\mu\text{V}$ ) | Error (%) |
|-------------------|---------------------------|------------------------------|-----------|
| Channel X + Input | 200024.85                 | -8.32                        | -0.00     |
| Channel X + Input | 20005.36                  | 0.39                         | 0.00      |
| Channel X - Input | -20003.50                 | 2.72                         | -0.01     |
| Channel Y + Input | 200030.06                 | -2.90                        | -0.00     |
| Channel Y + Input | 20004.14                  | -0.70                        | -0.00     |
| Channel Y - Input | -20008.00                 | -1.63                        | 0.01      |
| Channel Z + Input | 200034.52                 | 1.89                         | 0.00      |
| Channel Z + Input | 20005.02                  | 0.16                         | 0.00      |
| Channel Z - Input | -20007.28                 | -0.87                        | 0.00      |

| Low Range         | Reading ( $\mu\text{V}$ ) | Difference ( $\mu\text{V}$ ) | Error (%) |
|-------------------|---------------------------|------------------------------|-----------|
| Channel X + Input | 2000.94                   | 0.03                         | 0.00      |
| Channel X + Input | 200.94                    | 0.01                         | 0.01      |
| Channel X - Input | -198.93                   | 0.16                         | -0.08     |
| Channel Y + Input | 2000.58                   | -0.17                        | -0.01     |
| Channel Y + Input | 199.97                    | -0.81                        | -0.40     |
| Channel Y - Input | -200.24                   | -0.99                        | 0.50      |
| Channel Z + Input | 2000.83                   | 0.21                         | 0.01      |
| Channel Z + Input | 199.97                    | -0.67                        | -0.34     |
| Channel Z - Input | -199.90                   | -0.63                        | 0.32      |

### 2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

|           | Common mode Input Voltage (mV) | High Range Average Reading ( $\mu\text{V}$ ) | Low Range Average Reading ( $\mu\text{V}$ ) |
|-----------|--------------------------------|----------------------------------------------|---------------------------------------------|
| Channel X | 200                            | 23.26                                        | 21.16                                       |
|           | - 200                          | -21.29                                       | -22.70                                      |
| Channel Y | 200                            | -27.83                                       | -28.04                                      |
|           | - 200                          | 26.48                                        | 26.49                                       |
| Channel Z | 200                            | -11.47                                       | -11.06                                      |
|           | - 200                          | 9.80                                         | 9.70                                        |

### 3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

|           | Input Voltage (mV) | Channel X ( $\mu\text{V}$ ) | Channel Y ( $\mu\text{V}$ ) | Channel Z ( $\mu\text{V}$ ) |
|-----------|--------------------|-----------------------------|-----------------------------|-----------------------------|
| Channel X | 200                | -                           | 1.92                        | -3.40                       |
| Channel Y | 200                | 8.27                        | -                           | 3.32                        |
| Channel Z | 200                | 9.47                        | 5.42                        | -                           |

#### 4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

|           | High Range (LSB) | Low Range (LSB) |
|-----------|------------------|-----------------|
| Channel X | 15579            | 16774           |
| Channel Y | 16044            | 14871           |
| Channel Z | 16074            | 16518           |

#### 5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input 10M $\Omega$

|           | Average ( $\mu$ V) | min. Offset ( $\mu$ V) | max. Offset ( $\mu$ V) | Std. Deviation ( $\mu$ V) |
|-----------|--------------------|------------------------|------------------------|---------------------------|
| Channel X | 0.87               | -0.93                  | 1.98                   | 0.46                      |
| Channel Y | -0.62              | -1.71                  | 0.15                   | 0.38                      |
| Channel Z | -0.46              | -1.45                  | 0.52                   | 0.39                      |

#### 6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

#### 7. Input Resistance (Typical values for information)

|           | Zeroing (kOhm) | Measuring (MOhm) |
|-----------|----------------|------------------|
| Channel X | 200            | 200              |
| Channel Y | 200            | 200              |
| Channel Z | 200            | 200              |

#### 8. Low Battery Alarm Voltage (Typical values for information)

| Typical values | Alarm Level (VDC) |
|----------------|-------------------|
| Supply (+ Vcc) | +7.9              |
| Supply (- Vcc) | -7.6              |

#### 9. Power Consumption (Typical values for information)

| Typical values | Switched off (mA) | Stand by (mA) | Transmitting (mA) |
|----------------|-------------------|---------------|-------------------|
| Supply (+ Vcc) | +0.01             | +6            | +14               |
| Supply (- Vcc) | -0.01             | -8            | -9                |



Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **Sporton**

Certificate No: **EX3-3935\_May20**

## CALIBRATION CERTIFICATE

Object **EX3DV4 - SN:3935**

Calibration procedure(s) **QA CAL-01.v9, QA CAL-14.v5, QA CAL-23.v5, QA CAL-25.v7**  
**Calibration procedure for dosimetric E-field probes**

Calibration date: **May 27, 2020**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ )°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID               | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|------------------|-----------------------------------|------------------------|
| Power meter NRP            | SN: 104778       | 01-Apr-20 (No. 217-03100/03101)   | Apr-21                 |
| Power sensor NRP-Z91       | SN: 103244       | 01-Apr-20 (No. 217-03100)         | Apr-21                 |
| Power sensor NRP-Z91       | SN: 103245       | 01-Apr-20 (No. 217-03101)         | Apr-21                 |
| Reference 20 dB Attenuator | SN: CC2552 (20x) | 31-Mar-20 (No. 217-03106)         | Apr-21                 |
| DAE4                       | SN: 660          | 27-Dec-19 (No. DAE4-660_Dec19)    | Dec-20                 |
| Reference Probe ES3DV2     | SN: 3013         | 31-Dec-19 (No. ES3-3013_Dec19)    | Dec-20                 |
| Secondary Standards        | ID               | Check Date (in house)             | Scheduled Check        |
| Power meter E4419B         | SN: GB41293874   | 06-Apr-16 (in house check Jun-18) | In house check: Jun-20 |
| Power sensor E4412A        | SN: MY41498087   | 06-Apr-16 (in house check Jun-18) | In house check: Jun-20 |
| Power sensor E4412A        | SN: 000110210    | 06-Apr-16 (in house check Jun-18) | In house check: Jun-20 |
| RF generator HP 8648C      | SN: US3642U01700 | 04-Aug-99 (in house check Jun-18) | In house check: Jun-20 |
| Network Analyzer E8358A    | SN: US41080477   | 31-Mar-14 (in house check Oct-19) | In house check: Oct-20 |

|                                                                                                                 | Name          | Function              | Signature            |
|-----------------------------------------------------------------------------------------------------------------|---------------|-----------------------|----------------------|
| Calibrated by:                                                                                                  | Leif Klysner  | Laboratory Technician |                      |
| Approved by:                                                                                                    | Katja Pokovic | Technical Manager     |                      |
|                                                                                                                 |               |                       | Issued: June 1, 2020 |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. |               |                       |                      |





Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: **SCS 0108**

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

## Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\phi$      | $\phi$ rotation around probe axis                                                                                                                       |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

## Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the  $E^2$ -field uncertainty inside TSL (see below *ConvF*).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for *ConvF*. A frequency dependent *ConvF* is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).



# DASY/EASY - Parameters of Probe: EX3DV4 - SN:3935

## Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc (k=2) |
|-----------------------------------------------------------|----------|----------|----------|-----------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.49     | 0.53     | 0.48     | ± 10.1 %  |
| DCP (mV) <sup>B</sup>                                     | 102.6    | 103.2    | 102.3    |           |

## Calibration Results for Modulation Response

| UID | Communication System Name |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C   | D<br>dB | VR<br>mV | Max<br>dev. | Unc <sup>E</sup><br>(k=2) |
|-----|---------------------------|---|---------|------------------------------|-----|---------|----------|-------------|---------------------------|
| 0   | CW                        | X | 0.0     | 0.0                          | 1.0 | 0.00    | 137.0    | ± 3.3 %     | ± 4.7 %                   |
|     |                           | Y | 0.0     | 0.0                          | 1.0 |         | 150.1    |             |                           |
|     |                           | Z | 0.0     | 0.0                          | 1.0 |         | 141.6    |             |                           |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Page 5).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3935

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | 44.5       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3935

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-----------|
| 750                  | 41.9                               | 0.89                            | 10.58   | 10.58   | 10.58   | 0.57               | 0.80                    | ± 12.0 %  |
| 835                  | 41.5                               | 0.90                            | 10.31   | 10.31   | 10.31   | 0.38               | 0.93                    | ± 12.0 %  |
| 900                  | 41.5                               | 0.97                            | 10.16   | 10.16   | 10.16   | 0.40               | 0.88                    | ± 12.0 %  |
| 1750                 | 40.1                               | 1.37                            | 8.60    | 8.60    | 8.60    | 0.27               | 0.86                    | ± 12.0 %  |
| 1900                 | 40.0                               | 1.40                            | 8.35    | 8.35    | 8.35    | 0.24               | 0.86                    | ± 12.0 %  |
| 2000                 | 40.0                               | 1.40                            | 8.25    | 8.25    | 8.25    | 0.34               | 0.86                    | ± 12.0 %  |
| 2300                 | 39.5                               | 1.67                            | 7.86    | 7.86    | 7.86    | 0.35               | 0.90                    | ± 12.0 %  |
| 2450                 | 39.2                               | 1.80                            | 7.60    | 7.60    | 7.60    | 0.33               | 0.90                    | ± 12.0 %  |
| 2600                 | 39.0                               | 1.96                            | 7.43    | 7.43    | 7.43    | 0.37               | 0.90                    | ± 12.0 %  |
| 5250                 | 35.9                               | 4.71                            | 5.04    | 5.04    | 5.04    | 0.40               | 1.80                    | ± 14.0 %  |
| 5600                 | 35.5                               | 5.07                            | 4.76    | 4.76    | 4.76    | 0.40               | 1.80                    | ± 14.0 %  |
| 5750                 | 35.4                               | 5.22                            | 4.67    | 4.67    | 4.67    | 0.40               | 1.80                    | ± 14.0 %  |


<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4-9 MHz, and ConvF assessed at 13 MHz is 9-19 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies up to 6 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide: R22)




Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )





## Receiving Pattern ( $\phi$ ), $\vartheta = 0^\circ$

f=600 MHz,TEM




f=1800 MHz,R22




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

## Dynamic Range $f(\text{SAR}_{\text{head}})$ (TEM cell, $f_{\text{eval}} = 1900 \text{ MHz}$ )




Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \vartheta$ ),  $f = 900 \text{ MHz}$





---

## ***Appendix E. Conducted RF Output Power Table***

The detailed power table are shown as follows.



| LTE Band 2      |            |         |           |                       |                       |                        |                     |          |  |
|-----------------|------------|---------|-----------|-----------------------|-----------------------|------------------------|---------------------|----------|--|
| BW [MHz]        | Modulation | RB Size | RB Offset | Power Low Ch. / Freq. | Power Med Ch. / Freq. | Power High Ch. / Freq. | Tune-up limit (dBm) | MPR (dB) |  |
| Channel         |            |         |           | 18700                 | 18800                 | 19100                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1860                  | 1880                  | 1900                   |                     |          |  |
| 20              | QPSK       | 1       | 0         | 22.65                 | 22.89                 | 22.68                  | 24                  | 0        |  |
| 20              | QPSK       | 1       | 49        | 22.85                 | 22.56                 | 22.82                  |                     |          |  |
| 20              | QPSK       | 1       | 99        | 22.80                 | 22.55                 | 22.56                  |                     |          |  |
| 20              | QPSK       | 50      | 0         | 21.52                 | 21.67                 | 21.58                  | 23                  | 1        |  |
| 20              | QPSK       | 50      | 24        | 21.42                 | 21.53                 | 21.64                  |                     |          |  |
| 20              | QPSK       | 50      | 50        | 21.46                 | 21.32                 | 21.66                  |                     |          |  |
| 20              | QPSK       | 100     | 0         | 21.18                 | 21.64                 | 21.62                  | 23                  | 1        |  |
| 20              | 16QAM      | 1       | 0         | 21.35                 | 21.30                 | 21.09                  |                     |          |  |
| 20              | 16QAM      | 1       | 49        | 21.31                 | 21.29                 | 21.32                  |                     |          |  |
| 20              | 16QAM      | 1       | 99        | 21.20                 | 21.32                 | 21.12                  | 22                  | 2        |  |
| 20              | 16QAM      | 50      | 0         | 20.52                 | 20.63                 | 20.72                  |                     |          |  |
| 20              | 16QAM      | 50      | 24        | 20.56                 | 20.68                 | 20.75                  |                     |          |  |
| 20              | 16QAM      | 50      | 50        | 20.59                 | 20.47                 | 20.64                  | Tune-up limit (dBm) | MPR (dB) |  |
| 20              | 16QAM      | 100     | 0         | 20.46                 | 20.62                 | 20.75                  |                     |          |  |
| Channel         |            |         |           | 18675                 | 18800                 | 19125                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1857.5                | 1880                  | 1902.5                 |                     |          |  |
| 15              | QPSK       | 1       | 0         | 22.45                 | 22.69                 | 22.43                  | 24                  | 0        |  |
| 15              | QPSK       | 1       | 37        | 22.39                 | 22.30                 | 22.36                  |                     |          |  |
| 15              | QPSK       | 1       | 74        | 22.21                 | 22.21                 | 22.50                  |                     |          |  |
| 15              | QPSK       | 36      | 0         | 21.67                 | 21.75                 | 21.82                  | 23                  | 1        |  |
| 15              | QPSK       | 36      | 20        | 21.58                 | 21.83                 | 21.81                  |                     |          |  |
| 15              | QPSK       | 36      | 39        | 21.53                 | 21.61                 | 21.23                  |                     |          |  |
| 15              | QPSK       | 75      | 0         | 21.55                 | 21.82                 | 21.78                  | 23                  | 1        |  |
| 15              | 16QAM      | 1       | 0         | 21.45                 | 21.63                 | 21.72                  |                     |          |  |
| 15              | 16QAM      | 1       | 37        | 21.48                 | 21.49                 | 21.49                  |                     |          |  |
| 15              | 16QAM      | 1       | 74        | 21.50                 | 21.40                 | 21.63                  | 22                  | 2        |  |
| 15              | 16QAM      | 36      | 0         | 20.69                 | 20.66                 | 20.59                  |                     |          |  |
| 15              | 16QAM      | 36      | 20        | 20.87                 | 20.65                 | 20.66                  |                     |          |  |
| 15              | 16QAM      | 36      | 39        | 20.84                 | 20.55                 | 20.62                  | Tune-up limit (dBm) | MPR (dB) |  |
| 15              | 16QAM      | 75      | 0         | 20.77                 | 20.66                 | 20.57                  |                     |          |  |
| Channel         |            |         |           | 18650                 | 18800                 | 19150                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1855                  | 1880                  | 1905                   |                     |          |  |
| 10              | QPSK       | 1       | 0         | 22.56                 | 22.76                 | 22.87                  | 24                  | 0        |  |
| 10              | QPSK       | 1       | 25        | 22.57                 | 22.83                 | 22.56                  |                     |          |  |
| 10              | QPSK       | 1       | 49        | 22.58                 | 22.71                 | 22.58                  |                     |          |  |
| 10              | QPSK       | 25      | 0         | 21.40                 | 21.52                 | 21.74                  | 23                  | 1        |  |
| 10              | QPSK       | 25      | 12        | 21.52                 | 21.62                 | 21.70                  |                     |          |  |
| 10              | QPSK       | 25      | 25        | 21.43                 | 21.70                 | 21.75                  |                     |          |  |
| 10              | QPSK       | 50      | 0         | 21.34                 | 21.71                 | 21.70                  | 23                  | 1        |  |
| 10              | 16QAM      | 1       | 0         | 21.45                 | 21.43                 | 21.49                  |                     |          |  |
| 10              | 16QAM      | 1       | 25        | 21.52                 | 21.53                 | 21.64                  |                     |          |  |
| 10              | 16QAM      | 1       | 49        | 21.24                 | 21.48                 | 21.45                  | 22                  | 2        |  |
| 10              | 16QAM      | 25      | 0         | 20.38                 | 20.71                 | 20.75                  |                     |          |  |
| 10              | 16QAM      | 25      | 12        | 20.50                 | 20.69                 | 20.72                  |                     |          |  |
| 10              | 16QAM      | 25      | 25        | 20.45                 | 20.62                 | 20.65                  | Tune-up limit (dBm) | MPR (dB) |  |
| 10              | 16QAM      | 50      | 0         | 20.39                 | 20.73                 | 20.74                  |                     |          |  |
| Channel         |            |         |           | 18625                 | 18800                 | 19175                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1852.5                | 1880                  | 1907.5                 |                     |          |  |
| 5               | QPSK       | 1       | 0         | 22.64                 | 22.47                 | 22.58                  | 24                  | 0        |  |
| 5               | QPSK       | 1       | 12        | 22.66                 | 22.58                 | 22.65                  |                     |          |  |
| 5               | QPSK       | 1       | 24        | 22.67                 | 22.67                 | 22.77                  |                     |          |  |
| 5               | QPSK       | 12      | 0         | 21.66                 | 21.46                 | 21.45                  | 23                  | 1        |  |
| 5               | QPSK       | 12      | 7         | 21.65                 | 21.57                 | 21.49                  |                     |          |  |
| 5               | QPSK       | 12      | 13        | 21.64                 | 21.58                 | 21.51                  |                     |          |  |
| 5               | QPSK       | 25      | 0         | 21.55                 | 21.54                 | 21.58                  | 23                  | 1        |  |
| 5               | 16QAM      | 1       | 0         | 21.67                 | 21.81                 | 21.71                  |                     |          |  |
| 5               | 16QAM      | 1       | 12        | 21.61                 | 21.82                 | 21.64                  |                     |          |  |
| 5               | 16QAM      | 1       | 24        | 21.57                 | 21.54                 | 21.87                  | 22                  | 2        |  |
| 5               | 16QAM      | 12      | 0         | 20.98                 | 20.98                 | 21.03                  |                     |          |  |
| 5               | 16QAM      | 12      | 7         | 21.14                 | 20.90                 | 20.85                  |                     |          |  |
| 5               | 16QAM      | 12      | 13        | 21.22                 | 20.88                 | 20.82                  | Tune-up limit (dBm) | MPR (dB) |  |
| 5               | 16QAM      | 25      | 0         | 21.11                 | 21.01                 | 21.21                  |                     |          |  |
| Channel         |            |         |           | 18615                 | 18800                 | 19185                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1851.5                | 1880                  | 1908.5                 |                     |          |  |
| 3               | QPSK       | 1       | 0         | 22.67                 | 22.60                 | 22.72                  | 24                  | 0        |  |
| 3               | QPSK       | 1       | 8         | 22.88                 | 22.86                 | 22.84                  |                     |          |  |
| 3               | QPSK       | 1       | 14        | 22.88                 | 22.88                 | 22.87                  |                     |          |  |
| 3               | QPSK       | 8       | 0         | 21.57                 | 21.60                 | 21.62                  | 23                  | 1        |  |
| 3               | QPSK       | 8       | 4         | 21.61                 | 21.56                 | 21.68                  |                     |          |  |
| 3               | QPSK       | 8       | 7         | 21.62                 | 21.51                 | 21.63                  |                     |          |  |
| 3               | QPSK       | 15      | 0         | 21.60                 | 21.52                 | 21.64                  | 23                  | 1        |  |
| 3               | 16QAM      | 1       | 0         | 21.66                 | 21.50                 | 21.62                  |                     |          |  |
| 3               | 16QAM      | 1       | 8         | 21.55                 | 21.66                 | 21.65                  |                     |          |  |
| 3               | 16QAM      | 1       | 14        | 21.54                 | 21.64                 | 21.45                  | 22                  | 2        |  |
| 3               | 16QAM      | 8       | 0         | 20.44                 | 20.57                 | 20.69                  |                     |          |  |
| 3               | 16QAM      | 8       | 4         | 20.46                 | 20.72                 | 20.84                  |                     |          |  |
| 3               | 16QAM      | 8       | 7         | 20.45                 | 20.49                 | 20.61                  | Tune-up limit (dBm) | MPR (dB) |  |
| 3               | 16QAM      | 15      | 0         | 20.77                 | 20.86                 | 20.53                  |                     |          |  |
| Channel         |            |         |           | 18607                 | 18800                 | 19193                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1850.7                | 1880                  | 1909.3                 |                     |          |  |
| 1.4             | QPSK       | 1       | 0         | 22.88                 | 22.52                 | 22.65                  | 24                  | 0        |  |
| 1.4             | QPSK       | 1       | 3         | 22.53                 | 22.66                 | 22.67                  |                     |          |  |
| 1.4             | QPSK       | 1       | 5         | 22.87                 | 22.57                 | 22.58                  |                     |          |  |
| 1.4             | QPSK       | 3       | 0         | 22.66                 | 22.66                 | 22.59                  | 23                  | 1        |  |
| 1.4             | QPSK       | 3       | 1         | 22.43                 | 22.60                 | 22.54                  |                     |          |  |
| 1.4             | QPSK       | 3       | 3         | 22.64                 | 22.64                 | 22.51                  |                     |          |  |
| 1.4             | QPSK       | 6       | 0         | 21.84                 | 21.54                 | 21.47                  | 23                  | 1        |  |
| 1.4             | 16QAM      | 1       | 0         | 21.65                 | 21.55                 | 21.66                  |                     |          |  |
| 1.4             | 16QAM      | 1       | 3         | 21.55                 | 21.56                 | 21.65                  |                     |          |  |
| 1.4             | 16QAM      | 1       | 5         | 21.66                 | 21.65                 | 21.74                  | 23                  | 1        |  |
| 1.4             | 16QAM      | 3       | 0         | 21.65                 | 21.45                 | 21.65                  |                     |          |  |
| 1.4             | 16QAM      | 3       | 1         | 21.63                 | 21.60                 | 21.65                  |                     |          |  |
| 1.4             | 16QAM      | 3       | 3         | 21.66                 | 21.66                 | 21.45                  | 22                  | 2        |  |
| 1.4             | 16QAM      | 6       | 0         | 20.66                 | 20.65                 | 20.64                  |                     |          |  |

| LTE Band 4      |            |         |           |                       |                       |                        |                     |          |  |
|-----------------|------------|---------|-----------|-----------------------|-----------------------|------------------------|---------------------|----------|--|
| BW [MHz]        | Modulation | RB Size | RB Offset | Power Low Ch. / Freq. | Power Med Ch. / Freq. | Power High Ch. / Freq. | Tune-up limit (dBm) | MPR (dB) |  |
| Channel         |            |         |           | 20050                 | 20175                 | 20300                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1720                  | 1732.5                | 1745                   |                     |          |  |
| 20              | QPSK       | 1       | 0         | 22.44                 | 22.52                 | 22.51                  | 24                  | 0        |  |
| 20              | QPSK       | 1       | 49        | 22.31                 | 22.36                 | 22.39                  |                     |          |  |
| 20              | QPSK       | 1       | 99        | 22.18                 | 22.21                 | 22.28                  |                     |          |  |
| 20              | QPSK       | 50      | 0         | 21.21                 | 21.25                 | 21.33                  | 23                  | 1        |  |
| 20              | QPSK       | 50      | 24        | 21.24                 | 21.19                 | 21.24                  |                     |          |  |
| 20              | QPSK       | 50      | 50        | 21.22                 | 21.12                 | 21.14                  |                     |          |  |
| 20              | QPSK       | 100     | 0         | 21.19                 | 21.09                 | 21.18                  | 23                  | 1        |  |
| 20              | 16QAM      | 1       | 0         | 21.38                 | 21.32                 | 21.48                  |                     |          |  |
| 20              | 16QAM      | 1       | 49        | 21.42                 | 21.32                 | 21.48                  |                     |          |  |
| 20              | 16QAM      | 1       | 99        | 21.36                 | 21.28                 | 21.42                  | 22                  | 2        |  |
| 20              | 16QAM      | 50      | 0         | 20.22                 | 20.23                 | 20.32                  |                     |          |  |
| 20              | 16QAM      | 50      | 24        | 20.32                 | 20.15                 | 20.17                  |                     |          |  |
| 20              | 16QAM      | 50      | 50        | 20.25                 | 20.16                 | 20.21                  | Tune-up limit (dBm) | MPR (dB) |  |
| 20              | 16QAM      | 100     | 0         | 20.14                 | 20.19                 | 20.29                  |                     |          |  |
| Channel         |            |         |           | 20025                 | 20175                 | 20325                  |                     |          |  |
| Frequency (MHz) |            |         |           | 1717.5                | 1732.5                | 1747.5                 |                     |          |  |
| 15              | QPSK       | 1       | 0         | 22.61                 | 22.37                 | 22.31                  | 24                  | 0        |  |
| 15              | QPSK       | 1       | 37        | 22.31                 | 22.42                 | 22.35                  |                     |          |  |
| 15              | QPSK       | 1       | 74        | 22.36                 | 22.28                 | 22.19                  |                     |          |  |
| 15              | QPSK       | 36      | 0         | 21.12                 | 21.18                 | 21.32                  | 23                  | 1        |  |
| 15              | QPSK       | 36      | 20        | 21.19                 | 21.12                 | 21.15                  |                     |          |  |
| 15              | QPSK       | 36      | 39        | 21.14                 | 21.08                 | 21.15                  |                     |          |  |
| 15              | QPSK       | 75      | 0         | 21.15                 | 21.06                 | 21.14                  | 23                  | 1        |  |
| 15              | 16QAM      | 1       | 0         | 21.32                 | 21.36                 | 21.45                  |                     |          |  |
| 15              | 16QAM      | 1       | 37        | 21.29                 | 21.34                 | 21.45                  |                     |          |  |
| 15              | 16QAM      | 1       | 74        | 21.35                 | 21.26                 | 21.36                  | 22                  | 2        |  |
| 15              | 16QAM      | 36      | 0         | 20.36                 | 20.20                 | 20.19                  |                     |          |  |
| 15              | 16QAM      | 36      | 20        | 20.36                 | 20.24                 | 20.35                  |                     |          |  |

LTE Band 66

| BW [MHz]        | Modulation | RB Size | RB Offset | Power Low Ch. / Freq. | Power Middle Ch. / Freq. | Power High Ch. / Freq. | Tune-up limit (dBm) | MPR (dB) |
|-----------------|------------|---------|-----------|-----------------------|--------------------------|------------------------|---------------------|----------|
| Channel         |            |         |           | 132072                | 132322                   | 132572                 |                     |          |
| Frequency (MHz) |            |         |           | 1720                  | 1745                     | 1770                   |                     |          |
| 20              | QPSK       | 1       | 0         | 22.45                 | 22.61                    | 22.60                  | 24                  | 0        |
| 20              | QPSK       | 1       | 49        | 22.28                 | 22.29                    | 22.37                  |                     |          |
| 20              | QPSK       | 1       | 99        | 22.21                 | 22.25                    | 22.35                  |                     |          |
| 20              | QPSK       | 50      | 0         | 21.48                 | 21.50                    | 21.35                  |                     |          |
| 20              | QPSK       | 50      | 24        | 21.42                 | 21.49                    | 21.33                  | 23                  | 1        |
| 20              | QPSK       | 50      | 50        | 21.45                 | 21.35                    | 21.28                  |                     |          |
| 20              | QPSK       | 100     | 0         | 21.31                 | 21.37                    | 21.36                  |                     |          |
| 20              | 16QAM      | 1       | 0         | 21.43                 | 21.40                    | 21.53                  |                     |          |
| 20              | 16QAM      | 1       | 49        | 21.45                 | 21.42                    | 21.48                  | 23                  | 1        |
| 20              | 16QAM      | 1       | 99        | 21.54                 | 21.45                    | 21.55                  |                     |          |
| 20              | 16QAM      | 50      | 0         | 20.64                 | 20.53                    | 20.58                  |                     |          |
| 20              | 16QAM      | 50      | 24        | 20.35                 | 20.34                    | 20.25                  |                     |          |
| 20              | 16QAM      | 50      | 50        | 20.30                 | 20.25                    | 20.35                  | 22                  | 2        |
| 20              | 16QAM      | 100     | 0         | 20.34                 | 20.23                    | 20.26                  |                     |          |
| Channel         |            |         |           | 132047                | 132322                   | 132597                 | Tune-up limit (dBm) | MPR (dB) |
| Frequency (MHz) |            |         |           | 1717.5                | 1745                     | 1772.5                 |                     |          |
| 15              | QPSK       | 1       | 0         | 22.53                 | 22.49                    | 22.58                  |                     |          |
| 15              | QPSK       | 1       | 37        | 22.35                 | 22.40                    | 22.37                  |                     |          |
| 15              | QPSK       | 1       | 74        | 22.31                 | 22.28                    | 22.39                  |                     |          |
| 15              | QPSK       | 36      | 0         | 21.53                 | 21.48                    | 21.59                  | 23                  | 1        |
| 15              | QPSK       | 36      | 20        | 21.42                 | 21.43                    | 21.38                  |                     |          |
| 15              | QPSK       | 36      | 39        | 21.28                 | 21.25                    | 21.31                  |                     |          |
| 15              | QPSK       | 75      | 0         | 21.31                 | 21.30                    | 21.36                  |                     |          |
| 15              | 16QAM      | 1       | 0         | 21.48                 | 21.53                    | 21.55                  | 23                  | 1        |
| 15              | 16QAM      | 1       | 37        | 21.42                 | 21.38                    | 21.34                  |                     |          |
| 15              | 16QAM      | 1       | 74        | 21.45                 | 21.43                    | 21.44                  |                     |          |
| 15              | 16QAM      | 36      | 0         | 20.55                 | 20.48                    | 20.64                  |                     |          |
| 15              | 16QAM      | 36      | 20        | 20.38                 | 20.35                    | 20.50                  | 22                  | 2        |
| 15              | 16QAM      | 36      | 39        | 20.36                 | 20.42                    | 20.52                  |                     |          |
| 15              | 16QAM      | 75      | 0         | 20.34                 | 20.38                    | 20.45                  |                     |          |
| Channel         |            |         |           | 132022                | 132322                   | 132622                 | Tune-up limit (dBm) | MPR (dB) |
| Frequency (MHz) |            |         |           | 1715                  | 1745                     | 1775                   |                     |          |
| 10              | QPSK       | 1       | 0         | 22.51                 | 22.49                    | 22.45                  |                     |          |
| 10              | QPSK       | 1       | 25        | 22.36                 | 22.34                    | 22.29                  |                     |          |
| 10              | QPSK       | 1       | 49        | 22.29                 | 22.20                    | 22.24                  |                     |          |
| 10              | QPSK       | 25      | 0         | 21.43                 | 21.52                    | 21.52                  | 23                  | 1        |
| 10              | QPSK       | 25      | 12        | 21.34                 | 21.35                    | 21.27                  |                     |          |
| 10              | QPSK       | 25      | 25        | 21.15                 | 21.25                    | 21.18                  |                     |          |
| 10              | QPSK       | 50      | 0         | 21.23                 | 21.30                    | 21.29                  |                     |          |
| 10              | 16QAM      | 1       | 0         | 21.38                 | 21.43                    | 21.45                  | 23                  | 1        |
| 10              | 16QAM      | 1       | 25        | 21.32                 | 21.38                    | 21.40                  |                     |          |
| 10              | 16QAM      | 1       | 49        | 21.23                 | 21.45                    | 21.43                  |                     |          |
| 10              | 16QAM      | 25      | 0         | 20.65                 | 20.78                    | 20.82                  |                     |          |
| 10              | 16QAM      | 25      | 12        | 20.91                 | 20.80                    | 20.84                  | 22                  | 2        |
| 10              | 16QAM      | 25      | 25        | 20.52                 | 20.45                    | 20.43                  |                     |          |
| 10              | 16QAM      | 50      | 0         | 20.43                 | 20.49                    | 20.57                  |                     |          |
| Channel         |            |         |           | 131997                | 132322                   | 132647                 | Tune-up limit (dBm) | MPR (dB) |
| Frequency (MHz) |            |         |           | 1712.5                | 1745                     | 1777.5                 |                     |          |
| 5               | QPSK       | 1       | 0         | 22.44                 | 22.49                    | 22.51                  |                     |          |
| 5               | QPSK       | 1       | 12        | 22.26                 | 22.26                    | 22.33                  |                     |          |
| 5               | QPSK       | 1       | 24        | 22.01                 | 22.20                    | 22.27                  |                     |          |
| 5               | QPSK       | 12      | 0         | 21.38                 | 21.46                    | 21.47                  | 23                  | 1        |
| 5               | QPSK       | 12      | 7         | 21.35                 | 21.39                    | 21.34                  |                     |          |
| 5               | QPSK       | 12      | 13        | 21.15                 | 21.23                    | 21.17                  |                     |          |
| 5               | QPSK       | 25      | 0         | 21.14                 | 21.30                    | 21.21                  |                     |          |
| 5               | 16QAM      | 1       | 0         | 21.30                 | 21.39                    | 21.37                  | 23                  | 1        |
| 5               | 16QAM      | 1       | 12        | 21.22                 | 21.27                    | 21.38                  |                     |          |
| 5               | 16QAM      | 1       | 24        | 21.23                 | 21.22                    | 21.34                  |                     |          |
| 5               | 16QAM      | 12      | 0         | 20.33                 | 20.50                    | 20.53                  |                     |          |
| 5               | 16QAM      | 12      | 7         | 20.05                 | 20.24                    | 20.22                  | 22                  | 2        |
| 5               | 16QAM      | 12      | 13        | 20.11                 | 20.19                    | 20.27                  |                     |          |
| 5               | 16QAM      | 25      | 0         | 20.22                 | 20.27                    | 20.31                  |                     |          |
| Channel         |            |         |           | 131967                | 132322                   | 132657                 | Tune-up limit (dBm) | MPR (dB) |
| Frequency (MHz) |            |         |           | 1711.5                | 1745                     | 1778.5                 |                     |          |
| 3               | QPSK       | 1       | 0         | 22.40                 | 22.23                    | 22.42                  |                     |          |
| 3               | QPSK       | 1       | 8         | 22.22                 | 22.12                    | 22.24                  |                     |          |
| 3               | QPSK       | 1       | 14        | 22.06                 | 22.02                    | 22.16                  |                     |          |
| 3               | QPSK       | 8       | 0         | 21.36                 | 21.24                    | 21.43                  | 23                  | 1        |
| 3               | QPSK       | 8       | 4         | 21.24                 | 21.10                    | 21.30                  |                     |          |
| 3               | QPSK       | 8       | 7         | 21.13                 | 21.07                    | 21.19                  |                     |          |
| 3               | QPSK       | 15      | 0         | 21.16                 | 21.13                    | 21.24                  |                     |          |
| 3               | 16QAM      | 1       | 0         | 21.29                 | 21.24                    | 21.33                  | 23                  | 1        |
| 3               | 16QAM      | 1       | 8         | 21.20                 | 21.11                    | 21.30                  |                     |          |
| 3               | 16QAM      | 1       | 14        | 21.23                 | 21.22                    | 21.32                  |                     |          |
| 3               | 16QAM      | 8       | 0         | 20.40                 | 20.35                    | 20.47                  |                     |          |
| 3               | 16QAM      | 8       | 4         | 20.17                 | 20.14                    | 20.29                  | 22                  | 2        |
| 3               | 16QAM      | 8       | 7         | 20.12                 | 20.06                    | 20.24                  |                     |          |
| 3               | 16QAM      | 15      | 0         | 20.20                 | 20.15                    | 20.12                  |                     |          |
| Channel         |            |         |           | 131919                | 132322                   | 132605                 | Tune-up limit (dBm) | MPR (dB) |
| Frequency (MHz) |            |         |           | 1710.7                | 1745                     | 1775.3                 |                     |          |
| 1.4             | QPSK       | 1       | 0         | 22.44                 | 22.39                    | 22.34                  |                     |          |
| 1.4             | QPSK       | 1       | 3         | 22.28                 | 22.19                    | 22.15                  |                     |          |
| 1.4             | QPSK       | 1       | 5         | 22.31                 | 22.22                    | 22.36                  |                     |          |
| 1.4             | QPSK       | 3       | 0         | 22.51                 | 22.39                    | 22.44                  |                     |          |
| 1.4             | QPSK       | 3       | 1         | 22.43                 | 22.31                    | 22.26                  |                     |          |
| 1.4             | QPSK       | 3       | 3         | 22.19                 | 22.14                    | 22.24                  |                     |          |
| 1.4             | QPSK       | 6       | 0         | 21.17                 | 21.13                    | 21.22                  |                     |          |
| 1.4             | 16QAM      | 1       | 0         | 21.39                 | 21.24                    | 21.30                  | 23                  | 1        |
| 1.4             | 16QAM      | 1       | 3         | 21.33                 | 21.29                    | 21.16                  |                     |          |
| 1.4             | 16QAM      | 1       | 5         | 21.56                 | 21.66                    | 21.60                  |                     |          |
| 1.4             | 16QAM      | 3       | 0         | 21.39                 | 21.26                    | 21.30                  |                     |          |
| 1.4             | 16QAM      | 3       | 1         | 21.36                 | 21.27                    | 21.33                  |                     |          |
| 1.4             | 16QAM      | 3       | 3         | 21.32                 | 21.22                    | 21.26                  |                     |          |
| 1.4             | 16QAM      | 6       | 0         | 20.26                 | 20.20                    | 20.30                  |                     |          |
| 1.4             | 16QAM      | 6       | 0         | 20.26                 | 20.20                    | 20.30                  | 22                  | 2        |

LTE Band 71

| BW [MHz]        | Modulation | RB Size | RB Offset | Power Low Ch. / Freq. | Power Middle Ch. / Freq. | Power High Ch. / Freq. | Tune-up limit (dBm) | MPR (dB) |
|-----------------|------------|---------|-----------|-----------------------|--------------------------|------------------------|---------------------|----------|
| Channel         |            |         |           | 133222                | 133322                   | 133372                 |                     |          |
| Frequency (MHz) |            |         |           | 673                   | 683                      | 688                    |                     |          |
| 20              | QPSK       | 1       | 0         | 22.70                 | 22.73                    | 22.67                  | 24                  | 0        |
| 20              | QPSK       | 1       | 49        | 22.66                 | 22.54                    | 22.63                  |                     |          |
| 20              | QPSK       | 1       | 99        | 22.56                 | 22.65                    | 22.28                  |                     |          |
| 20              | QPSK       | 50      | 0         | 21.57                 | 21.59                    | 21.38                  |                     |          |
| 20              | QPSK       | 50      | 24        | 21.42                 | 21.34                    | 21.45                  | 23                  | 1        |
| 20              | QPSK       | 50      | 50        | 21.33                 | 21.28                    | 21.36                  |                     |          |
| 20              | QPSK       | 100     | 0         | 21.41                 | 21.42                    | 21.39                  |                     |          |
| 20              | 16QAM      | 1       | 0         | 21.54                 | 21.17                    | 21.64                  |                     |          |
| 20              | 16QAM      | 1       | 49        | 21.45                 | 21.54                    | 21.34                  | 23                  | 1        |
| 20              | 16QAM      | 1       | 99        | 21.56                 | 21.55                    | 21.56                  |                     |          |
| 20              | 16QAM      | 50      | 0         | 20.48                 | 20.45                    | 20.41                  |                     |          |
| 20              | 16QAM      | 50      | 24        | 20.39                 | 20.27                    | 20.32                  |                     |          |
| 20              | 16QAM      | 50      | 50        | 20.27                 | 20.45                    | 20.28                  | 22                  | 2        |
| 20              | 16QAM      | 100     | 0         | 20.35                 | 20.32                    | 20.35                  |                     |          |
| Channel         |            |         |           | 133197                | 133297                   | 133397                 | Tune-up limit (dBm) | MPR (dB) |
| Frequency (MHz) |            |         |           | 670.5                 | 680.5                    | 690.5                  |                     |          |
| 15              | QPSK       | 1       | 0         | 22.55                 | 22.56                    | 22.54                  |                     |          |
| 15              | QPSK       | 1       | 37        | 22.45                 | 22.56                    | 22.55                  |                     |          |
| 15              | QPSK       | 1       | 74        | 22.45                 | 22.40                    | 22.55                  |                     |          |
| 15              | QPSK       | 36      | 0         | 21.69                 | 21.58                    | 21.63                  | 23                  | 1        |
| 15              | QPSK       | 36      | 20        | 21.71                 | 21.59                    | 21.65                  |                     |          |
| 15              | QPSK       | 36      | 39        | 21.65                 | 21.63                    | 21.77                  |                     |          |
| 15              | QPSK       | 75      | 0         | 21.52                 | 21.54                    | 21.66                  |                     |          |
| 15              | 16QAM      | 1       | 0         | 21.55                 | 21.47                    | 21.55                  | 23                  | 1        |
| 15              | 16QAM      | 1       | 37        | 21.67                 | 21.70                    | 21.65                  |                     |          |
| 15              | 16QAM      | 1       | 74        | 21.52                 | 21.48                    | 21.32                  |                     |          |
| 15              | 16QAM      | 36      | 0         | 20.83                 | 20.69                    | 20.59                  |                     |          |
| 15              | 16QAM      | 36      | 20        | 20.73                 | 20.68                    | 20.72                  | 22                  | 2        |
| 15              | 16QAM      | 36      | 39        | 20.56                 | 20.76                    | 20.82                  |                     |          |
| 15              | 16QAM      | 75      | 0         | 20.66                 | 20.61                    | 20.72                  |                     |          |
| Channel         |            |         |           | 133172                | 133272                   | 133422                 | Tune-up limit (dBm) | MPR (dB) |
| Frequency (MHz) |            |         |           | 668                   | 678                      | 683                    |                     |          |
| 10              | QPSK       | 1       | 0         | 22.66                 | 22.56                    | 22.67                  |                     |          |
| 10              | QPSK       | 1       | 25        | 22.67                 | 22.63                    | 22.72                  |                     |          |
| 10              | QPSK       | 1       | 49        | 22.65                 | 22.56                    | 22.63                  |                     |          |
| 10              | QPSK       | 25      | 0         | 21.74                 | 21.75                    | 21.78                  | 23                  | 1        |
| 10              | QPSK       | 25      | 12        | 21.45                 | 21.78                    | 21.85                  |                     |          |
| 10              | QPSK       | 25      | 25        | 21.65                 | 21.84                    | 21.83                  |                     |          |
| 10              | QPSK       | 50      | 0         | 21.80                 | 21.45                    | 21.52                  |                     |          |
| 10              | 16QAM      | 1       | 0         | 21.69                 | 21.46                    | 21.68                  | 23                  | 1        |
| 10              | 16QAM      | 1       | 25        | 21.70                 | 21.37                    | 21.55                  |                     |          |
| 10              | 16QAM      | 1       | 49        | 21.77                 | 21.28                    | 21.47                  |                     |          |
| 10              | 16QAM      | 25      | 0         | 21.56                 | 21.55                    | 21.56                  |                     |          |
| 10              | 16QAM      | 25      | 12        | 21.65                 | 21.65                    | 21.56                  | 22                  | 2        |
| 10              | 16QAM      | 25      | 25        | 21.76                 | 21.54                    | 21.64                  |                     |          |
| 10              | 16QAM      | 50      | 0         | 21.67                 | 21.65                    | 21.67                  |                     |          |
| Channel         |            |         |           | 133147                | 133247                   | 133447                 |                     |          |
| Frequency (MHz) |            |         |           | 665.5                 | 675.5                    | 685.5                  |                     |          |
| 5               | QPSK       | 1       | 0         | 22.56                 | 22.64                    | 22.71                  |                     |          |
| 5               | QPSK       | 1       | 12        | 22.43                 | 22.58                    | 22.65                  |                     |          |
| 5               | QPSK       | 1       | 24        | 22.05                 | 22.69                    | 22.56                  |                     |          |
| 5               | QPSK       | 12      | 0         | 21.67                 | 21.73                    | 21.52                  | 23                  | 1        |
| 5               | QPSK       | 12      | 7         | 21.74                 | 21.79                    | 21.63                  |                     |          |
| 5               | QPSK       | 12      | 13        | 21.68                 | 21.67                    | 21.75                  |                     |          |
| 5               | QPSK       | 25      | 0         | 21.56                 | 21.58                    | 21.74                  |                     |          |
| 5               | 16QAM      | 1       | 0         | 21.49                 | 21.35                    | 21.65                  | 23                  | 1        |
| 5               | 16QAM      | 1       | 12        | 21.56                 | 21.30                    | 21.56                  |                     |          |
| 5               | 16QAM      | 1       | 24        | 21.42                 | 21.19                    | 21.36                  |                     |          |
| 5               | 16QAM      | 12      | 0         | 20.65                 | 20.65                    | 20.78                  |                     |          |
| 5               | 16QAM      | 12      | 7         | 20.76                 | 20.87                    | 20.99                  | 22                  | 2        |
| 5               | 16QAM      | 12      | 13        | 20.66                 | 20.56                    | 20.82                  |                     |          |
| 5               | 16QAM      | 25      | 0         | 20.65                 | 20.67                    | 20.78                  |                     |          |