telefication bv The Netherlands Chamber of Commerce 51565536 www.telefication.com



FCC SAR Exclusion Report

Product name : SWORD Health Motion tracker Alfa

Applicant : SWORD Health S.A.

FCC ID : 2AQ4A-AA1

Test report No.: 180700049 FCC RF exposure Ver 1.00



Laboratory information

Documentation

The test report must always be reproduced in full; reproduction of an excerpt only is subject to written approval of the testing laboratory. The documentation of the testing performed on the tested devices is archived for 10 years at Telefication Netherland.

Testing Location

| Test Site | Telefication BV | | |
|--------------------|-------------------|--|--|
| Test Site location | Edisonstraat 12a | | |
| | 6902 PK Zevenaar | | |
| | The Netherlands | | |
| | | | |
| | Tel. +31889983600 | | |
| | Fax. +31316583189 | | |



Revision History

| Version | Date | Remarks | Ву |
|---------|------------|-----------------|----|
| v1.00 | 02-08-2019 | Release version | KR |
| | | | |



Table of Contents

| ₹(| evision History2 | | | |
|----|------------------|--------------------------------------|---|--|
| 1 | Gene | eral Description | 4 | |
| | 1.1 | Applicant | 4 | |
| | 1.2 | Manufacturer | | |
| | 1.3 | Tested Equipment Under Test (EUT) | 4 | |
| | 1.4 | SAR Measurement Evaluation | | |
| | 1.4.1 | Maximum Output Power | | |
| | | SAR Testing Exclusions, Portable use | | |
| | 1.5 | Summary | | |



1 General Description

1.1 Applicant

Client name: SWORD Health S.A.

Address Rua Sá da Bandeira, 651, Porto, Portugal

Contact name: Ms. S. Francisco

E-mail: Sara@swordhealth.com

1.2 Manufacturer

Client name: SWORD Health S.A.

Address Rua Sá da Bandeira, 651, Porto, Portugal

Contact name: Ms. S. Francisco

E-mail: Sara@swordhealth.com

1.3 Tested Equipment Under Test (EUT)

Product name: SWORD Health Motion Tracker Alfa
Brand name: SWORD Health Motion Trackers
Product type: Inertial sensor for motion tracking

FCC ID: 2AQ4A-AA1 IC ID: 24618-AA1

Model(s):AlfaSoftware version:1.3.0Hardware version:1.0



1.4 SAR Measurement Evaluation

1.4.1 Maximum Output Power

The maximum radiated power including tune-up tolerance is shown as below.

| Mode | Output power (dBm) | |
|---------------|--------------------|--|
| Bluetooth 4.2 | 0.33* | |

^{*} from Telefication report 180700049 002 v1.00.

1.4.2 SAR Testing Exclusions, Portable use

According to KDB 447498 D01, the SAR test exclusion condition is based on source-based time-averaged maximum conducted output power, adjusted for tune-up tolerance, and the minimum test separation distance required for the exposure conditions. The SAR exclusion threshold is determined by the following formula.

1. For the test separation distance <= 50 mm

$$\frac{\text{Max.Tune up Power}_{(mW)}}{\text{Min.Test Separation Distance}_{(mm)}} \times \sqrt{f_{(GHz)}} \leq 3.0$$

When the minimum test separation distance is < 5 mm, a distance of 5 mm is applied to determine SAR test exclusion.

2. For the test separation distance > 50 mm, and the frequency at 100 MHz to 1500 MHz

$$\left[\text{(Threshold at 50 mm in Step 1)} + \text{(Test Separation Distance} - 50 \text{ mm)} \times \left(\frac{f_{\text{(MHz)}}}{150} \right) \right]_{\text{(mW)}}$$

3. For the test separation distance > 50 mm, and the frequency at > 1500 MHz to 6 GHz

[(Threshold at 50 mm in Step 1) + (Test Separation Distance -50 mm) $\times 10$]_(mW)

| | Max. Tune-up | Max. Tune-up | | | Require SAR |
|---------------|--------------|--------------|--------------|------------|-------------|
| Mode | Power | Power | Ant. to | Calculated | Testing? |
| | (dBm) | (mW) | Surface (mm) | Result | - |
| Bluetooth 4.2 | 0.33 | 1.1 | 5 | 0.34 | No |

Note:

- When separation distance <= 50 mm and the calculated result shown in above table is <= 3.0, the SARtesting exclusion is applied.
- When separation distance > 50 mm and the device output power is less than the calculated result (power threshold, mW) shown in above table, the SAR testing exclusion is applied.



1.5 Summary

Since the SAR testing for all device orientations apply SAR test exclusion per KDB 447498, SAR testing for this device is not required.