

PRODUCT LABELING

Proposed Label Location on EUT



<p>Nerve and Muscle Stimulator (Foot Drop System)</p> <p>XFT-2001E</p> <p>Input DC 3.7V, rechargeable lithium battery</p> <p>Wave form: Biphasic asymmetric balanced wave</p> <p>Frequency: 16-50Hz</p> <p>Pulse Width: 100-300μs</p> <p>FCC ID: 2AS2I2001E</p> <p>UDI</p>	<p>Output: 0-90mA</p> <p>Size: 419×103×13mm</p> <p>Weigh: 139g</p> <p>Service Life: 5 years</p> <p></p>	<p> Shenzhen XFT Medical Limited</p> <p>Room203, Building 1, Biomedicine Innovations Industrial Park, #14 Jinhai Road, Pingshan New District, Shenzhen, China</p> <p>US Agent Yang LingYun</p> <p>605 Sherwood Forest Place, Cary, North Carolina, Cary, NC 27519</p> <p>Phone: 919 4751044 Ext</p> <p> Federal law restricts this device to sale by or on the order of a practitioner licensed by the law of the State in which he/she practices to use or order the use of the device.</p> <p> </p>
---	--	---