



Date: 2022-12-14

TO:

Federal Communication Commission
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046

SUBJECT: Model Difference Letter for FCC: 2AS4N000002

To Whom It May Concern:

Due to the business consideration, we hereby declare that our product of Yushan X-ray Flat Panel Detector, Model: V17C, V17G, V14C, V14G, F14C, F14G have as following difference

Model Description:

Model		PCBA X-Board	ROIC	Scintillator	Other
Main	V17C	different size	12	CsI	Marketing Differences
Series	V17G		12	GOS	
	V14C		10	CsI	
	V14G		10	GOS	
	F14C		10	CsI	
	F14G		10	GOS	

Any questions regarding this declaration, please don't hesitate to contact us.


Sincerely,

Sean Liang

InnoCare Optoelectronics Corp
Yung Hsiang Liang / PM Director
Phone: +88665051889
E-Mail: Sean.YS.Liang@innolux.com

Model difference:

The only difference is the size of the peripheral Board as marked red column in the following figure, 14 series has 14 ROIC, 17 series has 17 ROIC, the rest main PCBA are the same. The following figure is for 17 series with 17 ROIC.



Product	Model name	Width	Length (mm)	Height (mm)	Weight (Kg)
14 series	V14C	460	383	15	2.7
	V14G	460	383	15	2.7
	F14C	460	383	15	2.5
	F14G	460	383	15	2.3
17 series	V17C	460	460	15	3.2
	V17G	460	460	15	3.2