

LABELING

Original Freigabeformular in TR archiviert.
Original form stored in the technical writing department.

Label sticks inside of the device

FA. Nr. ##### - ##
SN YYMMCE####
Removing this label causes loss of warranty.
10000021607 01

MAX. IRRIGATION RATE
PERFORMANCE MAX. IRRIGATION

3.0 l/min

MAX. PUMPING PRESSURE
PRESSION MAX. POMPE

450 mmHg

MAINS VOLTAGE RANGE
TENSION SECTEUR

AC 100-240 V~

SUPPLY FREQUENCY
FREQUENCE DU RESEAU

50/60 Hz

POWER CONSUMPTION
CONSOMMATION DE PUISSANCE

AC 100 V~ 0.72 A
AC 240 V~ 0.38 A

WARNING! Risk of fire. Replace fuse as marked.

99

YYYY-MM-DD

IP 21

COVIDIEN

LAP-Pump

Distributed by: Covidien llc,
15 Hampshire Street,
Mansfield, MA 02048 USA.
W.O.M. WORLD OF MEDICINE GmbH
Salzufer 8, 10587 Berlin, Germany
Phone: +49 30 39981-550
Fax: +49 30 39981-545,
Email: info.berlin@wom.group
PT00101905 Issuing Agency: GS1

REF PP120

SN YYMMCE####

TYPE PP120

Contains FCC ID: 2A5SK-TSHW42
Contains IC ID: 25004-TSHW42A
HVIN: 1430

CE0197

GS1 EAN
Datamatrix
01/11/21

(01)#####
(11)YYMMDD
(21)YYMMCE####

Rear of device

Diagram of the rear of the device showing various components and their part numbers:

- 10000017229 XX (Top center)
- 1109291 (Top left)
- 10000016395 XX (Top right)
- 1600041 (Right side, top)
- 1600385 (Right side, middle)
- 1601793 (Right side, bottom)
- 1600374 (Bottom center)

Box

Image of the device box with part numbers 10000022524 XX and 10000021608 XX.

Revision history

Rev. WOM	ECR	Change
01	2019107	added: RFID info on type label; replaced document no. with material no. for stickers on the rear of the device; update: design drawing of „Rear of Device“

Product Management / Customer

Project Management (R&D)

Regulatory Affairs

WORLD OF MEDICINE / Approval

Overview device label PP120 LAP-Pump Covidien

W.O.M. Item number
10000021607 01

File FMT 10000021607 01.fmt	File CDR 10000021607 01.cdr
Name Leyer	Date 11.06.2019