

LABELING

Original Freigabeformular in TR archiviert.
Original form stored in the technical writing department.

Label sticks inside of the device

FA. Nr. ##### - ##
SN YYYMCE###

Removing this label causes loss of warranty.
10000013000 04

MAX. IRRIGATION RATE1.3 l/min

MAX. PUMPING PRESSURE150 mmHg

MAINS VOLTAGE RANGEAC 100-240 V-

SUPPLY FREQUENCY50/60 Hz

POWER CONSUMPTIONAC 100 V~ 1.15 A
AC 240 V~ 0.59 A

FUSE2x T 3,15 A
250 V UL-recognized

Contains FCC ID: 2A55K-TSHW42
Contains IC: 25004-TSHW42A
HVIN: 1430

W.O.M. WORLD OF MEDICINE GmbH
Säcular 8, 10587 Berlin, Germany
Phone: +49 30 39981-850
EAN DataMatrix DB 01/1/21 (01)#####(11)YYMMDD (21)YYMCE###

CE0197
Issuing Agency: GSI

TYPE A127
SN YYYMCE###
REF REF No.

YYYYMM-DD

IP 41

Bottom of device

Rear of device

Top of device

Position Label horizontal: mittig, ca. 57,5 mm Abstand zum Geräterand (links/rechts)
vertikal: 70 mm Abstand von Geräterand, Front

Box

10000010325 XX
(für Kartonhersteller / for box producer)

Front

10000018845 XX

Rear

Revision history

Legend: Main Change request(s): Technical Writing Change Request(s)

Rev. WOM	ECR	Change
01	2013 357	W.O.M. GmbH
02	2014573	Notified body 0124 -> 0197 Symbol for AC voltage; notation of frequency indication
03	2015518	Technical specifications; UDI; new package label
04	2019304	Added: Non-ionizing electromagnetic radiation symbol, FCC information Removed: German translation, Warning: Risk of fire

Product Management / Customer

Project Management (R&D)

Regulatory Affairs

WORLD OF MEDICINE / Approval

Overview device labels A127 neutral

W.O.M. Item number

10000013000 04

File FMT

10000013000 04.fmt

Name

TR

File CDR

10000013000 04.cdr

Date

30.01.2020