

**EMC TEST REPORT****FCC 47 CFR Part 15B, ISED ICES-003 Issue 6**

<b>Report Reference No</b>	G0M-1912-8673-EF0115B-V01
<b>Testing Laboratory</b>	Eurofins Product Service GmbH
Address	Storkower Str. 38c 15526 Reichenwalde Germany
Accreditation	    DAkkS - Registration number : D-PL-12092-01-03 (ISED) ISED Testing Laboratory site: 3470A-3 DAkkS - Registration number : D-PL-12092-01-04 (FCC) FCC Filed Test Laboratory, Reg.-No.: 96970
<b>Applicant</b>	W.O.M. WORLD OF MEDICINE GmbH
Address	Salzufer 8 10587 Berlin GERMANY
<b>Test Specification</b> Standard(s)	47 CFR Part 15 Subpart B ISED ICES-003 Issue 6 ANSI C63.4:2014+A1:2017
Non-Standard Test Method	None
<b>Equipment under Test (EUT):</b>	
Product Description	Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures
Model(s)	FM300
Additional Model(s)	None
Brand Name(s)	PNEUMOCLEAR™
Hardware Version(s)	hardware traceability is constituted by serial number of device (assembled with processor board Toradex PXA320) SN: 1902CE0537
Software Version(s)	1.0.34.7
FCC-ID	2AS5K-TSHW42
IC	25004-TSHW42A
<b>Test Result</b>	<b>PASSED</b>

<b>Possible test case verdicts:</b>		
required by standard but not tested	N/T	
not required by standard	N/R	
required by standard but not appl. to test object	N/A	
test object does meet the requirement	P(PASS)	
test object does not meet the requirement	F(FAIL)	
<b>Testing:</b>		
Date of receipt of test item	2020-04-16	
<b>Report:</b>		
Compiled by	Ruslan Colbasiuc	
Tested by (+ signature) (Responsible for Test)	Ruslan Colbasiuc	
Approved by (+ signature) (Deputy Head of Lab)	Jens Marquardt	
Date of Issue	2020-09-25	
Total number of pages	44	
<b>General Remarks:</b>		
<p><b>The test results presented in this report relate only to the object tested.</b></p> <p><b>The results contained in this report reflect the results for this particular model and serial number. It is the responsibility of the manufacturer to ensure that all production models meet the intent of the requirements detailed within this report.</b></p> <p>This report shall not be reproduced, except in full, without the written approval of the issuing testing laboratory.</p>		
<b>Additional Comments:</b>		

**ABBREVIATIONS AND ACRONYMS**

<b>Acronyms</b>	
Acronym	Description
EUT	Equipment Under Test
FCC	Federal Communications Commission
ISED	Innovation, Science and Economic Development Canada
T <sub>NOM</sub>	Nominal operating temperature
V <sub>NOM</sub>	Nominal supply voltage

## VERSION HISTORY

Version History			
Version	Issue Date	Remarks	Revised By
01	2020-09-25	Initial Release	-

**REPORT INDEX**

<b>1</b>	<b>Equipment (Test Item) Under Test.....</b>	<b>6</b>
1.1	Equipment Ports.....	7
1.2	Equipment Photos - Internal.....	8
1.3	Equipment Photos – External.....	18
1.4	Support Equipment.....	24
1.5	Operational Modes.....	25
1.6	EUT Configuration.....	26
1.7	Sample emission level calculation.....	27
<b>2</b>	<b>Result Summary.....</b>	<b>28</b>
2.1	Test Conditions and Results - Radiated emissions acc. to ANSI C63.4.....	29
2.2	Test Conditions and Results - Conducted emissions acc. to ANSI C63.4.....	38



## 1 Equipment (Test Item) Under Test



Description	Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures	
Model	FM300	
Additional Model(s)	None	
Brand Name(s)	PNEUMOCLEAR™	
Serial Number(s)	1902CE0537	
Hardware Version(s)	hardware traceability is constituted by serial number of device (assembled with processor board Toradex PXA320) SN: 1902CE0537.	
Software Version(s)	1.0.34.7	
FCC-ID	2AS5K-TSHW42	
IC	25004-TSHW42A	
Class	Class B	
Equipment type	Table top	
Highest internal frequency [MHz]	806	
Dimension (cm)	31.8x14.9x42.9	
Radio Module	Type	RFID
	Model	TS-AH20
	Manufacturer	GiS GmbH
	FCC-ID	2AS5K-TSHW42
	IC	25004-TSHW42A
Supply Voltage	V <sub>NOM</sub>	120 VAC / 60 Hz
Manufacturer	W.O.M. WORLD OF MEDICINE GmbH Salzufer 8 10587 Berlin GERMANY	



## 1.1 Equipment Ports


Name	Type	Attributes	Comment
Mains Connection	AC	Count: 1 Direction: In Service only: No	-
SIDNE	IO	Count: 1 Direction: IO Service only: No	USB Type B
USB	IO	Count: 1 Direction: IO Service only: Yes	USB Type A
USB	IO	Count: 1 Direction: IO Service only: Yes	USB Type Mini B
Tube Heater	DC;IO	Count: 1 Direction: IO Service only: No	-
Gas Supply Port	NE	Count: 1 Direction: In Service only: No	-
Gas Insufflation Port	NE	Count: 1 Direction: Out Service only: No	-
Description:			
AC	AC mains power input/output port		
DC	DC power input/output port		
BAT	DC power input port connected to external battery		
IO	Input/Output port		
TP	Telecommunication port		
NE	Non-electrical port		

## 1.2 Equipment Photos - Internal






**Internal Architecture View 3****Internal Architecture View 4**



**Internal Architecture View 5****Internal Architecture View 6**


**Internal Architecture View 7****Internal Architecture View 8**


**Internal Architecture View 9****Internal Architecture View 10**

**Internal Architecture View 11****Internal Architecture View 12**



**Internal Architecture View 13****Internal Architecture View 14**

**Internal Architecture View 15****Internal Architecture View 16**

**Internal Architecture View 17****Internal Architecture View 18**



### 1.3 Equipment Photos – External



## EUT Right Side View



## EUT Back Side View



Test Report No.: G0M-1912-8673-EF0115B-V01

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany



**EUT Label****EUT Left Side View**


---

---

Test Report No.: GOM-1912-8673-EF0115B-V01

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany

**EUT Bottom View****Test Dummy**



**Gas Tube View 1****Gas Tube View 2 (with RFID TAG)**

---

---

Test Report No.: G0M-1912-8673-EF0115B-V01

Eurofins Product Service GmbH  
Storkower Str. 38c, D-15526 Reichenwalde, Germany



#### 1.4 Support Equipment

Product Type	Device	Manufacturer	Model	Comment
AE	PNEUMOCLEAR HEATED HUMIDIFIED SMOKE EVACUATION TUBE SET	W.O.M. WORLD OF MEDICINE GmbH	ST298	-
CABL	Power Cord EU 2m	-	1102131	-
CABL	Power Cord cable main US 2.5m	-	1102134	-
CABL	High pressure tube	W.O.M. WORLD OF MEDICINE GmbH	ST216	-
SIM	Abdominal cavity	W.O.M. WORLD OF MEDICINE GmbH	LapDummy Plus	-
AE	CO2 gas tank	-	-	Supply pressure > 15 bar is required
SIM	SIDNE DEVICE CONTROL	STRYKER	SDC3	SN: 16D001594
CABL	USB Cable	-	-	shielded
Description:				
AE	Auxiliary Equipment			
SIM	Simulator			
MON	Monitoring Equipment			
CBL	Connecting Cable			
Comment:				

### 1.5 Operational Modes

Mode #	Description
1	Advance Mode, Set pressure 15 mmHg; Set flow 20 l/min; Smoke evacuation level 4
Comment:	

## 1.6 EUT Configuration

Configuration #	Description
1	Insufflator connected with the gas bottle, and gas tube with heater system and RFID Tag. Gas tube connected to the test dummy with mechanical pressure sensor. USB connection with the SYDNE System established. Device powered up with 120 V / 60 Hz.
Comment:	

## 1.7 Sample emission level calculation

The following is a description of terms and a sample calculation, as appears in the radiated emissions data table. The numbers used in the calculation are for example only. There is no direct correlation to the specific data taken for the product described in this document:

Reading:

This is the reading obtained on the spectrum analyser in dB $\mu$ V. Any external preamplifiers used are taken into account through internal analyser settings.

A.F.:

This is the antenna factor for the receiving antenna. It is a conversion factor, which converts electric fields strengths to voltages, which can be measured directly on the spectrum analyser. It is treated as a loss in dB. Cable losses have been included with the A.F. to simplify the calculations. The antenna factor is used in calculations as follows:

$$\text{Reading on Analyser (dB}\mu\text{V)} + \text{A.F. (dB/m)} = \text{Net field strength (dB}\mu\text{V/m)}$$

Net:

This is the net field strength measurement (as shown above).

Limit:

This is the FCC Class B radiated emission limit (in units of dB $\mu$ V/m). The FCC limits are given in units of  $\mu$ V/m. The following formula is used to convert the units of  $\mu$ V/m to dB $\mu$ V/m:

$$\text{Limit (dB}\mu\text{V/m)} = 20 * \log (\mu\text{V/m})$$

Margin:

This is the margin of compliance below the FCC limit. The units are given in dB. A negative margin indicates the emission was below the limit. A positive margin indicates that the emission exceeds the limit.

Example only:

Reading + AF	= Net Reading	:	Net reading - FCC limit	= Margin
+21.5 dB $\mu$ V + 26 dB/m	= 47.5 dB $\mu$ V/m	:	47.5 dB $\mu$ V/m - 57.0 dB $\mu$ V/m	= -9.5 dB

## 2 Result Summary

FCC 47 CFR Part 15B, ISED ICES-003 Issue 6				
Reference	Requirement	Reference Method	Result	Remarks
Emission				
FCC 15.109 ICES-003, 6.2	Radiated emissions	ANSI C63.4:2014 +A1:2017	PASS	-
FCC 15.107 ICES-003, 6.1	AC power line conducted emissions	ANSI C63.4:2014 +A1:2017	PASS	-
Comment:				



Possible Test Case Verdicts	
PASS	Test object does meet the requirements
FAIL	Test object does not meet the requirements
N/T	Required by standard but not tested
N/R	Not required by standard for the test object



## 2.1 Test Conditions and Results - Radiated emissions acc. to ANSI C63.4

### 2.1.1 Information

Test Information	
Reference	FCC 15.109, ICES-003, 6.2
Reference method	ANSI C63.4:2014+A1:2017 Section 8
Equipment class	Class B
Equipment type	Table top
Highest internal frequency [MHz]	806
Measurement range	30 MHz to 5 GHz
Temperature [°C]	25
Humidity [%]	55
Operator	Ruslan Colbasiuc
Date	2020-08-13

### 2.1.2 Setup



**Equipment placement - Table top**

**Test Setup**

**2.1.3 Equipment**
**Test Software**

Description	Manufacturer	Name	Version
EMC Software	DARE Instruments	Radimation	2016.1.10

**Test Equipment**

Description	Manufacturer	Model	Identifier	Cal. Date	Cal. Due
Anechoic chamber	Frankonia	AC6	EF00910	2020-02	2023-02
EMI Test Receiver	R&S	ESU26	EF00887	2020-07	2021-07
TRILOG Broadband Antenna	Schwarzbeck	VULB 9162	EF00978	2019-10	2022-10
Horn Antenna	ETS-Lindgren	3117	EF00976	2019-03	2022-03
Climatic Sensor	Embedded Data Systems, LLC.	9A00100000254 77E	EF01124	2020-03	2021-03

#### 2.1.4 Procedure

<b>Exploratory measurement</b>
<ol style="list-style-type: none"> <li>1. The EUT was placed on a non-conductive table at a height of 0.8m.</li> <li>2. The EUT and support equipment, if needed, were set up to simulate typical usage.</li> <li>3. Cables, of type and length specified by the manufacturer, were connected to at least one port of each type and were terminated by a device or simulating load of actual usage.</li> <li>4. The antenna was placed at a distance of 3 or 10 m.</li> <li>5. The received signal was monitored at the measurement receiver.</li> <li>6. This procedure has to be performed in both antenna polarizations, horizontal and vertical.</li> <li>7. The arrangement of the equipment with the maximum emission level is shown on the setup picture at item 1.3</li> </ol>

<b>Final measurement</b>
<ol style="list-style-type: none"> <li>1. The EUT was placed on a 0.8 m non-conductive table at a 3 m distance from the receive antenna. The antenna output was connected to the measurement receiver.</li> <li>2. A biconical antenna was used for the frequency range 30 – 200 MHz, a logarithmic periodical antenna was used for the frequency range from 200 – 1000 MHz. Above one 1 GHz a Double Ridged Broadband Horn antenna was used. The antenna was placed on an adjustable height antenna mast.</li> <li>3. The EUT and cable arrangement were based on the exploratory measurement results.</li> <li>4. Emissions were maximized at each frequency by rotating the EUT and adjusting the receive antenna height and polarization. The maximum values were recorded.</li> <li>5. The test data of the worst-case conditions were recorded and shown on the next pages.</li> </ol>



#### 2.1.5 Limits


<b>Class B @ 3 m</b>		
Frequency [MHz]	Detector	Limit [dB $\mu$ V/m]
30 - 88	Quasi-peak	40
88 - 216	Quasi-peak	43.5
216 - 960	Quasi-peak	46
960 - 1000	Quasi-peak	54
> 1000	Peak Average	74 54

#### 2.1.6 Results

<b>Test Results</b>			
Operational mode	EUT Configuration	Verdict	Remark
1	1	PASS	

2.1.7 Setup Photos





## 2.1.8 Records

## Radiated emissions according to FCC 15B

Project Number: G0M-1912-8673  
 Applicant: W.O.M. WORLD OF MEDICINE GmbH  
 Model Description: Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures  
 Model: FM300  
 Test Sample ID: 29984 / 29985  
 Test Site: Eurofins Product Service GmbH  
 Operator: Mr. Colbasiuc  
 Test Date: 2020-08-13  
 Operating Conditions: ambient temperature: 25 °C  
 power input: 120 VAC / 60 Hz  
 Antenna: Schwarzbeck VULB 9162, Horizontal  
 Measurement Distance: 10 m calculated to 3 m  
 Mode: 1  
 Note 1:

Index 1




Peak Number	Frequency	Quasi-Peak	Quasi-Peak	Quasi-Peak	Quasi-Peak	Angle	Height
			Limit	Difference	Status		
1	440.46 MHz	37.1 dBμV/m	46 dBμV/m	-9.0 dB	Pass	-130 Degree	2.3 m
2	279.9 MHz	31.9 dBμV/m	46 dBμV/m	-14.2 dB	Pass	-130 Degree	2.3 m
3	363.852 MHz	29.2 dBμV/m	46 dBμV/m	-16.9 dB	Pass	-130 Degree	2.3 m
4	587.28 MHz	32.1 dBμV/m	46 dBμV/m	-13.9 dB	Pass	-130 Degree	2.3 m
5	818.58 MHz	43 dBμV/m	46 dBμV/m	-3.0 dB	Pass	-60 Degree	1 m

## Radiated emissions according to FCC 15B

Project Number: G0M-1912-8673  
 Applicant: W.O.M. WORLD OF MEDICINE GmbH  
 Model Description: Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures  
 Model: FM300  
 Test Sample ID: 29984 / 29985  
 Test Site: Eurofins Product Service GmbH  
 Operator: Mr. Colbasiuc  
 Test Date: 2020-08-13  
 Operating Conditions: ambient temperature: 25 °C  
 power input: 120 VAC / 60 Hz  
 Antenna: Schwarzbeck VULB 9162, Vertical  
 Measurement Distance: 10 m calculated to 3 m  
 Mode: 1  
 Note 1:

Index 2



Peak Number	Frequency	Quasi-Peak	Quasi-Peak Limit	Quasi-Peak Difference	Quasi-Peak Status	Angle	Height
1	32.7 MHz	27.6 dB $\mu$ V/m	40 dB $\mu$ V/m	-12.4 dB	Pass	-110 Degree	1 m
2	34.98 MHz	28.3 dB $\mu$ V/m	40 dB $\mu$ V/m	-11.7 dB	Pass	-110 Degree	1 m
3	49.296 MHz	25.4 dB $\mu$ V/m	40 dB $\mu$ V/m	-14.6 dB	Pass	-110 Degree	1 m
4	144.42 MHz	24.6 dB $\mu$ V/m	43.5 dB $\mu$ V/m	-18.9 dB	Pass	-110 Degree	1 m
5	196.572 MHz	29 dB $\mu$ V/m	43.5 dB $\mu$ V/m	-14.6 dB	Pass	-110 Degree	1 m
6	431.634 MHz	35 dB $\mu$ V/m	46 dB $\mu$ V/m	-11.0 dB	Pass	-110 Degree	1 m
7	709.158 MHz	35.7 dB $\mu$ V/m	46 dB $\mu$ V/m	-10.3 dB	Pass	-110 Degree	1 m


Test Report No.: G0M-1912-8673-EF0115B-V01

 Eurofins Product Service GmbH  
 Storkower Str. 38c, D-15526 Reichenwalde, Germany

## Radiated emissions according to FCC 15B

Project Number: G0M-1912-8673  
 Applicant: W.O.M. WORLD OF MEDICINE GmbH  
 Model Description: Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures  
 Model: FM300  
 Test Sample ID: 29984 / 29985  
 Test Site: Eurofins Product Service GmbH  
 Operator: Mr. Colbasiuc  
 Test Date: 2020-08-13  
 Operating Conditions: ambient temperature: 25 °C  
 power input: 120 VAC / 60 Hz  
 Antenna: ETS-Lindgren 3117, Vertical  
 Measurement Distance: 3 m  
 Mode: 1  
 Note 1: Table position 165°, Antenna high 1m


Index 3



## Radiated emissions according to FCC 15B

Project Number: G0M-1912-8673  
 Applicant: W.O.M. WORLD OF MEDICINE GmbH  
 Model Description: Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures  
 Model: FM300  
 Test Sample ID: 29984 / 29985  
 Test Site: Eurofins Product Service GmbH  
 Operator: Mr. Colbasiuc  
 Test Date: 2020-08-13  
 Operating Conditions: ambient temperature: 25 °C  
 power input: 120 VAC / 60 Hz  
 Antenna: ETS-Lindgren 3117, Horizontal  
 Measurement Distance: 3 m  
 Mode: 1  
 Note 1:

Index 4





Peak Number	Frequency	Peak	Peak Limit	Peak Difference	Peak Status	Angle	Height
1	1.362 GHz	45.4 dB $\mu$ V/m	74 dB $\mu$ V/m	-28.6 dB	Pass	-155 Degree	1 m
Peak Number	Frequency	Average	Average Limit	Average Difference	Average Status	Angle	Height
1	1.362 GHz	36.9 dB $\mu$ V/m	54 dB $\mu$ V/m	-17.1 dB	Pass	-155 Degree	1 m

## 2.2 Test Conditions and Results - Conducted emissions acc. to ANSI C63.4

### 2.2.1 Information

Test Information	
Reference	FCC 15.107, ICES-003, 6.1
Reference method	ANSI C63.4:2014+A1:2017 Section 12
Measurement range	150 kHz to 30 MHz
Equipment class	Class B
Equipment type	Table top
Temperature [°C]	25
Humidity [%]	55
Operator	Ruslan Colbasiuc
Date	2020-08-13

### 2.2.2 Setup



### 2.2.3 Equipment

<b>Test Software</b>			
Description	Manufacturer	Name	Version
EMC Software	DARE Instruments	Radimation	2016.1.10

<b>Test Equipment</b>					
Description	Manufacturer	Model	Identifier	Cal. Date	Cal. Due
AMN	Schwarzbeck	NSLK 8128	EF00975	2019-07	2021-07
EMI Test Receiver	R&S	ESU26	EF00887	2020-07	2021-07
Climatic Sensor	Embedded Data Systems, LLC.	9A00100000254 77E	EF01124	2020-03	2021-03

### 2.2.4 Procedure

<b>Exploratory measurement</b>
<ol style="list-style-type: none"> <li>1. The EUT was placed on a non conductive table 0.8 m above the reference ground plane and 0.4 m away from the vertical conducting plane (ANSI C63.4: 2014 item 7.3.1)</li> <li>2. The power cord that is normally supplied or recommended by the manufacturer was connected to the LISN.</li> <li>3. The distance between the outer edge of the EUT and the LISN shall be set to 0.8 m. A longer power cord shall be bundled to this length (bundling shall not exceed 40 cm in length).</li> <li>4. The LISN measurement port was connected to a measurement receiver</li> <li>5. I/O cables were bundled not longer than 0.4 m</li> <li>6. Measurement was performed in the frequency range 0.15 – 30MHz on each current-carrying conductor</li> <li>7. To maximize the emissions the cable positions were manipulated</li> <li>8. The worst configuration of EUT and cables is shown on a test setup picture at item 1.3</li> </ol>

<b>Final measurement</b>
<ol style="list-style-type: none"> <li>1. The EUT was placed on a non conductive table 0.8 m above the reference ground plane and 0.4 m away from the vertical conducting plane (ANSI C63.4: 2014 item 7.3.1)</li> <li>2. The power cord that is normally supplied or recommended by the manufacturer was connected to the LISN.</li> <li>3. The distance between the outer edge of the EUT and the LISN shall be set to 0.8 m. A longer power cord shall be bundled to this length (bundling shall not exceed 40 cm in length).</li> <li>4. The LISN measurement port was connected to a measurement receiver</li> <li>5. The EUT and cable arrangement were based on the exploratory measurement results</li> <li>6. The test data of the worst-case conditions were recorded and shown on the next pages</li> </ol>

### 2.2.5 Limits


<b>Class B</b>		
Frequency [MHz]	Quasi-peak Limit [dB $\mu$ V]	Average Limit [dB $\mu$ V]
0.15 - 0.5	66 - 56 *	56 - 46 *
0.5 - 5	56	46
5 - 30	60	50

\* Decreases with the logarithm of the frequency

## 2.2.6 Results

AC power line conducted emissions					
Port	Coupling	Operational mode	EUT Configuration	Verdict	Remark
Mains Connection	AMN	1	1	PASS	-

## 2.2.7 Setup Photos




## 2.2.8 Records

### Conducted emissions at the mains power port according to FCC 15B

Project Number: G0M-1912-8673  
 Applicant: W.O.M. WORLD OF MEDICINE GmbH  
 Model Description: Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures  
 Model: FM300  
 Test Sample ID: 29984 / 29985  
 Test Site: Eurofins Product Service GmbH  
 Operator: Mr. Colbasiuc  
 Test Date: 2020-08-13  
 Operating Conditions: ambient temperature: 25 °C  
 power input: 120 VAC / 60 Hz  
 LISN: Schwarzbeck NSLK 8128 (L)  
 Mode: 1  
 Applied to Port: Mains Connection  
 Note 1:

Index 7



Peak Number	Frequency	Quasi-Peak	Quasi-Peak Limit	Quasi-Peak Difference	Quasi-Peak Status	LISN
1	1.886 MHz	52.2 dB $\mu$ V	56 dB $\mu$ V	-3.8 dB	Pass	Line 1
2	1.961 MHz	51 dB $\mu$ V	56 dB $\mu$ V	-5.0 dB	Pass	Line 1
3	2.66 MHz	46.4 dB $\mu$ V	56 dB $\mu$ V	-9.6 dB	Pass	Line 1
4	5.438 MHz	51 dB $\mu$ V	60 dB $\mu$ V	-9.0 dB	Pass	Line 1
5	8.232 MHz	51.4 dB $\mu$ V	60 dB $\mu$ V	-8.6 dB	Pass	Line 1
6	5.163 MHz	51.2 dB $\mu$ V	60 dB $\mu$ V	-8.8 dB	Pass	Line 1

Test Report No.: G0M-1912-8673-EF0115B-V01


Eurofins Product Service GmbH  
 Storkower Str. 38c, D-15526 Reichenwalde, Germany

Peak Number	Frequency	Average	Average Limit	Average Difference	Average Status	LISN
1	1.886 MHz	32.9 dB $\mu$ V	46 dB $\mu$ V	-13.1 dB	Pass	Line 1
2	1.961 MHz	34.9 dB $\mu$ V	46 dB $\mu$ V	-11.1 dB	Pass	Line 1
3	2.66 MHz	29.8 dB $\mu$ V	46 dB $\mu$ V	-16.2 dB	Pass	Line 1
4	5.438 MHz	37.5 dB $\mu$ V	50 dB $\mu$ V	-12.5 dB	Pass	Line 1
5	8.232 MHz	36.3 dB $\mu$ V	50 dB $\mu$ V	-13.7 dB	Pass	Line 1
6	5.163 MHz	42 dB $\mu$ V	50 dB $\mu$ V	-8.0 dB	Pass	Line 1

## Conducted emissions at the mains power port according to FCC 15B

Project Number: G0M-1912-8673  
 Applicant: W.O.M. WORLD OF MEDICINE GmbH  
 Model Description: Insufflator for Laparoscopy, Vessel Harvesting, Colorectal and Cardiac Procedures  
 Model: FM300  
 Test Sample ID: 29984 / 29985  
 Test Site: Eurofins Product Service GmbH  
 Operator: Mr. Colbasiuc  
 Test Date: 2020-08-13  
 Operating Conditions: ambient temperature: 25 °C  
 power input: 120 VAC / 60 Hz  
 LISN: Schwarzbeck NSLK 8128 (N)  
 Mode: 1  
 Applied to Port: Mains Connection  
 Note 1:

Index 8



Peak Number	Frequency	Quasi-Peak	Quasi-Peak Limit	Quasi-Peak Difference	Quasi-Peak Status	LISN
1	1.851 MHz	52.7 dB $\mu$ V	56 dB $\mu$ V	-3.3 dB	Pass	Neutral
2	2.004 MHz	51.6 dB $\mu$ V	56 dB $\mu$ V	-4.4 dB	Pass	Neutral
3	2.116 MHz	50.2 dB $\mu$ V	56 dB $\mu$ V	-5.8 dB	Pass	Neutral
4	2.422 MHz	48.6 dB $\mu$ V	56 dB $\mu$ V	-7.4 dB	Pass	Neutral
5	8.39 MHz	51.1 dB $\mu$ V	60 dB $\mu$ V	-8.9 dB	Pass	Neutral
6	5.105 MHz	49 dB $\mu$ V	60 dB $\mu$ V	-11.0 dB	Pass	Neutral

Peak Number	Frequency	Average	Average Limit	Average Difference	Average Status	LISN
1	1.851 MHz	35.2 dB $\mu$ V	46 dB $\mu$ V	-10.8 dB	Pass	Neutral
2	2.004 MHz	34.1 dB $\mu$ V	46 dB $\mu$ V	-11.9 dB	Pass	Neutral
3	2.116 MHz	31.9 dB $\mu$ V	46 dB $\mu$ V	-14.1 dB	Pass	Neutral
4	2.422 MHz	32.2 dB $\mu$ V	46 dB $\mu$ V	-13.8 dB	Pass	Neutral
5	8.39 MHz	35.6 dB $\mu$ V	50 dB $\mu$ V	-14.4 dB	Pass	Neutral
6	5.105 MHz	32.9 dB $\mu$ V	50 dB $\mu$ V	-17.1 dB	Pass	Neutral

End of the Test Report