

June 03, 2019

FCC ID : 2ASNMTOUCHXWALLET

To whom it may concern,

We LTA LAB Inc. , hereby declare that CPR Add-on Kit/ DTS, model : TOUCH xWallet(FCC ID: 2ASNMTOUCHXWALLET)of SCTECHONE CO., Ltd.is exempt form RF exposure SAR evaluation as its output power meets the exclusion limits stated in KDB 447498D01(v06).

KDB 447498D01(v06) has the following exclusion for portable devices:

The 1g and 10g SAR test exclusion thresholds for 100MHz to 6 GHz at thest separation dsitansces \leq 50 mm are determined by :

$[(\text{max power of channel, including tune-up tolerance, mW}) / (\text{min. test separation distance, mm})]$

- $[\sqrt{f(\text{GHz})}] \leq 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR, where
- $F(\text{GHZ})$ is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation
- The result is rounded to one decimal place for comparison

The test exclusions are applicable only when the minium test separation distanceis \leq 50mm and for transmission frequencies between 100MHz and 6GHz. When the minimum test separation distance is $<$ 5 mm, a distance of 5 mm is applied to determine SAR test exclusion.

This device hs $f = 2.48$ Ghz and distance = 4 mm (minimum separation distance : 5 mm was used in the calculation) and the maximum average output power was 3 mW

So for this device:

$3 \text{ mW} / [5 \text{ mm}] * \sqrt{2.48} = 0.381$
This is less than 3.0, so no SAR is required.

Even taking imto account the tolerance, this device can be satisfied with the limits.



LTA LAB Inc.
Hee-Cheon, Kwon / Test Engineer