

EMF TEST REPORT

Test Report No. : OT-225-RWD-037
Reception No. : 2204001402
Applicant : EOFLOW Co.,Ltd
Address : Healthcare Innovation Park #2202 Dolma-ro 172 Bundang-gu Seongnam-si
Gyeonggi-do 13605, Korea
Manufacturer : EOFLOW Co.,Ltd
Address : Healthcare Innovation Park #2202 Dolma-ro 172 Bundang-gu Seongnam-si
Gyeonggi-do 13605, Korea
Type of Equipment : Insulin infusion pump
FCC ID. : 2AUGS-IDU200
Model Name : IDU200K
Multiple Model Name : N/A
Serial number : N/A
Total page of Report : 7 pages (including this page)
Date of Incoming : May 02, 2022
Date of issue : May 31, 2022

SUMMARY

The equipment complies with the regulation; **FCC PART 15 SUBPART C Section 15.247**

This test report only contains the result of a single test of the sample supplied for the examination.

It is not a generally valid assessment of the features of the respective products of the mass-production.

This report is not correlated with the "KS Q ISO/IEC 17025 and KOLAS accreditation" of Korean Laboratory Accreditation Scheme.

Tested by

/ Soon-Ki, Choi / Engineer
ONETECH Corp.

Reviewed by

/ Tae-Ho, Kim / General Manager
ONETECH Corp.

Approved by

/ Ki-Hong, Nam / General Manager
ONETECH Corp.

This Report is not correlated with the authentication of KOLAS

It should not be reproduced except in full, without the written approval of ONETECH Corp.

OTC-TRF-RF-001(0)

ONETECH Corp.: 43-14, Jinsaegol-gil, Chowol-eup, Gwangju-si, Gyeonggi-do, 12735, Korea (TEL: 82-31-799-9500, FAX: 82-31-799-9599)

CONTENTS

	PAGE
1. VERIFICATION OF COMPLIANCE	4
2. GENERAL INFORMATION.....	5
2.1 PRODUCT DESCRIPTION.....	5
2.2 ALTERNATIVE TYPE(S)/MODEL(S); ALSO COVERED BY THIS TEST REPORT.....	5
3. EUT MODIFICATIONS.....	5
4. RF EXPOSURE EVALUATION	6
4.1 RF EXPOSURE CALCULATION	6
4.2 EUT DESCRIPTION.....	6
4.3 TEST RESULT OF SAR EXCLUSION FOR DEVICES	7

Revision History

Rev. No.	Issue Report No.	Issued Date	Revisions	Section Affected
0	OT-225-RWD-037	May 31, 2022	Initial Release	All

1. VERIFICATION OF COMPLIANCE

Applicant : EOFLOW Co.,Ltd

Address : Healthcare Innovation Park #2202 Dolma-ro 172 Bundang-gu Seongnam-si Gyeonggi-do 13605, Korea

Contact Person : Junghoon Park / Manager

Telephone No. : +82-31-724-0338

FCC ID : 2AUGS-IDU200

Model Name : IDU200K

Brand Name : N/A

Serial Number : N/A

Date : May 31, 2022

EQUIPMENT CLASS	DTS – DIGITAL TRNSMISSION SYSTEM
E.U.T. DESCRIPTION	Insulin infusion pump
THIS REPORT CONCERNS	Original Grant
MEASUREMENT PROCEDURES	ANSI C63.10: 2013
TYPE OF EQUIPMENT TESTED	Pre-Production
KIND OF EQUIPMENT	Certification
AUTHORIZATION REQUESTED	
EQUIPMENT WILL BE OPERATED UNDER FCC RULES PART(S)	FCC PART 15 SUBPART C Section 15.247
Modifications on the Equipment to Achieve Compliance	None
Final Test was Conducted On	3 m, Semi Anechoic Chamber

- The above equipment was tested by ONETECH Corp. for compliance with the requirement set forth in the FCC Rules and Regulations. This said equipment in the configuration described in this report, shows the maximum emission levels emanating from equipment are within the compliance requirements.

2. GENERAL INFORMATION

2.1 Product Description

The EOFLOW Co.,Ltd, Model IDU200K (referred to as the EUT in this report) is a Insulin infusion pump. The product specification described herein was obtained from product data sheet or user's manual.

Device Type	Insulin infusion pump		
Operating Frequency	2 402 MHz ~ 2 480 MHz		
RF Output Power	Bluetooth LE	1 Mbps	0.54 dBm
		2 Mbps	0.57 dBm
Number of Channel	40 Channels		
Modulation Type	GFSK		
Antenna Type	Dielectric Chip Antenna		
Antenna Gain	-0.88 dBi		
List of each Osc. or crystal Freq.(Freq. >= 1 MHz)	48 MHz		
Rated Supply Voltage	DC 3.0 V		

2.2 Alternative type(s)/model(s); also covered by this test report.

- None

3. EUT MODIFICATIONS

- None

4. RF EXPOSURE EVALUATION

4.1 RF Exposure Calculation

According to the FCC rule §1.1310, the limit for General Population/Uncontrolled exposure is 1 mW/cm² for the device operating 1 500 ~ 100 000 MHz.

4.2 EUT Description

Kind of EUT	Insulin infusion pump
Device Category	<input checked="" type="checkbox"/> Portable (< 20 cm separation) <input type="checkbox"/> Mobile (> 20 cm separation) <input type="checkbox"/> Others
Exposure	<input type="checkbox"/> MPE
Evaluation Applied	<input type="checkbox"/> SAR <input checked="" type="checkbox"/> SAR Test Exclusion Evaluation

4.3 Test Result of SAR Exclusion for Devices

According to the procedure, KDB 447498 D01, the standalone SAR test exclusion threshold is

$$[(\text{Max. Power of channel, including tune-up tolerance, mW}) / (\text{Min. test separation distance, mm})] \times [\sqrt{f(\text{GHz})}] < 7.5$$
$$= (1.57/5) \times \sqrt{2.48} = 0.452$$

Conclusion: The SAR test exclusion threshold is less than 7.5, so the device meets the RF Exposure Requirement and excluded SAR Test.

Mode	Frequency (MHz)	Target Power W/tolerance (dBm)	Max tune up power (dBm)	Max tune up power (mW)	Separation distance (mm)	RF exposure
BLE 1 Mbps	2 480	0.54 ± 1.0	1.54	1.43	5	0.449
BLE 2 Mbps	2 480	0.57 ± 1.0	1.57	1.44	5	0.452