

Inventis SRL  
Corso Stati Uniti 1/3, 35127 - Padova (PD), Italy

Equipment Autorisation Division  
Federal Communications Commission  
7435 Oakland Mills Road  
Columbia, MD 21046

2020, 27<sup>th</sup> of April

The product referenced as:

|                      |   |
|----------------------|---|
| <b>Company:</b>      | Inventis SRL<br>Corso Stati Uniti 1/3, 35127 - Padova (PD), Italy |
| <b>Product Name:</b> | Real ear measurement system + Audiometer                          |
| <b>FCC ID:</b>       | 2AVOO-RE1RA   |

meets the requirements stated by KDB 680106 D01 v03 Section 5 (b):

| <b>KDB 680106 D01 v03 Section 5 (b) requirement</b>   | <b>Product Technical Specification</b>  | <b>Verdict</b> |
|---|---|----------------|
| Power transfer frequency is less than 1 MHz   | Operating band 0.110 - 0.205 MHz  | Complies       |
| Output power from each primary coil is less than or equal to 15 watts   | 2.1 W   | Complies       |
| The transfer system includes only single primary and secondary coils. This includes charging systems that may have multiple primary coils and clients that are able to detect and allow coupling only between individual pairs of coils | The product has a single primary and secondary coils                                | Complies       |
| Client device is placed directly in contact with the transmitter  | Probe (client devices) is placed directly in contact with the transmitter (trumpet) | Complies       |
| Mobile exposure conditions only (portable exposure conditions are not covered by this exclusion)  | The product is for mobile use only  | Complies       |
| The aggregate H-field strengths at 15 cm surrounding the device and 20 cm above the top surface from all simultaneous transmitting coils are demonstrated to be less than 50% of the MPE limit  | Test results prove conformity (see test report n. 381371-2 TRFEMF.pdf)              | Complies       |

Best Regards



(MASSIMO MARTINELLI)