


BT- CFEA-F

USER'S GUIDE



Product Information

Product name	CPR and First Aid Training Model
Model name	BT-CFEA-F
Manufacturer	BT Inc.
Address	A-313, Samsung Techno-valley, 140, Tongil-ro, Deogyang-Gu, Goyang-Si, Gyeonggi-do, 10594, Republic of Korea

Product Specifications

Dimensions	350 x 670x 215 mm
Weight	4.4kg
Battery	6EA of LR14 (C-type, 1.5V)
Power Converter	Input : AC 100-240V ~, 50/60Hz, 0.4A Output : DC 12V, 1.5A

After service

Website	www.btinc.co.kr
Address	A-313, Samsung Techno-valley, 140, Tongil-ro, Deogyang-Gu, Goyang-Si, Gyeonggi-do, 10594, Republic of Korea
Telephone	02-2038-2903
E-mail	btinc@btinc.co.kr



Cautions for battery disposal

Please follow the regulations of corresponding area for the disposal of used batteries.

■ FCC required warning statements

a. Rule Part 15.19(a)(3): This device complies with part 15 of the FCC Rules.

Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

b. Rule Part 15.21: The users manual or instruction manual for an intentional or unintentional radiator shall caution the user that changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

*Note: This equipment has been tested and found to comply with the limits for a Class B digital

*Note: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:


- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help

Cautions


Please read it carefully to ensure correct usage and user's safety through understanding of the product.

- ① Use only components provided or approved by BT Inc.
- ② Do not use the product for other purposes described in the manual.
- ③ If you disassemble or modify the product without any permission by BT Inc, you cannot get the warranty. Also, making imitated or altered products for commercial purposes are prohibited by law.
- ④ If the product is overheated or smelled something burning during the usage, turn off the power switch, unplug the power converter and call customer service center.
- ⑤ Keep the product away from water and moisture.
- ⑥ Do not use the product near strong magnetic fields or high voltage power lines.
- ⑦ Do not use the product and unplug the power converter under a weather of lightning. It can cause the malfunctions of the product and increase the risk of an electric shocks.
- ⑧ The product may malfunction under the extreme weather conditions.
(Operation temperature : 0°C~40°C (32°F~104°F), Storage temperature : -10°C~50°C (14°F~122°F))
- ⑨ Do not keep the product near hot or cold environments as possible. Deterioration of product or internal damage may occur by the extreme weather conditions
- ⑩ Never clean the product with chemicals such as solvent, benzene or thinner as these may cause fire ignition or the surface deterioration.


Components




Adult torso manikin



Magnetic papers





Power Converter



Lung airbag


*** Additional components – Storage case for manikin and User's Guide**

Product name




Power connections

<Battery power supply>





※ The battery holder is located at the backside of the manikin.



- ① Open the battery cover at the backside of the manikin by pressing the stoppers with both index fingers.
- ② Insert 6 batteries into the battery holder with correct polarities.
- ③ Use LR14 (C-type, 1.5V) standard batteries only.



<External AC Power Supply >



- ※ Use only the power converter supplied by manufacturer.
- ※ In case of using a power bank, please use an included USB cable.
- ※ Some models of power banks may not be compatible with product.

Use the magnetic papers


<AED Pads>



- ① Prepare the commercial adhesive AED Pads for training.
- ② Detach the white release papers from the magnetic papers.
- ③ Attach the magnetic papers onto the adhesive face of the AED pads.
- ④ Cut out the remains of magnetic papers from the AED pads.

Basic Functions


<Chest compression >




- ① Turn on the power switch after plug the power connector or install the batteries.
- ② When the display is turned on, select the display options and start the chest compression practice.
- ③ Compression depth, Compression rate, total compression counts and correct compression counts are displayed at the display panel.

Basic Functions

<Free Airway >




<Artificial respiration >




- ① Tilt the head of the manikin to open the airway.
- ② Pinch the nose of the manikin and try mouth to mouth resuscitation.
(Pay attention to the chest inflation.)

Basic Functions

<AED>





<Bleeding Control>



Software

<Bluetooth Connection>

- ① Turn on manikin and tablet PC.
- ② Select application,




- ③ A Bluetooth pop-up screen will appear, select the Bluetooth number (Bluetooth number Sample : CFAF0001)
- ④ Once connected, it will automatically connect next time.
- ⑤ When it is now automatically connected, Pair Bluetooth by selecting Bluetooth icon.

- * Bluetooth connection may be limited depend on the tablet.
- * Please use a recommended tablet by BT.

Software

<Orientation>



Software

<Training>



CPR



AED



Chocking



Bleeding Control

Software

<Evaluation>

