

Federal Communications Commission  
Authorization and Evaluation Division  
7435 Oakland Mills Road,  
Columbia, MD 21046

Attn: Director of Certification

**SUBJECT: BONTECH Co., Ltd.**  
**FCC ID: 2AVRHBT-DA22W**  
**DATE: February 09, 2020**

This is a description of similarities and differences between the versions of the X-ray Detectors.

Variant models are identical to basic model except for Scintillator type.

|                          | <b>Basic Model</b>                 | <b>Variant model</b>               |
|--------------------------|------------------------------------|------------------------------------|
| <b>Product Name</b>      | Flat Panel Digital X-ray Detectors | Flat Panel Digital X-ray Detectors |
| <b>Model Name</b>        | BT-DA22W                           | BT-DB22W                           |
| <b>Scintillator type</b> | CsI                                | Gdox                               |

Should you have any questions or comments concerning the above, please contact the undersigned.

Sincerely Yours,



---

Kichan An / Manager  
Title: General Manager  
Company: BONTECH Co.,Ltd