



Document number 0000718 Rev. 4
Document name Type Label Artwork - Docking Station
Product(s) Wireless intra oral sensor WIOS-S2
Author(s) Linda Kellberg
Created 7.1.2021 14.07
Last Modified By (M-Files Server), 7.1.2021 14.07

****STRICTLY CONFIDENTIAL****

Type Label Artwork – Docking Station

Version History:	Author:	Date:	Description:
1.0	Linda Kellberg	24.11.2020	First version.
2.0	Linda Kellberg	4.12.2020	Labeling specification 0000666 updated. No changes to current label artwork are required.
3.0	Linda Kellberg	16.12.2020	Labeling specification 0000666 updated. No changes to current label artwork are required.
4.0	Linda Kellberg	7.1.2021	Illustration of the label location added. No changes to label content.

TABLE OF CONTENTS

1	INTRODUCTION	3
1.1	Purpose and Scope	3
1.2	References	3
1.3	Terms and Abbreviations.....	3
2	LABEL ARTWORK.....	4
3	LABEL LOCATION	5

1 INTRODUCTION

1.1 Purpose and Scope

Purpose of this document is to describe the type label artwork of the Docking Station (DS) model WIOS-DS2 which forms a part of the Wireless IntraOral Sensor System (WIOSS) by Athlos Oy. Requirements related to the label are defined in document 0000666 [1] in sections 4.1, 4.2, and 4.4.

NOTE: Label artwork is for example purposes only. The size and positioning of symbols and text in the actual printed label may vary.

1.2 References

#	ID:	Definition:	Revision:
1	0000666	Requirement Specifications (Wireless intra oral sensor WIOS-S2) Labeling Specification	Rev. 3

1.3 Terms and Abbreviations

Term, Abbreviation:	Definition:
WIOSS	Wireless IntraOral Sensor System
DS	Docking Station

2 LABEL ARTWORK


Material:	Brady Material Type: Polyester, Brady Material No.: B-423
Size:	50.8 x 25.4 mm2

DC-Air Docking Station		Rx ONLY		
Docking Station				
#	WIOS-DS2		(01)01234567890036	
SN	00000006		(21)00000006	
	2020-11 Athlos Oy Klovinpellontie 1-3, Tower 2 01280 Espoo, Finland		FCC ID: 2AX53DCDS1	
				

1

¹ Serial number and UDI described on the label are samples only describing the length (number of characters) of the number with respect to the size of the label and positioning of the data. They do not apply to any actual product.

3 LABEL LOCATION



Signatures:

<u>Document Reviewed:</u>	I have reviewed the contents of this document	
Name:	Vasileios Grammatikakis ATHLOS\vasileios.grammatika <i>Vasileios Grammatikakis</i>	Title:
Electronically Signed in		Timestamp
		2021-01-07 10:30:59 (UTC+00:00)

<u>Document Approved:</u>	I approve this document to be immediately released for use	
Name:	Konstantinos Spartiotis ATHLOS\konstantinos.spartio <i>Konstantinos Spartiotis</i>	Title: CEO
Electronically Signed in		Timestamp
		2021-01-07 12:07:49 (UTC+00:00)