

Federal Communications Commission
Equipment Authorization Division
7435 Oakland Mills Road
Columbia, MD 21048
U.S.A.

14 March 2025

Manufacturer's Statement FCC ID location justification FCC ID: 2AXDT-RFM017

To whom it may concern

We, the undersigned, hereby submit the justification for placing the FCC-ID in the associated Instruction for User Manual (IFU) for the Hearing Aids containing the module with the above-mentioned FCC-ID.


According to KDB Publication 784748 D01, when the device is so small, or for such use that it is impracticable to label with a font size that is four points or larger (and the device does not utilize electronic labelling), then the FCC-ID shall be placed in the User Manual, and the device packaging, or on a removable label attached to the device.

Hearing Aids having the above-mentioned FCC-ID are so small and considering the dimensions of the equipment and the label area, it is impracticable to label FCC-ID with a font size required on the devices.

Therefore, the FCC-ID label is placed in the Instruction for user manual (IFU) of the end product, and on the device packaging. Instruction for module integration regarding the FCC-ID labelling is included in the integration manual.

If you have any questions, feel free to contact us at the address shown above.

Sincerely,


(date/signature) 18/03/2025

Dr. Mrinalinee Pandey
Manager Wireless System Integration
+45 44 35 56 00

mrinalinee.pandey@wsa.com


(signature)

Kristine Pedersen
Senior Director Global Regulatory Affairs
+45 21 37 09 38

kristine.pedersen@wsa.com