


A0295_Label Sample and Label Location



Compliance label location on meter



NOTES:

1. All dims. on this drawing are for reference only. Refer to M0676 for dimensional information.
2. Refer to M0676 for material information.

3. All printing to be letter pressed.

4. Broadwell standard font for 'SENTIA' branding.

5. Roboto bold font for all other text.

6. White coloured branding, text, and symbols (CMYK 0/0/0/0) or equivalent.

7. Pantone/PMS 032C

C: 0% R: 245

M: 93% G: 51

Y: 76% B: 63

K: 0%

Hex#: F5333F.

8. Pantone/PMS 403C

C: 46% R: 140

M: 42% G: 131

Y: 48% B: 123

K: 7%

Hex#: 8C837B.

9 UDI Barcode printed here in Pantone Black C. Manufacturer to build barcode variably. Remove "FPO" text, once real barcode is created and placed on artwork. UDI Barcode Build Details, see below:

Barcode Print Details

Barcode Dimension: 8.890 mm Square

Barcode Type: 2D Data Matrix

Verification Grade Required: C Or Better

Minimum Module Size: 0.380 mm

Maximum Module Size: 0.495 mm

Print Requirements: Direct Print

Variable Print Details

Model: A0295

Serial Controlled By Supplier: XXXXXX

10 Do not print magenta. Area reserved for Serial Number (S/N), which is variably printed in Pantone Black C. Serial Number will start and increment as follows: XXXXXX is incrementing, for example, 000001, 000002.

REV.	DESCRIPTION	DATE
1.1	INITIAL RELEASE	22-JUN-20
1.2	ADDED FCC ID#, IC#, HVIN, BATTERY CAPACITY, DEVICE DESCRIPTION, 'DESIGNED BY', REFERENCE DIMENSIONS / UPDATE TO POWER RATING, SERIAL NUMBER CONFIGURATION	02-NOV-20

Universal Biosensors Pty Ltd
1 CORPORATE AVE, ROWVILLE, VIC 3178, AUSTRALIA
TEL: +61 3 9213 9000 FAX: +61 3 9213 9099

TITLE: Sentia Analyzer Label Artwork	
DRAWING NO.:	79002
VERS.	1.2
DRAWN DATE:	02-NOV-20
SCALE	1:1
SHEET SIZE	A4
Commercial in Confidence © Copyright 2020 Property of Universal Biosensors Pty Ltd	