

RADISEN CO., LTD
14F Gongduk B/D, 11, Saechang-ro, Mapo-gu, Seoul, 04168, Republic of Korea

Federal Communications Commission
Authorization and Evaluation Division
1435 Oakland Mills Road
Columbia, MD 21046

Date: October 15, 2020

Regarding FCC Application (FCC ID: 2AXRZPEDRA-1417M)

This is a description of similarities and differences between the versions of the X-ray Detectors.

Variant models are identical to basic model except for Scintillator type.

| | Basic Model | Variant model |
|--------------------------|-----------------------------------|---|
| Product Name | Digital Flat Panel X-ray Detector | Digital Flat Panel X-ray Detector |
| Model Name | PEDRA-1417MC | PEDRA-1417MG DET14-MCD1 DET14-MGD1 |
| Dimension (mm) | 384 x 460 x 15 | 384 x 460 x 15 |
| TFT Type | TFT –amorphous Silicon | TFT –amorphous Silicon |
| Scintillator type | CsI:TI (Direct) | CsI:TI (Direct) GD ₂ O ₂ S:Tb |

If you need any other information or have any comment, please free to contact me.

Signature



Name: YoungTae Choi

Title: RA Manager