

# Cover Letter-Wireless Charger Approval

Date: 2021-11-25

Gentlemen:

There's a wireless charger that would like to have your authorization as an Inductive wireless power transfer applications approval.

The specific product as below, Wireless Charger Module, with its designed features and specified description, meets special requirements for KDB 680106 D01 v03r01 section 5. b) requirements.

<b>Company:</b>	Amsino International, Inc.
<b>Product Name:</b>	Digital Surgical Suction System-Docking Station
<b>Model Number:</b>	iRD301
<b>FCC ID:</b>	2AYKB-IRD301


<b>KDB 680106 D01 Section 5. b) Requirements</b>	<b>Product Technical Specification</b>	<b>Result</b>
1) Power transfer frequency is less than 1 MHz	125kHz	Complied
2) Output power from each primary coil is less than or equal to 15 watts.	120 watts	No Complied
3) The system may consist of more than one source primary coils, charging one or more clients. If more than one primary coil is present, the coil pairs may be powered on at the same time.	One source primary coils and can be powered on at the same time.	Complied
4) Client device is placed directly in contact with the transmitter.		Complied
5) Mobile exposure conditions only	Fixed	No Complied
6) The aggregate H-field strengths at 15 cm surrounding the device and 20 cm above the top surface from all simultaneous transmitting coils are demonstrated to be less than 50% of the MPE limit.	Please refer to RF exposure report. 15cm H-field strengths meet 50% of the MPE limit and can demonstrate compliance .	Complied

Sincerely,

Print Name: Elyn Li

Title: RA

Signature:



On behalf of Company: Amsino International, Inc.

Telephone: 909 6265888

Email: elyn\_li@amsino.com