

Federal Communications Commission
Authorization and Evaluation Division
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046

20210607

Re: RF exposure information for:
Product name: Nirs™ DS3 with FCC ID 2AZ6MNIRSDS3

To Whom It May Concern:

Nirs™ DS3 is a portable device intended for use within 20 cm of humans and must demonstrate compliance to the either relevant SAR limits or demonstrate a maximum conducted output power less than the SAR exclusion thresholds pursuant to “KDB 447498 D01 General RF Exposure Guidance v06” at the specified distance to user and frequency range.

The properties of Nirs™ DS3 are as follows:

- Maximum average conducted output power: 215 mW Peak mW
- Distance to user: < 50 mm
- Frequency range: 13.110-14.010 MHz

Power threshold according to KDB 447498 §4.3.1.c.2: 443 mW

The power threshold for SAR exclusion is respected hence Nirs™ DS3 is compliant.

Please contact me if there is any information you may need.

Sincerely,



Gitte Stroem, Compliance Manager

ANALYTICS BEYOND MEASURE

FOSS Analytical A/S
Nils Foss Allé 1
P O Box 260
DK-3400 Hillerød

Phone +45 7010 3370
Fax +45 7010 3371
E-mail info@foss.dk
Web www.fossanalytics.com

Danske Bank, Copenhagen
Corporate ID No. DK73399815