

Medtronic

Medtronic

RGI
6165 Gunbarrel Ave
Boulder, CO 80301
USA
www.medtronic.com

Date 6th May 2021

EMC Technologies NZ Ltd

47 Mackelvie Street
Grey Lynn, Auckland 1021
NEW ZEALAND

SUBJECT: Change of Identification

FCCID: 2AZSMKF82J

I hereby advise the following:

The module installed in the Zephyr BioModule has been certified by the FCC with the following FCC ID and Grant issue dates: VPYCMABZ, 02/08/2017.

The module has been installed in the BioModule in accordance with the module manufacturer's instructions with the module being installed on the same circuit board. There is no change in the design, circuitry, or construction, and the original test results continue to be representative of and applicable to the equipment.

In addition to the RF transmitter module, power supply and signal processing circuitry has been added in order to operate as a BioModule.

Testing has been carried out by EMC Technologies New Zealand Ltd the results of which show that the original test results continue to be representative.

The results also show that the modules can operate simultaneously.

Respectfully,



Simran Varghese
Systems Engineer