



校准结果

CALIBRATION RESULT

证书编号：S423066282
Certificate No.

第 5 页 共 6 页
Page of

Short circuit reflex

频率	实测值
Frequency	Measured
(MHz)	/
10	0.999
50	0.999
100	0.997
200	0.997
500	0.995
1000	0.990
2000	0.988
3000	0.987
4000	0.988
5000	0.987
6000	0.989
7000	0.991
8000	0.991
9000	0.992

6、短路相位

Short phase

频率	实测值
Frequency	Measured
(MHz)	(°)
10	179.72
50	178.71
100	177.46
200	175.00
500	167.82
1000	155.88
2000	132.23



校准结果

CALIBRATION RESULT

证书编号：S423066282
Certificate No.

第 6 页 共 6 页
Page of

3000	108.78
4000	89.78
5000	62.69
6000	40.52
7000	18.02
8000	-4.03
9000	-26.07

说明(Notes)

1、本次校准的测量不确定度

Measurement Uncertainty in Calibration

1.1 依据JJF 1059.1-2012 测量不确定度评定与表示

Conform JJF 1059.1-2012 *Evaluation and Expression of Uncertainty in Measurement*.

1.2 本次测量结果的扩展不确定度 ($k=2$)

The Expanded Uncertainty of the Measurement Results

- 反射系数相位(Reflective Phase) $U = 1.7^{\circ}$
- 反射系数模值 (Reflective Properties) $U = 0.029$

以下空白

End of Report